Technical methods

Separation of haemoglobins Lepore and H from A and A₂ in starch or acrylamide gels buffered with tris-E.D.T.A.

C. C. CURTAIN *From the Baker Medical Research Institute, Melbourne, Australia*¹

Owen and Got (1957) first used the high resolving power of the starch gels of Smithies (1955) to separate foetal and adult haemoglobins. Unfortunately some haemoglobin separations, notably that of Lepore from A and A₂, are difficult to achieve in gels made from buffers containing high concentrations of borate, possibly because of some degree of binding between the haemoglobins and borate ions. Recently Lehmann and Sharih (1961) described separation of Lepore haemoglobin on paper in tris-E.D.T.A. buffer. We have found that this buffer, used at a lower ionic strength, gives an excellent resolution of Lepore in starch gels, which have the advantage of a much narrower zone on which to apply the samples than paper, and can, after clarification with anhydrous glycerol (Gratzer and Beaven, 1960), be scanned photometrically without staining. Alternatively water-clear gels can be made from a polymer of 5% N-N bis-acrylamide and 95% acrylamide similar to those described by Raymond and Weintraub (1959).

The separation of haemoglobins Lepore, A, and A₂

¹Postal address: Commercial Road, Prahran S1, Melbourne, Victoria, Australia.

Received for publication 3 October 1961

FIG. 1. Electrophoresis of (a) normal, (b) Lepore, and (c) H containing haemolysates in acrylamide gels in tris-E.D.T.A. buffer.

To facilitate photographic reproduction the gels have been stained with naphthalene black and photographed with a red filter. Normally the gels are scanned unstained.

is shown in Fig. 1 (b) and of H, A, and A₂ in Fig. 1 (c). The patterns on photometric scanning at 540 mμ are shown in Fig. 2A and B.

The starch gels were made from Starch-hydrolysed (lot 154) purchased from Connaught Laboratories, Toronto, Canada, and the tris-E.D.T.A. buffer described by Craddock-Watson, Fenton, and Lehmann (1959), diluted 1 in 3.

The acrylamide gels were made from chemical grout AM9 supplied by Cynamid (Australia). The gels were prepared by polymerizing 7.5 g. of monomer per 100 ml. of solution according to the manufacturer's instructions.

FIG. 2. A, Photometric scan at 540 mμ of the pattern shown in Fig. 1a, and B of that shown in Fig. 1b.

The gels were equilibrated with three changes of excess tris-E.D.T.A. buffer for 24 hours before using.

The samples were inserted in both types of gel on slips of Whatman 3 mm. filter paper. The runs were carried out at a potential gradient of 8 volts per cm. at room temperature (approximately 16°C.) for six hours in an apparatus based on that of Smithies (1955). The top and bottom 1.5 mm. of the starch gels were sliced off and the centre section clarified in anhydrous glycerol and scanned between two microscope slides in a recording microdensitometer (Curtain, 1960). A 540 m μ second order Bausch and Lomb interference filter was placed in the light source of the low-powered ($\times 80$) optical system used. The clear acrylamide gels were scanned uncut.

The haemoglobin samples used in these experiments were supplied by Dr. B. Ryan, General Hospital, Port Moresby, and Dr. K. Dowell, Eroro Mission Station, Territory of Papua and New Guinea. I wish to acknowledge the technical assistance of Miss Joan Harris.

This work was supported in part by a grant from the National Health and Medical Research Council, Canberra, Australia.

REFERENCES

- Cradock-Watson, J. E., Fenton, J. C. B., and Lehmann, H. (1959). *J. clin. Path.*, **12**, 372.
 Curtain, C. C. (1960). *J. Sci. Instrum.*, **37**, 190.
 Gratzner, W. B., and Beaven, G. H. (1960). *Clin. Chim. Acta*, **5**, 577.
 Lehmann, H., and Sharih, A. (1961). *J. clin. Path.*, **14**, 200.
 Owen, J. A., and Got, C. (1957). *Clin. Chim. Acta*, **2**, 588.
 Raymond, S., and Weintraub, L. (1959). *Science*, **130**, 711.
 Smithies, O. (1955). *Biochem. J.*, **61**, 629.

A portable water-bath for use when investigating the coagulation mechanism

E. B. MITCHELL *From the Department of Pathology, St. Helier Hospital, Carshalton*

The recent production by Luckham Ltd. of a portable water-bath has overcome the previous difficulty of controlling anticoagulant therapy at the bedside. The bath has been used at St. Helier Hospital for some time

Received for publication 5 February 1962

FIG. 1. Apparatus without lid and handles in carrying position.

FIG. 2. Apparatus without lid and handles retracted ready for use.