Disinfection of heat-sensitive material by low-temperature steam and formaldehyde

V. G. ALDER, ANNE M. BROWN, AND W. A. GILLESPIE

From the Bristol Royal Infirmary and University of Bristol

SYNOPSIS Steam under subatmospheric pressure at temperatures below 90°C. rapidly killed non-sporing organisms after air had been removed by a high-vacuum pump. Most bacterial spores were killed but small proportions of the populations were very resistant. The destruction of spores was not logarithmic.

The addition of formaldehyde vapour to the steam greatly increased its sterilizing power, with deep penetration into fabrics and destruction of spores. Penetration into wide tubes was good, but was poor in narrow tubes. Most fabrics, plastics, and instruments were unharmed. Low-temperature steam with formaldehyde is probably as efficient a sterilizing agent as ethylene oxide.

Steam at 90°C. has been used to disinfect blankets in this hospital for the past five years (Alder and Gillespie, 1961; Alder and Leitch, 1963). This paper deals with the action of steam at temperatures between 70°C. and 90°C. on sporing and non-sporing bacteria, vaccinia virus, and bacteriophage, and with the increased sporicidal action obtained by adding formaldehyde to the steam.

METHODS

A horizontal, rectangular jacketed autoclave, of 2½ cu. ft. capacity, was used for most experiments. It was modified by the manufacturers (Drayton-Castle Limited) to operate with steam between 70° and 100°C. at pressures down to 20 in. Hg below atmospheric (Fig. 1). The steam was admitted after preliminary evacuation of air to 29·2 in. Hg below atmospheric (20 mm. Hg absolute pressure). A Drayton 'Dial set 50' controller maintained a constant subatmospheric pressure of steam inside the chamber at a pre-arranged value, by controlling the flow of steam through the inlet valve. When the pressure fell below the preset value the inlet valve opened automatically and closed again when the correct pressure was reached. In this way, temperatures were usually kept within $\pm 2^\circ\text{C}$. of the desired values. Steam and condensate escaped intermittently through a chamber discharge valve into an evacuated condenser. The other valves were operated manually and controlled by readings from temperature gauges and barometrically-compensated absolute pressure gauges.

When formaldehyde was required, formalin A.R. (38% formaldehyde w/v) was passed through a steam-heated vapouriser into the steam-to-chamber inlet pipe

before admitting steam. Subsequent evacuation to 29 in. Hg below atmospheric pressure (25 mm. absolute) removed the steam and all but a trace of formaldehyde.

Large articles were disinfected in a 55 cu. ft. jacketed high-temperature automatic dressings autoclave, which could be operated at high pressure and also, with a separate cycle, at subatmospheric pressure.

PROBLEMS ENCOUNTERED IN OPERATING THE AUTOCLAVES In earlier experiments, condensation was often excessive, but was avoided later by keeping the jacket temperature at 100°C. with the chamber at 80°C. The contents were covered with towels or cardboard to shield them from radiant heat. Thermocouple readings showed that the shielding was effective.

Temperature control was disturbed on a few occasions when air leaked into the chamber.

TEST ORGANISMS Suspensions of the following bacteria were made in sterile serum and saline to give viable counts of 10^5 to 10^6 per ml. Overnight cultures of *Staphylococcus aureus*, *Escherichia coli* and *Streptococcus faecalis*, and spore crops of *Bacillus subtilis*, *Bacillus stearothermophilus* (N.C.A. 1518) and *Clostridium sporogenes* (N.C.T.C. 276). *Cl. sporogenes* spores were prepared as described by Ingram and Handford (1957) and the others as in the Report (1958a).

Volumes, each of 0·1 ml., of all suspensions were freeze dried in small loosely plugged tubes.

Heavy suspensions of all the bacteria in serum were also dried in air on glass chips, aluminium foil, and filter paper. Oxoid spore strips of *B. stearothermophilus* were also used.

Survival of non-sporers after disinfection was investigated by incubating in broth for two days at 37°C. Spores

FIG. 1. Diagram of autoclave modified for use with subatmospheric steam and formaldehyde.

in tubes were re-suspended in distilled water and viable counts performed on the surface of nutrient agar at 37°C. (56°C. for *B. stearothermophilus*). Untreated controls were counted simultaneously. Survival of spores on paper, glass, and foil was determined by three days' incubation in broth at 37°C. (tryptone dextrose broth at 56°C. for *B. stearothermophilus*).

Vaccinia virus was prepared and tested by Dr. S. K. R. Clarke. Volumes, each of 0.2 ml., of a crude chorio-allantoic membrane preparation of titre 10^6 EID₅₀/ml. were freeze dried in 1 ml. ampoules kept at -70°C. and opened just before disinfection. The contents of each ampoule was then suspended in Earle's saline by a Waring blender, and entirely inoculated into two HeLa cell tubes and incubated for 11 days. Control ampoules contained 10^6 TCD₅₀ of virus. Bacteriophage was freeze-dried from strong suspensions and tested for survival on propagating staphylococci (Blair and Williams, 1961).

NEUTRALIZATION OF RESIDUAL FORMALDEHYDE The addition of sodium sulphite to neutralize residual formaldehyde (Report, 1958b) delayed the growth of some organisms, especially *B. stearothermophilus*, on plates. Dilution in broth was employed instead, since experiments showed that the quantities of formaldehyde which might have remained in test objects did not inhibit the sporing bacteria. Germination of *B. stearothermophilus* spores was not impaired by adding 0.025% (v/v) formalin A.R. to tryptone dextrose broth. Spores in Oxoid strips germinated after moistening with 0.02 ml. of 0.1% formalin

A.R. before incubation in 15 ml. of the broth. *B. stearothermophilus* spores remained viable after storage for 24 hours at room temperature in 0.1% formalin (v/v).

TESTS OF PENETRATION Penetration of steam and formaldehyde vapour through fabrics was determined by placing test objects and thermocouples inside a standard pack of 32 cotton towels (size 73 × 68 cm.), each folded three times. Penetration through wide- and narrow-bore tubes was also investigated. The former was 3.4 metres of polyvinyl tubing (9.5 mm. bore) containing Oxoid spore strips at intervals of 30 cm., coiled and sealed into a bleached Kraft paper bag. The narrow tube was a 0.5 mm. bore ureteric catheter of braided nylon and was investigated as suggested by Dr. J. C. Kelsey (personal communication). The shafts of two 1 in. hypodermic needles (23 gauge) were inserted into one end of each of two 3 cm. lengths of the catheter tubing and the hubs fitted to the nozzles of two 2 ml. plastic syringe barrels, previously sawn in half transversely. After inserting test objects, the cut ends of the barrels were smeared with silicone stopcock grease and tightly joined end to end by rubber pressure tubing and adhesive tape, so as to constitute a chamber which steam and formaldehyde could enter only through the catheters and needles.

Penetration through two types of paper was studied. (1) Glassine (16 lb. D.C./480 sheets) when porosity by Bendtsen tester (Hardacker, Bobb, and Wink, 1958) was nil, and (2) M.G. Bleached Kraft (21 lb. D.C./480 sheets) when porosity by Bendtsen tester was 250 ml./minute.

DISINFECTION BY LOW-TEMPERATURE STEAM

NON-SPORING ORGANISMS In numerous experiments *Staph. aureus*, *Esch. coli*, and *Str. faecalis*, dried from serum and saline, were always sterilized by steam at 80°C. for 15 minutes and at 70°C. for 20 minutes. Vaccinia virus (three tests) and bacteriophage (25 tests) were sterile after exposure to 80°C. for 30 minutes. Shorter exposure periods were not investigated.

SPORES Steam between 85° and 90°C. slowly killed the majority of *B. subtilis* and *B. stearothermophilus* spore populations in freeze-dried preparations, but some always survived after three hours (Figures 2 and 3). The survival curve for *B. subtilis* was not logarithmic, and, as the exposure time lengthened, the additional time needed for a tenfold reduction of the remaining viable spores (the decimal reduction time, or 'D' value, of Katzyn, Sandholzer, and Strong, 1943) was also increased. The curve for *B. stearothermophilus* was not determined.

Water at the same temperature as the steam was much less efficient in killing spores (Figures 2 and 3). It was to be expected that steam by releasing latent heat would kill more quickly than hot water, but the size of the difference is difficult to explain. The explanation may lie in easier penetration of spores by steam. Another possibility is inhibition by anti-rust and anti-scale chemicals in boiler steam, and although this could not be demonstrated the matter needs further investigation.

DISINFECTION OF SPORES BY STEAM WITH FORMALDEHYDE

Addition of formaldehyde to the steam, in pro-

FIG. 2. Survival of *B. subtilis* spores at 85° to 90°C. in hot water, steam, and steam with formaldehyde.

portions varying from 5 ml. to 20 ml. of formalin per cubic foot of autoclave capacity, greatly increased the rate of destruction of the spores (Figures 2 and 3 and Table I). The improvement was greatest with *B. stearothermophilus* spores, none of which survived exposure to steam and formaldehyde for three hours at 80°-85°C., whereas no tubes treated with steam alone were sterilized (Table I). In steam and formaldehyde at a higher temperature, 85°-90°C., viable counts fell to zero in less than 40 minutes (Figures 2 and 3). All spores, including those dried from serum, were killed in two hours at 80°C. and in one and a half hours at 90°C.

PENETRATION THROUGH FABRIC AND PAPER

STEAM Subatmospheric steam quickly penetrated blankets after removing air (Alder and Gillespie, 1961). In experiments with towel packs at 80°C., thermocouples showed almost instantaneous heat penetration, and *Str. faecalis* was sterilized at the centre of the pack in less than 20 minutes.

STEAM WITH FORMALDEHYDE *B. stearothermophilus* spore strips in glassine envelopes were placed at various depths in standard towel packs. The strips were often sterilized by steam with formaldehyde at lower temperatures, and always above 80°C. (Table II). The position of the strips in the packs made no material difference to the results and it was evident that the formaldehyde vapour had penetrated deeply (Table III). But the glassine envelopes distinctly hindered penetration, as shown below.

Penetration through glassine and bleached Kraft papers was compared in several experiments in which *B. stearothermophilus* strips in pairs of envelopes, one of each kind, were placed in standard

FIG. 3. Survival of *B. stearothermophilus* spores at 85° to 90°C. in hot water, steam, and steam with formaldehyde.

TABLE I
SPORICIDAL ACTION OF STEAM AND STEAM-WITH-FORMALDEHYDE AT 80°C. TO 85°C.

Organism	Disinfectant	Exposure Period	Number of Tests	
			Exposed	Sterilized
<i>B. stearothermophilus</i> in tubes	Steam	3 hours	52	0
<i>B. stearothermophilus</i> in tubes	Steam with formaldehyde	3 hours	40	40
<i>B. stearothermophilus</i> in tubes	Steam with formaldehyde	1 hour	114	49
<i>B. subtilis</i> in tubes	Steam	3 hours	9	6
<i>B. subtilis</i> in tubes	Steam with formaldehyde	3 hours	9	9
<i>B. subtilis</i> in tubes	Steam with formaldehyde	1 hour	8	7
<i>Cl. sporogenes</i> in tubes	Steam	1 hour	7	4
<i>Cl. sporogenes</i> in tubes	Steam with formaldehyde	1 hour	7	7

TABLE II
INFLUENCE OF DURATION OF EXPOSURES, TEMPERATURE, AND COMPOSITION OF STEAM/FORMALDEHYDE MIXTURE ON STERILIZATION OF *B. STEAROTHERMOPHILUS* SPORE STRIPS IN TOWEL PACKS AND POLYTHENE TUBES

Position of Strips	Temperature Range (°C.)	Volume of Formalin Injected per Cubic Foot of Autoclave Capacity (ml.)	Exposure Time (hr.)	Number of Strips	
				Exposed	Sterilized
In glassine paper envelopes inside towel packs	60-70	20	3	28	8 (28%)
	70-78	20	1	56	9 (16%)
	70-78	20	2	56	17 (30%)
	70-78	20	3	95	63 (66%)
	70-78	20	4	5	5 (100%)
	74-77	6	3	30	13 (43%)
	80-85	20	1	28	28 (100%)
	80-85	20	3	28	28 (100%)
	80-89	6	3	99	98 (99%) ¹
Bare strips at 30 cm. intervals in Polythene tube Autoclave atmosphere	70-71	20	1	11	11 (100%)
	70-78	20	2½	11	11 (100%)
	70-78	20	1	6	6 (100%)
	70-78	20	3	9	9 (100%)

¹One failure in towel wetted by condensation.

TABLE III

STERILIZATION OF *B. STEAROTHERMOPHILUS* SPORE STRIPS PLACED AT DIFFERENT DEPTHS IN STANDARD TOWEL PACKS (SUMMARIZED FROM TABLE II)

	Position of Spore Strips													
No. of towels from top of pack	2	4	6	8	10	12	14	16	18	20	22	24	26	28
No. of strips exposed	15	16	15	19	16	16	16	21	12	16	16	18	18	16
No. of strips sterilized	9	5	7	7	9	3	9	10	3	8	4	8	8	10
No. of strips sterilized in groups of eight towels	21 (46%)			28 (42%)			25 (39%)			26 (50%)				

towel packs and exposed to steam with formaldehyde (5 ml. formalin per cu. ft.) at temperatures between 71° and 84°C. Below 80°C. spores survived for two hours in 27 of 30 glassine envelopes compared with 15 of 30 in bleached Kraft. At 81° to 84°C. for one hour, the survival rate in glassine was 14/30, compared with 0/30 in bleached kraft. The same mixture freely entered cardboard boxes. Survival rates of spore strips in bleached kraft envelopes in a Bri-pac carton treated at 79° to 84°C. were 1/24 after one and a half hours and 0/20 after two hours.

Glassine envelopes, being transparent, allow recognition of enclosed instruments. This advantage can be retained, without hindering penetration, in an envelope made of both papers such as the Window-Bag (E. S. & A. Robinson).

PENETRATION INTO TUBES

Steam with formaldehyde freely entered and killed *B. stearothermophilus* in the wide tube (Table II), but steam and steam with formaldehyde penetrated badly into the narrow ureteric catheter (Table IV).

Several non-sporing bacteria survived for an hour in steam between 68° and 80°C. and one survived in steam with formaldehyde. Spore strips often contained survivors after treatment in steam with formaldehyde, even after five hours on one occasion.

DAMAGE TO INSTRUMENTS AND FABRICS

Alder and Gillespie (1961) showed that 50 exposures to steam at 90°C. did not damage woollen blankets, and subsequent experience has confirmed this. The blankets became markedly creased when bundled unfolded into sacks and disinfected without unpacking. The creases disappeared during use, but if desired can be avoided altogether, though at some cost in time, by folding and packing blankets in shaped bags (Alder and Leitch, 1963). Cotton blankets were not creased.

Steam at 80°C. severely damaged leather and slightly blistered some painted metal surfaces, though others were unaffected. Polystyrene softens at 70°C. and articles made from this material are distorted in low temperature steam above this temperature.

TABLE IV

ACTION OF STEAM AND STEAM WITH FORMALDEHYDE ON BACTERIA IN URETERIC CATHETERS

Duration of Exposure (min.)	Temperature Range (°C.)	Proportion of Cultures Not Sterilized ¹				B. stearothermophilus Spores	
		Non-sporers				Catheters	Controls ²
		Catheters					
		S. aureus	E. coli	Str. faecalis			
<i>Steam</i>							
10	68-75	1/1	1/1	1/1	} Equal numbers of controls all sterile	—	—
30	67-80	0/16	2/16	3/16		—	—
60	79-80	1/20	1/20	3/20		—	—
120	77-80	0/12	0/12	0/12		—	—
<i>Steam with formaldehyde</i>							
10	75-76	1/1	1/1	0/1	} Equal numbers of controls all sterile	1/1	1/1
30	76-77	0/7	0/6	0/7		3/7	3/5
60	75-79	1/12	0/12	0/12		5/10	0/10
120	80	0/4	0/4	0/4		0/4	0/4
180	66-69	—	—	—		1/9	0/16
240	67-73	0/1	0/1	0/1		2/4	0/10
300	67-73	—	—	—		8/18	1/30
300	81-83	0/10	0/10	0/18		1/8	0/5

¹Each fraction shows $\frac{\text{Number cultures not sterilized}}{\text{Number exposed}}$

²Control tests were placed in Petri dishes in the autoclave.
— = not tested.

The following articles were not damaged by steam alone and with formaldehyde. (Those marked * were treated once and others several times.)

At 80°C.: Nylon fur; *nylon fabric; *terylene fabric; *acrilan fabric; *rayon fabric; polyurethane; teflon tubing; epoxy resins; silicone rubber; perspex (moulded and flat); cystoscopes and other endoscopic instruments with electric light bulbs; electric leads and switches and bakelite plugs; *an electric motor; *a capacitor; ball point pens without ink reservoirs (the treatment causes the ink to ooze); *a worsted suit; *clothing for disinfection; *a baby's carry-cot; gum elastic catheters; latex rubber catheters; cuffed endotracheal tubes; a Rubens valve for anaesthetic apparatus; a cyclator for anaesthetic apparatus; a Radcliffe humidifier; a Cape respirator with pressure gauge removed; *a packet of tea (in steam only).

At 90°C.: Rubber anaesthetic face masks and tubing; tracheostomy tubes; Portex translucent polyvinyl tubing for peritoneal dialysis and perfusions; books; magazines; documents with ordinary ink writing; surgeons' boots and nurses' theatre shoes.

REMOVAL OF FORMALDEHYDE AFTER DISINFECTION

Usually no smell of formaldehyde was detected in disinfected articles though sometimes there was a faint smell on first opening the autoclave. Phenylhydrazine tests revealed traces of formaldehyde in gum elastic catheters and rubber latex tubes. The amount was too small to determine quantitatively, but was less than 0.01% in 25 g. of gum elastic catheter material (H. E. Groves, personal communication).

Residual formaldehyde might have been removed more thoroughly by flushing with steam and evacuating the chamber again but this procedure was not investigated.

DISCUSSION

The growing use of complicated surgical apparatus and plastic equipment, some of which is damaged at high temperatures, has increased the need for alternative methods of sterilization. These, though inevitably less efficient than high temperature methods, must be acceptably safe and reliable in the circumstances for which they are intended. Ethylene oxide provides one such method but it is not an easy sterilizing technique to use. Kelsey (1963) stated that because of its many variables ethylene oxide is probably the one sterilizing agent that can be effectively controlled only by bacteriological means. The results presented here suggest that steam at

subatmospheric pressures is another suitable disinfecting agent, and that when mixed with formaldehyde vapour and used above 70°C., it is comparable in potency with ethylene oxide. However, the value and limitations of the method cannot be assessed until it has been more widely used. The apparatus can be controlled by pressure (as in the present series of experiments) or by the temperature at the bottom of the chamber. Leakage of air must be avoided since it would hinder penetration and also, by disturbing temperature control, might damage heat-sensitive instruments. This danger could be altogether avoided by arranging for the steam inlet valve to close automatically if the temperature in the upper part of the chamber rises to 5° above the operating temperature.

Subatmospheric steam at 80° to 85°C. rapidly penetrated porous fabrics from which air had been removed by an efficient pump, and destroyed vaccinia virus and non-spore-forming bacteria. The steam will disinfect articles such as bedding and some endoscopic instruments, where contamination by spores is unimportant. Although subatmospheric steam at 90°C. did not sterilize spore suspensions, it destroyed the majority of spores in the populations tested and was more efficient than water at the same temperature. The survival graphs of spores in heated suspensions were in accordance with the results of Reynolds and Lichtenstein (1952) and Vas and Prosz (1957). Most spores were quickly killed but a few remained viable after prolonged heating. Such variability within one population has been attributed to differences in the physiological development of spores in suspensions prepared by ordinary cultural methods (Halvorson, 1957). Powell (1957) also demonstrated heterogeneity in spore populations. Hence different preparations of the same strain might differ in heat resistance unless survival curves were plotted with suspensions of uniform physiological state, perhaps prepared by synchronous culture. It is the existence of the few very resistant members of spore populations which makes it necessary to employ high temperatures in heat sterilizers. The chance that some spores will survive at lower temperature partly depends on the size of the population. Hence even boiling in water, a manifestly imperfect method of sterilization, has in the past been reliable for lightly contaminated objects such as well-washed instruments.

The addition of formaldehyde to subatmospheric steam made it much more lethal to spores. *Stearothermophilus* was killed almost as quickly as *B. subtilis* and *Cl. sporogenes*. In suitable conditions of temperature and humidity the main limitation to the use of formaldehyde vapour as a sterilant is its poor penetrative power (Report, 1958b). Nordgren

(1939) showed that if the temperature was raised and the chamber evacuated before admitting formaldehyde penetration was improved and spores coated with blood or sputum could be sterilized in two hours. The present work showed that formaldehyde in steam penetrated well into non-woollen fabrics and wide tubes after air had been removed by a high-vacuum pump of the type now fitted to dressing sterilizers. Early experiments demonstrated some penetration into woollen fabric, but this was not investigated further. Penetration into narrow tubes was poor, perhaps because traces of air remaining after evacuation were not displaced by the steam. This fault is shared by ethylene oxide (J. C. Kelsey, personal communication).

Cardboard for use in a low-temperature steam disinfectant should first have been sterilized at high temperatures to destroy spores (Report, 1958a).

Some margin of safety should be allowed when disinfecting instruments and equipment, though they would rarely be as heavily contaminated as in the experiments reported here. The following exposures, probably unnecessarily long, have been used for routine disinfection:—Steam, 20 minutes at 80°C. or 30 minutes at 70°C.; steam with formaldehyde (5 ml. formalin A.R. per cu. ft. of autoclave space), two hours at 80°C. for deep penetration

through fabric and two hours at 70°C. for shallow penetration, *e.g.*, through bleached Kraft paper, and surface sterilization.

We gratefully acknowledge the financial assistance of the Nuffield Provincial Hospitals Trust, and the advice and help of Dr. J. C. Kelsey, Dr. S. K. R. Clarke, Dr. J. A. Clement, Mr. K. M. Henfrey (Drayton Castle Co. Ltd.), Mr. M. I. Lees, Miss G. Patient, Dr. R. Winter (E.S. & A. Robinson Ltd.), and Mr. T. W. G. Rowe (Edwards High Vacuum International Ltd.). We also thank Mr. H. Eardley Groves (R. A. Cripps & Sons Ltd., Brighton) for estimating formaldehyde.

REFERENCES

- Alder, V. G., and Gillespie, W. A. (1961). *J. clin. Path.*, **14**, 515.
 — and Leitch, C. W. (1963). *Hospital (Lond.)*, **59**, 662.
 Blair, J. E., and Williams, R. E. O. (1961). *Bull. Wld Hlth Org.*, **24**, 771.
 Halvorson, H. O. (1957). *J. appl. Bact.*, **20**, 305.
 Hardacker, K. W., Bobb, F. C., and Wink, W. A. (1958). *J. Tech. Ass. Pulp Paper Industr.*, **41**, 191.
 Ingram, M., and Handford, P. M. (1957). *J. appl. Bact.*, **20**, 442.
 Katzyn, L. I., Sandholzer, L. A., and Strong, M. E. (1943). *J. Bact.* **45**, 265.
 Kelsey, J. C. (1963). *J. clin. Path.*, **16**, 542.
 Nordgren, G. (1939). *Acta path. microbiol. scand.*, suppl. **40**.
 Powell, E. O. (1957). *J. appl. Bact.*, **20**, 342.
 Report (1958a). *Present Sterilizing Practice in Six Hospitals*. Nuffield Provincial Hospitals Trust, London.
 Report (1958b). *J. Hyg. (Lond.)*, **56**, 488.
 Reynolds, H., and Lichtenstein, H. (1952). *Bact. Rev.*, **16**, 126.
 Vas, K., and Proszt, F. (1957). *J. appl. Bact.*, **20**, 431.