FIG. 1. The effect of chloride on the gold bromide colour development.

The optical densities are recorded on increasing chloride concentration and maintaining the bromide level at: A, 200 mg./100 ml.; B, 100 mg./100 ml.; C, 0 mg./100 ml.

normal urine and standard lines prepared from both normal and the ascorbic-acid-supplemented urine were compared. The lines obtained were identical.

The chloride concentration has a marked effect on the gold bromide colour development. This effect was noted by Malamud, Mullins, and Brown (1933), who added sodium chloride to serum until the serum chloride concentration equalled that of the cerebrospinal fluid before measuring the bromide concentrations. Wolf and Eadie (1950) diluted urine, with no prior treatment, and plasma to a constant halide concentration, a technique requiring tedious preliminary total halide measurements and dilutions. Figure 1 shows that the colour development falls to a constant rate at a chloride concentration of 44 mEq./l. In the method described the inclusion of 2 ml. 0.5% NaCl produced the required chloride concentration and the large variations encountered in urinary halide levels caused no change in colour development. The NaCl was also found to prevent turbidity in the plasma bromide estimation.

Because of the 90% recovery for urinary bromide the standards were prepared from normal pooled urine. Urines of varying pH and pigment content were used to prepare standard lines which were identical. Similarly standards were prepared from the test urines, after treatment with Dowex 2-x8 (200–400 mesh) to remove the bromide. These standards replicated those obtained from the series of normal urines. Calibration is linear up to 300 mg. Br/100 ml.

An automated plasma bromide method based on the development of gold bromide colour has been advanced

Screening for metachromatic leucodystrophy¹

C. R. READ *From the Pathology Department, Queen Mary's Hospital for Children, Carshalton*

Lake (1965) proposed a diagnostic test for metachromatic leucodystrophy based on the detection of sulphatides within renal epithelial cells in urinary deposits. These sulphatides react with certain dyes to give a brown metachromasia.

The urinary deposit often contains much debris, consisting mainly of bacteria and phosphates desired and this obscures the few cells that may be present. The method described below overcomes these difficulties.

REAGENTS AND MATERIALS

1 1% aqueous solution of cresyl fast violet (Merck) the pH adjusted to 3.5–3.6 with acetic acid.

2 Hemming² filter containing Whatman No. 41 filter paper.

3 Glycerin albumin.

PROCEDURE

The urine, which must be fresh and should not be an early morning specimen, is centrifuged at about 1,500 r.p.m. for five minutes. Most of the supernatant is decanted and the sediment resuspended in the remaining

¹This work was supported by a grant from the National Society for Mentally Handicapped Children.

²Beaumaris Instrument Co. Ltd., Rosemary Lane, Beaumaris, Anglesey. It is described in the *Oxoid Manual* (2nd ed., 1963).

Received for publication 12 December 1966.

Simple colorimetric method for the determination of bromide in urine—concluded

by Technicon Corporation but cannot be directly applied to urine because of the variation in chloride concentration and interference by urinary pigments.

The assistance of Miss G. Carter, Mr. P. Underwood, and Mr. J. T. Wearne is gratefully acknowledged.

REFERENCES

- Barbour, R. F., Pilkington, F., and Sargent, W. (1936). *Brit. med. J.*, 2, 957.
- Gray, M. G., and Moore, M. (1942). *J. Lab. clin. Med.*, 27, 680.
- Hunter, G., Smith, H. V., and Taylor, L. M. (1954). *Biochem. J.*, 56, 588.
- Malamud, W., Mullins, B. M., and Brown, J. R. (1933). *Proc. Soc. exp. Biol. (N.Y.)*, 30, 1084.
- Wolf, R. L., and Eadie, G. S. (1950). *Amer. J. Physiol.*, 163, 436.
- Wooster, A. G., Dunlop, Merle, and Joske, R. A. (1967). *Amer. J. med. Sci.*, 253, 23.

supernatant. The suspension is transferred to a bijou bottle connected to the Hemming's filter containing a Whatman no. 41 filter paper. The bottles with the filter are centrifuged at about 2,000 r.p.m. for five minutes. This retains particles of diameter greater than $5\ \mu$, including epithelial cells, but allows bacteria and smaller debris, *e.g.*, phosphates, to pass through.

Glycerin albumin is spread thickly on two 3×1 in. slides. The moist filter paper is removed from the Hemming's filter and an imprint made on the first slide. The imprinted material is relatively fluid and is mixed by means of a slide end with the glycerin albumin on the slide. The mixture is drawn down with the slide edge into several parallel thick transverse lines, to make scanning for cells easier. The filter paper is then placed firmly on the second slide and the imprinted area ringed on the reverse of the slide with a diamond. The filter paper is then discarded.

The slides are fixed in formalin vapour at 60°C . for one hour, washed in water, and stained at room temperature for 10 minutes with cresyl fast violet. The smears are finally washed in water and mounted in glycerine jelly.

This technique removes much of the debris normally found even in a fresh urine. It is possible to use a cellulose acetate filter in place of the Whatman no. 41, but being a finer filter much debris is retained.

This technique may have some value in the examination of cells in urine deposits for malignancy.

REFERENCE

Lake, B. D. (1965). *Arch. Dis. Childh.*, **40**, 284.

Glove box system for micro-determination of protein-bound iodine in children's blood

R. W. H. EDWARDS AND J. H. MCCracken *From the Endocrine Laboratory, The Hospital for Sick Children and Institute of Child Health, London*

The use of blood samples from pricked fingers or heels is essential in paediatric practice. In central London, a commensurate reduction in the scale of quantities necessary for the well-established procedure for estimating protein-bound iodine (P.B.I.) of Foss, Hanks, and Van Slyke (1960) resulted only in excessive atmospheric contamination. To overcome this the procedure is now carried out in the glove box system to be described.

APPARATUS

The glove box, shown in the figure, consists of a Dexion-supported table and working chamber. The rear wall of the system is made of block-board to provide strength. The left hand wall of asbestos carries a central hole and short sheet iron tunnel connexion to the muffle furnace. The furnace door operates without hindrance and the electrical controls were detached and reinstalled below the table.

FIG. 1. The glove box is described in the text. Dimensions (cm.) are: table height 80; main chamber 90 wide, 55 deep, 80 high; air-lock $45 \times 45 \times 45$; glove holes 15 diam., 16 from table top, centres 45 apart. Perspex parts cut from 10 mm. sheet.

Received for publication 9 November 1966.