An evaluation of the Fisher Hem-Alyzer

T. R. J. LAPPIN, A. LAMONT, AND M. G. NELSON

From the Department of Haematology, Royal Victoria Hospital, Belfast

SYNOPSIS The Fisher Hem-Alyzer is a multitest sequential discrete analyser with automatic printout of digitized results on paper tape. The instrument is well designed, soundly constructed, and reliable in routine use. Cross contamination from the sampler probe or between the cuvettes is minimal. Accuracy and precision are both highly satisfactory. Although the standard rate of throughput is 32 specimens per hour the instrument can be used at irregular intervals to analyse smaller batches or even individual specimens. In a routine daily service the Hem-Alyzer was capable of handling a workload of 200 specimens per day, requiring three hours of laboratory technician time. The Hem-Alyzer produces only three parameters. By not determining the packed cell volume it is not possible to incorporate the derived indices (MCV and MCHC) into the system and this, in the long term, is bound to be a disadvantage.

For some time, we have conducted at this centre, a programme of evaluation of multichannel haematological equipment. The result of our experience with the Technicon SMA-4 has already been reported by Lappin, Lamont, and Nelson, and with the SMA-7A by Nelson (1969).

We have applied much the same methods of scientific evaluation to a study of the Fisher Hem-Alyzer. This equipment uses the principle of discrete sequential analysis to determine the haemoglobin concentration and the red and white cell counts. Approximately 1 ml of mixed whole blood is placed in each sample cup, which is then inserted into one of the 48 numbered slots in the turntable head. Efficient mixing is achieved by a magnetic device located under the turntable which agitates each cup for a few minutes before it is presented in turn to the sampler probe. The

probe automatically removes 20 μ l of mixed whole blood, which is diluted with 5 ml of isotonic saline in a cuvette. A further dilution is carried out for the red cell count, requiring the removal of 20 μ l of the primary dilution, the transfer to a second cuvette, and the addition of a further 5 ml of saline. Subsequently, the erythrocytes are lysed when 40 μ l of a surface acting material (Hemolyte, Fisher Limited) is injected into the first cuvette and aliquots of this haemolysed solution are used for the white cell and haemoglobin determinations. For aspiration and delivery of fluids the instrument uses high quality glass-barrelled syringes.

The diluted red and white cell suspensions are presented alternately to a flow cell by a reversing flow valve. The cells are counted by means of a light scattering device, which monitors the flashes of light refracted from the cells in a dark

Received for publication 22 October 1969.

field optical system. The haemoglobin is determined as oxyhaemoglobin by a split-beam dual photocell null balance filter photometer at 525 nm and the results are then read directly as g/100 ml of whole blood.

One and a half hours are required to complete the analysis of the 48 blood samples comprising the full turntable. Three minutes after sampling the results are presented in digital form printed on paper tape. Each set of results is identified by the number of the corresponding specimen position on the sample turntable and the nature of the analysis indicated by the letters 'W' for white cell count, 'H' for haemoglobin, and 'R' for red cell count.

The instrument measures 2 ft 10 in. by 1 ft 9 in. by 1 ft 4 in. in height, thus occupying approximately 6 sq ft of bench space. The accessory items, namely the printout and a sample mixer, each require approximately 1 sq foot of bench space.

Calibration

Before the instrument was subjected to evaluation it was tested to see if the calibration of the three channels was correct. In order to test the accuracy of the haemoglobin and the red cell count channels, we used the red cell suspension which we routinely employ for standardization of our equipment (Nelson, 1969; Carville and Lee, 1969).

It was found to be a simple matter to calibrate the haemoglobin channel. After an initial warming-up period the galvanometer was zeroed by adjusting in turn the sample and reference screws and their respective lock nuts until the minimum reading was obtained. At this point the sample and reference sides of the colorimeter are in balance. The haemoglobin channel must then be calibrated for the day's run and for this purpose five aliquots of a sample of known concentration are placed in the first five sample cups of the turntable and the instrument is set in motion. When the first result appears on the printout the operator then adjusts the position of a small screw situated below the galvanometer to raise or lower the observed haemoglobin concentration to the required value. After the first adjustment the value of each successive result is visualized and final adjustments are made until the calibration of the haemoglobin channel is correct.

The instrument which was delivered to us gave results on the red cell count channel which were 1.5×10^6 per c mm too high. This inaccuracy was present for samples of varying red cell concentration. It was found that recalibration of the red cell count channel was far from a simple matter and could only be done by altering the volume of either the sample or diluent syringes. Neither of these was fitted with any reference

point from which adjustment could readily be made. As the diluent syringe was already set to deliver its maximum possible volume, we had to alter the sample syringe in order to make it deliver a somewhat smaller volume. Although by this manoeuvre we were able to correct the red cell count there was a consequential fall in the observed white cell count. By making use of red and white cell characteristic curves at different red cell threshold settings it was eventually possible to set the white cell count channel to give the correct count value. This procedure was tedious and time-consuming but once satisfactory calibration had been achieved no resetting was found to be necessary. When this difficulty in calibration was reported to the agents, they subsequently evolved a step-by-step calibration procedure, which in future will be incorporated in the operating instructional manual and they will undertake to make the calibration as part of the initial installation.

When 'particle-free' diluent was passed through the instrument at high red cell threshold settings, it was noted that a high particle count was recorded. The presence of particles in an apparently particle-free fluid was related to the filtration process, which was used in the preparation of the diluent. Because a 0.45μ filter was used to prepare the diluent it is possible for particles of less than 0.45μ to be present in the fluid and at high threshold settings to be counted by the instrument. At the threshold settings used for blood cell counting, however, the presence of such small particles in the diluting fluid would not be counted.

Apart from calibration one other practical problem is the need to exclude the entry of air bubbles into the flow lines. This may occur through the reagent tubes or at the interface of the two sections of the reverse flow valve if this has not been properly sealed.

Also during operation it is important to see that the sample probe does not impinge on the rotating turntable. The cycle of movements may fall out of sequence due to operator error. The equipment should be protected from dust and reagents rendered free from contaminating particles greater than 0.5 microns.

Once the instrument had been properly calibrated the subsequent evaluation followed a predetermined schedule designed to determine the carryover, the precision and accuracy, the mechanical and electronic stability, the maintenance and servicing requirements, the rate of throughput, the problems of data handling, and the running costs.

Carryover

It is likely that the amount of carryover in the Fisher Hem-Alyzer will be largely determined by

contamination from the sample probe, which not only takes up the original aliquot of whole blood, but also removes a 20 μ l amount of diluted blood and transfers it from the first to the second mixing cuvette. In order to reduce cross-contamination the instrument has been so designed that the stainless steel needle probe is cleaned by passing it through a filter paper disc, both before and after each sampling or transfer manoeuvre. The actual degree of contamination was investigated using a radioisotope dilution method with radioiodinated albumin as a marker. The contamination was found to be less than 0.01% which is so minimal that it is unlikely to cause any significant carryover effect.

Precision

The precision was determined as previously described (Lappin, Lamont, and Nelson, 1969) by repeated analysis of the same blood sample and calculation of the standard deviation and coefficient of variation. Twenty replicate determinations were carried out on aliquots made from the same blood specimen (Table I).

The mean coefficients of variation obtained were 2.18, 1.20, and 1.88% for leucocyte count, haemoglobin concentration, and erythrocyte count respectively.

Accuracy

To determine the accuracy 100 blood samples were divided into two aliquots and submitted to analysis on the Hem-Alyzer and by an acceptable standard method. The standard methods used were as follows: a Coulter model A cell counter was used for white cell counts by the method of Richar and Breakell (1959) and for red cell counts by the method of Brecher, Schneiderman, and Williams (1956). The haemoglobin content was determined as cyanmethaemoglobin at 540 nm in a Linson Junior colorimeter which had been calibrated with a cyanmethaemoglobin reference standard, marketed by British Drug Houses Limited, and which conforms to the ICSM International Reference Standard. Each pair of results was used for calculation of the statistical data shown in Table II.

As is shown in Table II there is no significant ($P < 0.05$) average difference between the results of the two methods for white cell counting ($t = 1.36$, degrees of freedom = 99, $0.2 > P > 0.1$) and for haemoglobin concentration ($t = 0.63$, degrees of freedom = 99, $0.6 > P > 0.5$). In order to test that the results from each pair of determinations were strongly associated, in a positive sense, the correlation coefficient (r) and regression equations were calculated. For white cell counting we found:

(1) $r = 0.9878$, degrees of freedom = 99, $P < 0.001$

and (2) $y = 0.9103x + 0.654$ (standard deviation from regression $Sy.x = 0.591$)

For haemoglobin concentration we found:

(1) $r = 0.9961$, degrees of freedom = 99, $P < 0.001$

and (2) $y = 0.9717x + 0.408$ (standard deviation from regression $Sy.x = 0.202$)

The results of the comparison of the two independent red cell counting methods show a significant value for 't' ($t = 4.39$, degrees of freedom = 99, $P < 0.001$). However, there was a very small difference between the means of the two methods, and there is strong positive correlation. For red cell counting we found:

	Range of Determination	Coefficient of Variation (%)	
		Range	Mean
Leucocytes	3.56-20.50 $\times 10^9/c\ mm$	1.21-3.24	2.18
Haemoglobin	5.74-18.70 g/100 ml	0.95-1.72	1.20
Erythrocytes	1.89-6.29 $\times 10^6/c\ mm$	1.77-2.06	1.88

Table I Precision of the Fisher Hem-Alyzer

	White Cell Count ($\times 10^9/c\ mm$)	Haemoglobin (g/100 ml)	Red Cell Count ($\times 10^6/c\ mm$)
Mean value			
Hem-Alyzer	8.363	13.212	4.602
Other methods	8.267	13.246	4.663
Paired t test			
Degrees of freedom	99	99	99
t	1.36	0.63	4.39
P	0.1-0.2	0.5-0.6	<0.001
Correlation test			
Degrees of freedom	98	98	98
Correlation coefficient	0.9878	0.9961	0.9652
P	<0.001	<0.001	<0.001
Regression equation	$y = 0.9103x + 0.654$ $Sy.x = 0.591$	$y = 0.9717x + 0.408$ $Sy.x = 0.202$	$y = 0.9651x + 0.202$ $Sy.x = 0.165$

Table II Comparison of results with Hem-Alyzer and other methods

(1) $r = 0.9652$, degrees of freedom = 98,
 $P < 0.001$
and (2) $y = 0.9651x + 0.222$ (standard deviation from regression $S_{y.x} = 0.165$).

Mechanical and Electronic Reliability

The Fisher Hem-Alyzer is of pleasing design: indeed it won an award in America in 1968 as a technical product of merit and significance. At the same time the instrument is soundly engineered and constructed of good quality materials to a high standard. As a consequence it performed reliably in routine daily use.

Provided the flow lines were regularly washed with detergent solution after each period of use, mechanical blockage was eliminated during the three-month period of study.

The electronic components of the main instrument appeared to be stable and the only electronic fault occurred in the accessory automatic printer, which intermittently failed to record one of the digits.

Maintenance and Servicing

If recalibration is necessary after replacement of

mechanical components then the new calibration procedure should be effective, although we have had no opportunity to confirm this.

The routine maintenance requirements are easily achieved and involve the cleaning of the flow lines and the replacement of small items, such as light bulbs. The provision of a 'Trouble Shooter's Manual' has been most helpful in assisting laboratory technicians to detect and correct faults in the equipment.

Rate of Throughput

A warming-up period of 20 minutes is required for the instrument to reach electronic stability. Several sample cups filled with saline, followed by at least five blood samples of known haemoglobin concentration, are placed on the sample turntable, which is then set in motion.

The rate of analysis of the instrument is 32 samples per hour. However, the number of test samples which are routinely examined will be fractionally less than this because of the introduction of a control sample in place of every twentieth test sample.

At the end of a day's run some detergent, followed by saline, is passed through the instrument and the cuvettes are then covered to protect them from dust.

Fig. 1 Comparison of duplicate white cell counts (● = single observation; ● = more than one coinciding observation).

Fig. 2 Comparison of duplicate haemoglobin determinations (● = single observation; ● = more than one coinciding observation).

Fig. 3 Comparison of duplicate red cell counts (● = single observation; ● = more than one coinciding observation).

Data Handling

While the printout of digitized results eliminates some of the problems involved in data handling, the necessity to transfer the results on to a standard hospital document from the original paper tape list is still required. The automatic production of record-compatible results would eliminate the possibility of transcription error introduced by this manual manoeuvre. It should be pointed out, however, that the equipment is provided with a computer interfaced circuit and plug, making it possible to relay the results directly to a more sophisticated data processing system.

Running Costs

The Fisher Hem-Alyzer uses only two reagent solutions, particle-free saline and Hemolyte. Sample cups and printout paper are also required and the combined cost of these expendable items is currently £1 12s 4d per hundred specimens. The total running costs are presented in Table III, which include the expendables, the staff component, and a nominal servicing charge. Amortization over a five-year period, while not necessarily related to the life of the instrument, was chosen for comparative purposes, as it was

Item	Total (£)
Staff	
Technician, part-time	75
Materials	
Reagents	325
Plastic sample cups	600
Printout paper	45
Electricity	
Maintenance and servicing	2,000
Amortization Over five years	£3,880

Table III Annual running cost of Fisher Hem-Alyzer based on a workload of 60,000 specimens/year

¹Nominal.

originally used in the cost analysis of the other items of multichannel equipment (Nelson, 1969). The costing of the Hem-Alyzer is based on a workload of 60,000 specimens per year and worked out at 1.3 shillings per specimen, which is a little cheaper than the cost per specimen obtained on both the SMA-7 and the Coulter model S, which we calculated to be approximately 1.6 shillings per specimen. However, both the SMA-4 and Coulter model S produce several parameters, whereas the Hem-Alyzer produces only three.

This study was undertaken at the request of the Laboratory Equipment and Methods Advisory Group of the Department of Health and Social Security, London.

We are grateful to Messrs Fisons Scientific Apparatus Limited for making the Hem-Alyzer available for evaluation.

References

- Brecher, G., Schneiderman, M., and Williams, G. Z. (1956). Evaluation of electronic red blood cell counter. *Amer. J. clin. Path.*, **26**, 1439-1449.
- Carville, J. M., and Lee, D. (1969). Red cell suspension as a working standard. *J. clin. Path.*, **22**, 738.
- Lappin, T. R. J., Lamont, A., and Nelson, M. G. (1969). An evaluation of the AutoAnalyzer SMA-4. *J. clin. Path.*, **22**, 11-18.
- Nelson, M. G. (1969). Multichannel continuous flow analysis on the SMA-4/-7A. *J. clin. Path.*, **22**, suppl. (Coll. Path.), 230-25.
- Richar, W. J., and Breakell, E. S. (1959). Evaluation of an electronic particle counter for the counting of white blood cells. *Amer. J. clin. Path.*, **31**, 384-393.