An assessment of the Coulter counter model S

P. H. PINKERTON, I. SPENCE, J. C. OGILVIE, W. A. RONALD, PATRICIA MARCHANT, AND P. K. RAY

From the Department of Laboratory Haematology, University of Toronto Clinic, Sunnybrook Hospital and the Institute of Applied Statistics, University of Toronto, Toronto, Canada

SYNOPSIS An assessment of the Coulter model S automatic blood counter has been carried out. The standard deviations for the haemoglobin concentration, haematocrit, red cell count and white cell count are, respectively, ± 0.15 g/100 ml, $\pm 0.45\%$, ± 0.04 m/c mm, and ± 0.4 thous/c mm. These results are clearly more accurate than careful manual estimates, performed for comparison on the same samples. Details of the comparisons are presented. A comparison is also made with routine daily estimations.

'Carryover' from one sample to the next was found to be about 2%, and tests in the range likely to be found in practice showed good linearity for the haemoglobin estimation, haematocrit, red cell count, and white cell count. A brief account of instrument failures is given.

The performance of the Coulter model S compares well with that of other automated equipment for which detailed evaluations are available.

It is well known that, in the last two decades, there has been an unremitting increase in the work load carried by hospital laboratories but in the wake of this increase have come developments in medical laboratory technology which have permitted automation of many routine methods. A large proportion of the work load in haematology involves three measurements: haemoglobin (Hb) concentration, haematocrit, and white cell count, and, to a much lesser extent, red cell count; recently automated instruments which determine these parameters have become available, eg, the Technicon AutoAnalyzer SMA-4 (see, for example, Lappin, Lamont, and Nelson, 1969) and the Coulter Electronics Inc. model S. The latter also directly measures or computes the mean cell volume (MCV), mean corpuscular

haemoglobin (MCH), and mean corpuscular haemoglobin concentration (MCHC). We now report our experience in assessing the Coulter model S for routine use in this laboratory.

Apparatus and Methods

The Coulter model S employs established principles in particle counting and size analysis (Brecher, Schneiderman, and Williams, 1956; Dacie and Lewis, 1963). The instrument consists, in effect, of six Coulter counters, three counting and sizing red cells and three counting white cells, together with a photosensitive device for measuring Hb concentration. The blood and its dilutions, and the diluting fluids, are moved through the system by a pneumatic power supply. The info-

mation from the various measuring devices is gathered and computed electronically and presented digitally.

The instrument measures the Hb concentration, red cell count, MCV and white cell count directly, and computes the haematocrit from the MCV and the red cell count. The MCH and MCHC are also calculated. The time from aspiration of the sample into the instrument until presentation of results is 40 seconds, and the machine can accept a sample every 20 seconds, counting one while the dilutions of the next are in preparation.

A flow diagram for the Coulter model S is shown in Figure 1. The instrument, which is manually fed and simple to operate once standardized, aspirates about 1.3 ml of anti-coagulated blood, most of which is used to flush away the previous sample. Only 44.7 μ l is actually used for the estimations. This volume is diluted 1:224 in isotonic particle-free fluid and

Fig. 1 Flow diagram of the Coulter model S. The blood sample is presented as indicated by the tube (top left).

the dilution split into two aliquots. The first is used for red cell counting and sizing after a further dilution to 1:50,000. The second aliquot is mixed with lysing agent, which converts haemoglobin to cyanmethaemoglobin, and destroys the red cells, leaving white cells intact; the final dilution of this aliquot is 1:250. After a fixed interval for red cell lysis and haemoglobin release and conversion to cyanmethaemoglobin, this aliquot passes through the photosensitive cells where the Hb concentration is measured, and also to the white cell counting chamber. The red cell and white cell counts are then carried out simultaneously and in triplicate. The three results for each cell count are averaged and printed out. If one result disagrees with the other two by more than 3 standard deviations of the mean (SD) it is rejected and the other two averaged and printed out. If there is a discrepancy of more than 3 SD between all three results, all three are rejected. The mean cell volume is measured similarly in duplicate and the mean figure printed out. The size distribution of both red and white cells can be observed on twin oscilloscope screens, and the patency of the holes in the aperture tubes can also be checked visually during counting.

A prediluted sample, 44.7 μ l of blood in 10 ml of diluting fluid, can be presented to the instrument through a separate aspirator which bypasses the first diluting step referred to above. This is of value in processing capillary blood samples.

The output from each of the six counters and the haemoglobinometer is monitored electronically, correction for coincidence is made automatically, and the results are fed out through an automatic digital printout. An interface can be fitted to permit direct transmission of results to a computer. (More detailed descriptions of the instrument are available from the manufacturer.)

The machine was calibrated according to the manufacturers' instructions, using their preserved blood standard 4C, before each group of measurements was carried out.

In the first group of experiments, 25 large samples of blood were taken. Each was counted by the model S 25 times consecutively. Simultaneously and independently, two technicians measured the haemoglobin concentration, the microhaematocrit, the red cell count, and the white cell count using standard manual methods (Dacie and Lewis, 1963). Haemoglobin concentrations were measured on a direct reading haemophotometer (Fisher Scientific Co.) standardized with the International Reference Standard haemoglobinocyanide solution produced for the ICSH by the Rijks Instituut voor de Volksgezondheid, Utrecht. The haemoglobinometer was standardized before each batch of results was read. Each technician performed the manual technique in duplicate and recorded the mean value. In addition, white cell counts were performed, using a Coulter model B particle

counter with high threshold set at 100, the low threshold at 11, the reciprocal of the aperture current at $\frac{1}{2}$, and the reciprocal of the amplification at $\frac{1}{3}$.

The results from the model S were analysed for reproducibility and were compared with the manual results.

In the second group of experiments, samples, on which Hb concentration, haematocrit, MCHC, and white cell count had been performed manually as part of the routine day's work, were also counted by the model S. The results were arbitrarily divided into three groups, of 135, 122, and 129 pairs, for comparison and analysis.

The degree of contamination of a sample by the previous sample ('carryover') was assessed. The model S was prepared by flushing three times with isotonic 'particle-free' solution and then the following procedure carried out:

- 1 Isotonic solution was 'counted' and the result for the 'red cell count' recorded; this was taken to represent the 'background' count and was subtracted from all subsequent counts.

- 2 A blood sample was put through the machine and the result rejected; a second count was performed on the same sample and the red cell count recorded.

- 3 Isotonic solution was then counted again and the result recorded.

- 4 'Carryover' was calculated by expressing the count obtained in step 3 as a percentage of that obtained in step 2.

- 5 Steps 2 to 4 were repeated once.

In all, this procedure, ie, steps 1-5, was carried out 10 times, using a total of 10 blood samples with red cell counts between 3.4 and 8.2 million c mm, giving 20 estimates of carryover.

Linearity of haemoglobin concentration, red and white cell counts, and haematocrit, as measured by the model S were assessed by adjusting the red cell/plasma ratio of samples of blood with added compatible plasma or by removing plasma, employing careful volumetric methods. The result obtained on the unaltered blood sample is taken as the 100% 'reference point'; the 'expected' values were then calculated and compared with the observed values obtained by counting for each of the adjusted samples.

Statistical Analysis

In analysing the first set of experiments, unless otherwise noted, we are referring to the machine performance statistics. Sample means and variances for each of the seven parameters measured were computed in the usual fashion (Mood and Graybill, 1963) and 95% confidence intervals were calculated for each parameter over all samples. In order to obtain a more compact set of measures descriptive of the precision of the model S, seven parameter means were calculated

by combining all samples, and the associated pooled within-sample variance estimates were obtained. These statistics form the basis for the calculation of the overall confidence intervals based on all 25 samples, but ignoring between sample variation (see Table V).

In addition to visual inspection, two simple statistical tests were employed to detect the presence of outliers ('wild' observations): (1) by checking that no more than one or two observations were outside the computed 95% confidence intervals for each sample, and, (2) by computing the Studentized range statistic (Scheffé, 1959). This also proved useful for detecting and eliminating transcriptional errors in handling data.

In order to test for 'drift' over time, the ratio (mean square successive difference)/(sample variance) was computed over all 25 consecutive observations for all parameters in each sample. The distribution of this statistic has been derived by Hart (1942); this measure is extremely sensitive to departures from randomness.

It is somewhat difficult to test the significance of the difference between the machine means and the technician means. This is partly because of the small number of technician estimates resulting in relatively few degrees of freedom for the technician variances, and partly because the machine variances are very low in comparison with the technician variances (see variance ratios in Table VI). Since a test based on either variance would be unsatisfactory, and since there are strong statistical arguments against pooling results in this kind of situation, a conventional *t* test was not used to determine whether the machine estimates were biased. A non-parametric binomial sign test (Bradley, 1968) was used instead, to determine whether the direction of the differences, over all samples, showed any consistent tendency.

In the second set of experiments, the degree of agreement between man and machine was assessed by computing product-moment correlation coefficients (Mood and Graybill, 1963) between the model S estimates and a technician's results.

Results

THE PRECISION OF RESULTS

The agreement between a series of estimations on each of 25 samples has been determined for each of the seven parameters measured by the Coulter model S. The results of these estimations reflect the consistency of the instrument, and are presented in Tables I-IV. The mean value for the Coulter model S and the 95% confidence limits are shown with the mean manual value for comparison. The presence of statistically significant 'drift' is indicated. In most instances,

Sample No.	Haemoglobin (g/100 ml)		Haematocrit (%)	
	Machine	Manual	Machine	Manual
1	16.30 ¹ ± 0.52	16.80	48.84 ¹ ± 1.34	48.00
2	13.56 ± 0.11	13.80	40.62 ± 0.81	39.50
3	13.92 ± 0.17	13.85	40.40 ± 0.79	39.50
4 ^a	6.55 ¹ ± 0.85	6.45	20.29 ¹ ± 2.67	20.50
5	13.77 ± 0.23	14.00	40.54 ± 0.80	41.50
6	16.32 ± 0.18	16.50	48.92 ± 0.76	48.50
7	8.53 ± 0.14	8.70	27.17 ± 0.66	27.50
8	8.04 ± 0.20	7.80	23.38 ± 0.55	23.50
9	7.13 ± 0.17	7.10	23.73 ± 0.66	25.50
10	12.27 ± 0.19	12.35	36.37 ¹ ± 0.65	42.00
11	7.36 ± 0.10	7.10	21.66 ± 0.74	22.50
12	18.83 ¹ ± 0.25	17.80	55.38 ¹ ± 0.74	54.20
13	10.08 ± 0.18	10.15	29.11 ± 0.55	29.75
14	12.64 ¹ ± 0.29	12.75	37.32 ± 0.75	37.00
15	13.26 ± 0.21	12.80	37.67 ± 0.52	37.00
16	13.60 ± 0.18	13.90	39.54 ¹ ± 0.81	40.50
17	15.76 ± 0.17	16.10	48.09 ± 0.93	54.00
18	8.79 ± 0.14	8.55	26.68 ± 0.67	28.50
19	18.76 ± 0.17	19.60	55.76 ± 0.80	55.50
20	10.75 ¹ ± 0.17	10.00	33.64 ± 0.64	35.25
21	9.35 ± 0.16	8.75	26.80 ± 0.64	30.00
22	12.62 ¹ ± 0.88	12.15	37.50 ± 0.75	40.35
23	13.96 ± 0.18	14.35	40.00 ± 0.65	41.00
24	10.49 ± 0.17	10.40	31.34 ± 0.86	35.50
25	18.82 ± 0.18	18.95	54.39 ± 0.44	53.50

Table I Precision of haemoglobin and haematocrit estimations expressed as means and 95% confidence intervals by machine compared with mean manual estimates

¹Significant drift down.

^aOne set of results in this sample was out of keeping with the other 24 (Hb = 4.6 g/100 ml); no explanation is now apparent, but this anomalous type of result was not subsequently observed in any other group of experiments.

Sample No.	Red Cell Count (m/c mm)		Mean Cell Volume (cu)	
	Machine	Manual	Machine	Manual
1	5.36 ± 0.11	5.46	91.44 ± 2.24	87.0
2	4.55 ¹ ± 0.07	5.22	87.72 ± 1.40	82.50
3	4.50 ± 0.08	4.71	89.84 ± 0.98	83.50
4 ^a	2.35 ¹ ± 0.21	2.04	86.12 ± 1.72	99.50
5	4.49 ± 0.06	4.51	89.60 ± 1.62	92.00
6	4.97 ± 0.09	5.45	98.52 ± 1.99	90.50
7	3.11 ± 0.04	3.24	87.44 ± 1.20	85.00
8	2.80 ¹ ± 0.07	2.65	84.96 ± 1.74	89.50
9	3.81 ± 0.07	3.98	60.44 ± 1.34	65.00
10	4.24 ¹ ± 0.06	4.56	84.64 ± 1.67	92.00
11	2.22 ± 0.05	2.40	97.52 ¹ ± 1.70	90.50
12	6.23 ± 0.06	6.29	88.28 ± 1.52	86.00
13	3.19 ± 0.06	3.03	90.48 ± 1.70	98.00
14	4.38 ± 0.06	4.52	84.24 ± 1.49	80.00
15	4.09 ¹ ± 0.06	4.01	90.04 ± 2.02	92.50
16	4.21 ± 0.09	4.27	93.64 ± 1.32	95.50
17	6.10 ± 0.10	6.67	77.40 ± 1.19	81.00
18	3.32 ± 0.09	3.59	78.44 ± 1.80	79.00
19	5.90 ± 0.08	6.10	92.92 ± 1.68	90.50
20	4.70 ¹ ± 0.07	4.43	70.44 ^a ± 1.53	79.00
21	3.45 ¹ ± 0.08	3.33	78.96 ± 1.93	88.50
22	4.76 ¹ ± 0.08	5.12	77.92 ± 1.57	80.50
23	4.34 ± 0.06	4.45	90.76 ± 1.49	92.00
24	3.62 ± 0.08	3.85	85.68 ± 1.66	92.50
25	5.57 ± 0.06	5.90	95.80 ± 1.84	88.50

Table II Precision of red cell count and mean cell volume estimations expressed as means and 95% confidence intervals by machine compared with mean manual estimates

¹, ² see Table I.

^a Significant drift up.

Sample No.	MCH (pg)		MCHC (%)	
	Machine	Manual	Machine	Manual
1	30.78 ¹ ± 1.48	30.50	33.50 ¹ ± 1.67	34.00
2	30.22 ± 0.79	27.80	33.90 ± 0.70	34.80
3	31.23 ± 0.44	29.25	34.55 ± 0.76	35.00
4	28.00 ± 0.75	34.50	32.13 ± 0.85	31.20
5	30.67 ± 0.55	30.45	33.83 ± 0.78	34.00
6	33.02 ± 0.54	30.20	33.32 ± 0.48	33.95
7	28.01 ± 0.50	27.75	32.10 ± 0.76	31.10
8	29.21 ¹ ± 0.87	29.15	34.50 ¹ ± 0.98	33.00
9	19.12 ± 0.32	18.15	30.74 ± 0.99	27.45
10	29.13 ¹ ± 0.61	27.10	34.08 ¹ ± 0.82	29.50
11	33.70 ± 0.70	29.60	33.40 ± 1.00	31.60
12	34.10 ± 0.48	31.05	30.04 ± 0.33	32.95
13	31.87 ± 0.65	31.40	35.16 ± 0.68	33.90
14	28.93 ¹ ± 0.77	27.80	34.24 ± 1.04	34.20
15	31.90 ± 0.54	31.80	35.27 ± 0.54	34.45
16	32.41 ± 0.56	32.05	34.44 ± 0.62	34.30
17	25.95 ± 0.47	24.05	33.30 ± 0.58	29.75
18	26.75 ± 0.77	23.95	33.84 ± 0.90	29.85
19	31.96 ± 0.44	30.50	34.00 ± 0.45	34.70
20	22.91 ± 0.34	22.45	32.10 ± 0.57	28.65
21	27.11 ² ± 0.57	27.40	34.40 ± 0.79	29.15
22	26.10 ² ± 1.78	24.20	33.71 ² ± 2.20	30.05
23	32.16 ± 0.49	32.05	35.12 ± 0.60	34.85
24	29.05 ± 0.58	27.10	33.69 ± 1.35	29.50
25	33.71 ± 0.40	32.15	34.70 ± 0.42	35.40

Table III Precision of mean corpuscular haemoglobin and mean corpuscular haemoglobin concentration estimations expressed as means and 95% confidence intervals by machine compared with mean manual estimations

¹, ² see Table I.

³ see Table II.

Sample No.	White Cell Count (1,000/c mm)		
	Model S	Manual	Model B
1	6.79 ¹ ± 1.70	5.95	5.30
2	6.36 ³ ± 0.46	6.50	6.50
3	6.96 ³ ± 1.10	6.80	6.50
4 ²	10.27 ± 2.22	10.40	12.25
5	4.46 ± 0.22	4.70	5.35
6	9.04 ± 0.28	10.45	9.50
7	6.59 ± 2.42	7.60	6.25
8	5.40 ³ ± 1.12	6.30	5.25
9	7.18 ± 0.24	7.70	7.70
10	4.95 ± 0.26	6.35	5.80
11	3.99 ± 0.23	3.90	4.05
12	9.20 ± 0.21	8.45	9.00
13	8.54 ¹ ± 0.34	7.75	8.10
14	7.58 ± 0.26	7.85	7.95
15	5.51 ± 0.25	5.85	5.80
16	6.93 ± 0.29	7.75	7.75
17	7.38 ± 0.46	6.85	7.50
18	30.66 ± 2.06	37.65	31.70
19	10.67 ± 0.40	12.20	11.40
20	8.13 ± 0.25	8.95	8.45
21	5.02 ¹ ± 0.29	5.20	4.55
22	5.76 ± 0.57	6.15	5.80
23	7.00 ± 0.23	6.05	6.85
24	9.28 ± 0.27	9.25	8.85
25	9.02 ± 0.15	8.80	9.10

Table IV Precision of white cell count estimations expressed as means and 95% confidence intervals by machine compared with mean manual estimates using a haemocytometer and a Coulter model B

¹, ² see Table I.

³ see Table II.

without highly sensitive statistical analysis, drift would not have been detected, and was rarely apparent on casual inspection of results. In Table V, the overall 95% confidence interval for each parameter by manual and automatic methods is presented, giving a compact assessment of the reproducibility of results. In only one sample (number 4) was a 'wild' set of results obtained. This was readily detected by simple inspection of the results.

Blood Test Parameter	95% Confidence Interval	
	Coulter Model S	Manual Methods
Hb (g/100 ml)	12.16-12.76	12.01-12.84
PCV (%)	36.11-37.87	36.70-39.90
RCC (m/c mm)	4.16- 4.35	3.64- 5.10
MCV (cμ)	84.60-87.66	76.53-97.98
MCH (pg)	28.84-30.20	24.27-32.65
MCHC (%)	32.66-34.54	30.39-34.16
WBC (1,000/c mm)	7.18- 9.04	7.03- 9.88

Table V Overall 95% confidence intervals computed over all samples ignoring between-sample variation

Blood Test Parameter	Pooled Within-sample Variances and Standard Deviations						Variance Ratios Technician: Machine
	Machine			Technician			
	Mean	Variance	SD	Mean	Variance	SD	
Hb (g/100 ml)	12.46	0.024	0.15	12.43	0.042	0.20	1.79
PCV (%)	36.99	0.200	0.45	38.02	0.408	0.64	2.04
RCC (m/c mm)	4.25	0.002	0.04	4.39	0.130	0.36	3.10
MCV (c μ)	86.13	0.610	0.78	87.22	26.833	5.18	43.96
MCH (pg)	29.52	0.125	0.35	28.46	4.112	2.03	32.81
MCHC (%)	33.60	0.228	0.48	32.29	0.851	0.92	3.74
WBC (1,000/c mm)	8.11	0.223	0.47	8.44	0.464	0.46	2.01

Table VI Comparison of the reproducibility of machine and manual estimates of seven blood test parameters

THE ACCURACY OF RESULTS

This may be defined as the measure of closeness of observed and 'true' values. The determination of 'true' values for haematological measurements for comparison with those obtained by the

Coulter model S is difficult, as manual techniques are hard to standardize accurately and fraught with potential sources of error. Furthermore, the Coulter model S is itself calibrated using a standard of preserved blood in which the parameters have been determined either manually or by another instrument. Thus, an assessment of accuracy of the Coulter model S is rather a comparison of the Coulter model S with the results of manual estimations done (a) especially for comparison with the instrument and (b) as part of the daily routine work of the Department. The results for (a) are presented in Tables I-VI where the mean manual and mean machine results are compared. The degree of within-sample variation by the model S and by the manual methods is shown in Table VI, which also gives the manual : machine variance ratio.

In Table VII an analysis of the differences between the mean results for the manual and machine estimates of the seven parameters is presented. By the manual technique, the red cell counts are significantly higher than those obtained by the model S and the MCH values are significantly lower. This relationship would be expected in view of the close correlation between manual and machine haemoglobin estimations (see Tables I, VII, and VIII).

No correlation was detected between the Hb concentration, haematocrit, red cell count, MCV, MCH, or MCHC and the variance of these parameters. However, when the white cell count results were examined, they were found to be significantly ($P < 0.05$) less reproducible when the white cell count was high.

In Table VIII the results of comparison of the ordinary routine work and the instrument counts are shown, together with their correlation coefficients. The degree of correlation for the haemoglobin concentration, white cell count, and haematocrit is high, but that for the MCHC is comparatively poor.

CARRYOVER

The mean background count in the red-cell channel for isotonic solution was 0.023×10^6 per c mm with a range of 0 to 0.04×10^6 . The mean of 20 estimates of carryover was 1.96% with a range of 1.1 to 2.7%.

Blood Test Parameter	No. of Positive Differences in 25 Samples	Probability of Occurrence
HB	13	Not significant
PCV	16	Not significant
RCC	19	< 0.05
MCV	15	Not significant
MCH	2	< 0.01
MCHC	8	Not significant
WBC	15	Not significant

Table VII Binomial sign test of significance of differences between machine and technician means

Group	Haemoglobin Concentration (g/100 ml)			Haematocrit (%)		
	Machine Mean	Manual Mean	Correlation Coefficient	Machine Mean	Manual Mean	Correlation Coefficient
I (135) ¹	13.85	14.04	0.971	40.23	41.74	0.853
II (122)	13.93	14.02	0.972	41.31	42.09	0.764
III (129)	13.95	14.19	0.977	41.21	41.70	0.807

Table VIIIa Comparison between routine work performed manually and by Coulter model S

¹Numbers in each group.

Group	Mean Corpuscular Haemoglobin Concentration (%)			White Cell Count (1,000/c mm)		
	Machine Mean	Manual Mean	Correlation Coefficient	Machine Mean	Manual Mean	Correlation Coefficient
I (135) ¹	34.46	33.25	0.437	7.92	8.52	0.923
II (122)	33.95	32.99	0.478	8.59	8.98	0.870
III (129)	33.54	33.15	0.776	8.71	9.41	0.931

Table VIIIb Comparison between routine work performed manually and by Coulter model S

¹Numbers in each group.

Parameter		Sample 1					Sample 2				
		Tube Number ¹					Tube Number ¹				
		1	2	3	4	5	1	2	3	4	5
Hb	Observed	16.9	8.7	6.4	4.3	2.6	—	10.5	8.0	5.4	3.2
	Expected	16.9	8.5	6.4	4.2	2.4	—	10.5	7.9	5.2	3.2
PCV	Observed	50.1	26.6	19.1	12.5	7.7	—	30.2	23.0	15.5	9.5
	Expected	51.1	25.5	19.2	12.8	7.7	—	30.8	23.1	14.4	9.2
RCC	Observed	5.55	2.89	2.11	1.39	0.85	—	3.17	2.40	1.62	1.0
	Expected	5.60	2.90	2.10	1.40	0.85	—	3.13	2.35	1.51	0.99
WCC	Observed	8.1	4.3	3.1	2.2	1.3	—	37.9	29.9	20.1	11.3
	Expected	8.1	4.1	3.1	2.0	1.2	—	37.8	28.3	18.9	11.3
MCV	Observed	91	93	92	92	92	—	95	96	96	95
MCH	Observed	29.8	29.6	29.7	30.7	30.1	—	33.9	32.9	33.4	32.0
MCHC	Observed	33.6	32.8	33.4	34.6	33.6	—	35.0	34.7	35.2	34.1

Table IX The results of tests of linearity on two samples of blood diluted or concentrated by the addition or removal of plasma (see methods)

¹Tube number 1 contains blood concentrated to 133½% of the original by removal of plasma. Tubes 2, 3, 4, and 5 contain blood diluted with plasma to 66½, 50, 33½, and 20% respectively of the original. The observed values for the unaltered blood sample are taken as the 100% 'reference' point.

Fig. 2 Linearity for haemoglobin concentration and haematocrit. The line represents perfect correlation between 'observed' and 'expected' values, and the symbols give the results actually obtained.

LINEARITY

The results of estimates of linearity for two samples are presented in Table IX and Figures 2 and 3 illustrate a typical set of results for a third sample. The results are linear in the ranges tested, ie, Hb 2-24 g/100 ml; haematocrit, 6-70%; red cell count 0.7-8.0 m/c mm; white cell count, 1.0-50,000/cu mm.

As can be seen from Table IX the results for

Fig. 3 Linearity for red cell count and white cell count. The line represents perfect correlation between 'observed' and 'expected' values, and the symbols give the results actually obtained.

MCV, MCH, and MCHC remain constant within the limits of error of the machine. Results obtained from three further samples were similar to those presented.

INSTRUMENT FAILURES

Minor faults such as blocked aperture tubes and drainage tubes are quite frequent and usually easily detected and remedied. Less easily detected

and less frequent are loose electronic circuit cards, but once found this fault is also easily remedied.

More major instrument failures occur occasionally and generally require replacement of parts. While these faults are often easily diagnosed, it is usually necessary to seek assistance from the manufacturer, at least in obtaining the necessary spare parts. We have observed in six months' operation: failure of the blood sampling valve, requiring replacement; failure of the mechanical mechanism in one digital printout system; failure of several electronic circuit cards requiring replacement; looseness of the joints between aperture tubes and the flange on which they are mounted; failure of the lamp in the automatic haemoglobinometer.

In no instance has the model S been out of commission for more than one day, and has usually been repaired within three hours of discovery of the fault.

Discussion

The reproducibility achieved by the Coulter model S in this study is comparable with that found by Barnard, Carter, Crosland-Taylor, and Stewart (1969), also using a Coulter model S, and with that reported by Nelson and his colleagues in their evaluation of the Technicon Auto-Analyzer SMA-4 (Lappin *et al*, 1969), and the SMA-7A (Nelson, 1969a).

It is clear from the variance ratios (see Table VI) that the results using the Coulter model S are much more reproducible than those obtained by manual methods. The only directly measured parameter which showed a significant and consistent difference between manual and machine counts was the red cell count which tended to be higher when counted by hand. Since the haemoglobin estimations were in good agreement by both techniques, a significant difference, in the opposite direction, was seen in the MCH value.

When results obtained during routine daily work were compared with those obtained from the model S, there was good correlation for the haemoglobin concentration. However, the correlation for the haematocrit determination was less good and this may reflect the well known inaccuracies of microhaematocrit estimations. It may also in part reflect the fact that the haematocrit estimation performed by the model S is computed from two other measurements (red cell count and MCV) thus compounding the opportunity for error. The correlation coefficients for the MCHC are rather poor; this, again, may reflect technical inaccuracies in the manual methods and the use of three parameters by the model S (Hb, red cell count, MCV) to calculate the mean corpuscular haemoglobin concentration. The correlation coefficients for the white cell counts indicate good agreement.

The potential range of error in any parameter

except the white cell count is not influenced by the value obtained for that parameter in the ranges tested here. However, when the white cell count is high, the degree of imprecision increases, although not to a degree that is important in practice.

The results provided by the model S are linear within the ranges met within clinical practice for the Hb concentration, haematocrit, red cell count, and white cell count (see Table IX and Figs. 2 and 3). There is a tendency for very low white counts in the 'observed' group to be higher than expected. This may in part be due to a relative increase in the background count compared with the white cell count. In general the background count in the white cell channels has been in the range 0-300/cmm, which is clearly insufficient to be of significance in practice. We have been able to detect total white cell counts of less than 500/c mm on several occasions, confirmed by manual counting. The effects of a high white cell count on other parameters have not been studied (see Barnard, Carter, Crosland-Taylor, and Stewart, 1969). The amount of carryover from one specimen to the next is about 2%, an amount negligible for practical purposes. This is similar to the carryover found for the Coulter model S by Barnard *et al* and for the Technicon SMA-4 (Lappin *et al*, 1969), but greater than the carryover reported by Nelson (1969a) for the SMA-7A.

The rate of operation of the model S is about three times that of the SMA-4. The theoretical maximum rate is 180 samples per hour, but the usual maximum working rate in practice is about 150 samples per hour. One technician working alone can comfortably achieve 120 samples per hour.

The Coulter model S is stable for considerable periods (days or weeks) in the absence of an instrument failure (see below). Recalibration is seldom required more than once weekly. In view of the expense of the standard calibrating reagent, '4C', we use samples with high and low haemoglobin concentrations and high and low white cell counts each morning, taken from the previous day's work, to check that the instrument does not require recalibration. In addition a large sample of blood (about 30 ml) is taken daily and counted at arbitrary intervals during the day to check the stability of the calibration and for the absence of drift. Very high and very low (but feasible) values on any parameter are confirmed by recounting. Discrepancies have seldom been encountered and were probably a consequence of inadequate mixing, and an automatic rocking mixer is now used. Routine inspection of blood smears helps to prevent the reporting of inaccurate white cell counts and red cell indices. Only one gross error was detected in this study (see sample 4, Tables I-IV). The commonest cause of detectable error is now misidentification of a sample at the time of withdrawal. Obviously

incorrect results (resulting perhaps from electronic or printer failure) are occasionally observed.

The Coulter model S, as presently marketed, requires to be operated manually. While this has the disadvantage of using staff, there are advantages in that a batch of non-urgent blood counts can immediately be set aside to perform emergency work on a 'while-you-wait' basis. Furthermore, the oscilloscope screens, counting chambers, diluting chambers, exhaust chambers, and other working parts are under constant skilled surveillance, allowing prompt detection of defects.

Instrument failures are essentially of two types: minor and readily remediable by trained laboratory staff, and major, requiring highly skilled assistance or spare parts. In general, repairs have been carried out within two or three hours of discovery of a major fault and within minutes for most minor faults; the model S has not been out of use for more than 24 hours on any occasion. As a precaution, the equipment required to perform an entire day's work by hand is kept in readiness in case of a major failure. Proximity to a Coulter service facility is obviously an advantage. The importance of careful attention to cleaning and maintenance of the machine is stressed.

It is clear that the Coulter model S is at least comparable with other automated methods in respect of accuracy and reproducibility and very much more accurate and reproducible than manual methods (see Table VI). It is faster than any other instrument so far developed and offers the opportunity for already overworked laboratories to handle quickly and accurately their increasing workload and, at the same time, permit redeployment of skilled staff to perform other, less easily automated, tasks. The broader implications of automation in the clinical pathology laboratory have been discussed more fully elsewhere (Nelson, 1969b).

References

- Barnard, D. F., Carter, A. B., Crosland-Taylor, P. J., and Stewart, J. W. (1969). An evaluation of the Coulter model S. *J. clin. Path.*, **22**, suppl. (Coll. Path.), 3, 26-33.
- Bradley, J. V. (1968). *Distribution-Free Statistical Tests*. Prentice Hall, Englewood Cliffs, New Jersey.
- Brecher, G., Schneiderman, M., and Williams, G. Z. (1958). Evaluation of electronic red blood cell counter. *Amer. J. clin. Path.*, **26**, 1439-1449.
- Dacie, J. V., and Lewis, S. M. (1963). *Practical Haematology*, 3rd ed. Churchill, London.
- Hart, B. I. (1942). *Ann. math. Statist.*, **13**, 445.
- Lappin, T. R. J., Lamont, A., and Nelson, M. G. (1969). An evaluation of the AutoAnalyzer SMA-4. *J. clin. Path.*, **22**, 11-18.
- Mood, A. M., and Graybill, F. A. (1963). *Introduction to the Theory of Statistics*, 2nd ed. McGraw-Hill, New York.
- Nelson, M. G. (1969a). Multichannel continuous flow analysis on the SMA-4/-7A. *J. clin. Path.*, **22**, suppl. (Coll. Path.), 20-25.
- Nelson, M. G. (1969b). Automation in the laboratory. *J. clin. Path.*, **22**, 1-10.
- Scheffé, H. (1959). *The Analysis of Variance*. Wiley, New York.