Technical methods

6 COST OF REAGENTS

Sephadex

Allowing some 6% for wastage, each test would cost approximately 1.5 pence (0.625 N.P.).

Liothyronine

Allowing 500-600 tests (out of a possible 13,00-1,600 tests) for the 200 μ Ci, cost per test would be 3p.

Buffer

Cost per test 0.01 pence (0.0042 N.P.).

Total cost per test is 8.71 pence (3.63 N.P.) which compares favourably with the resin sponge method costing 135 pence (56.3 N.P.) per test in kit form.

7 ACCURACY OF THE METHOD

One hundred and eight consecutive sera were tested in duplicate and the standard error of this method was found to be 2.221%.

8 AGREEMENT WITH RESIN SPONGE METHOD (TRIOSORB-125)

Figure 3 shows a comparison of results obtained by the Sephadex and resin-sponge methods. From this the normal limits of the Sephadex uptake ratio are 0.85 to 1.34 compared with 0.86 to 1.20 for the resin-sponge method.

Fig. 3 The results of duplicate estimations done by the Sephadex method compared with single estimations on the same sera done by the resin method. 'Uptake ratio' indicates the uptake of the unknown divided by the uptake of normal serum.

Demonstration of sensitized lymphocytes in blood

E. J. FIELD and E. A. CASPARY *From the Medical Research Council, Demyelinating Diseases Unit, Newcastle General Hospital, Newcastle upon Tyne*

Under certain circumstances, both physiological and pathological, it becomes of interest to determine the occurrence of sensitized lymphocytes in blood. The macrophage migration inhibition test has been adapted for this purpose (Soborg and Bendixen, 1967) but presents certain technical features which make it, in the hands of some workers (Kaltreider, Soghor, Taylor, and Decker, 1969), unreliable and open to error. (This has been our own experience to date (Hughes, unpublished data).) A simple, reliable and reproducible method has now been developed both for human and animal work and offers quantitative results suitable for statistical analysis. In principle it depends upon the interaction of sensitized lymphocytes with their sensitizing antigen to produce a factor which causes inhibition of macrophage migration and forms the basis of the macrophage migration inhibition test (Bloom and Bennett, 1966). The same or a similar factor also reduces electrophoretic mobility of macrophages (Sundaram, Phondke, and Ambrose, 1967; Hughes, Caspary, and Field, 1970)—an action which can be blocked by antilymphocytic serum (Field, Hughes, and Caspary, 1969; Caspary, Hughes, and Field, 1970). In the reduction of macrophage mobility by specific antigen (Sundaram *et al*, 1967; Diengdoh and Turk, 1968; Field *et al*, 1969; Hughes *et al*, 1970) the latter interacts with those lymphocytes (10-15%) present in the peritoneal exudate of the sensitized animal. In the method described here, antigen is allowed to

Received for publication 2 July 1971

Simple, inexpensive micro method for the estimation of T_3 uptake—continued

I am grateful to Dr W. G. Dangerfield of the North Middlesex Hospital and to Dr P. Marshall-Jones of Whipps Cross Hospital for some useful suggestions.

References

- Cavalieri, R. R., Castle, J. N., and Searle, G. L. (1969). A simplified method for estimating free-thyroxine fraction in serum. *J. nucl. Med.*, 10, 565-570.
- Clark, F. (1967). The estimation of thyroid hormone binding by plasma proteins and of unbound levels of thyroxine in plasma. In *Symposium on the thyroid gland*, 1966 (Suppl., *J. clin. Path.*, 20, 344-351).

react with the blood lymphocytes under test and the effect upon guinea pig macrophage electrophoretic mobility measured, ie, the guinea pig macrophages are used purely as an indicator system of lymphocyte-antigen interaction. Preliminary experiment showed, as anticipated, that simple mixture of blood lymphocytes with guinea pig peritoneal exudate resulted in a mixed lymphocyte reaction. In order to eliminate this the guinea pig peritoneal exudate (with its 10-20% lymphocyte content) was irradiated. Trial showed that 100 rads (Cobalt 60 γ rays, at 56 cm from a 4,000 Curie source, dose rate 100r in 54 sec) was adequate to incapacitate peritoneal lymphocytes from taking part in a mixed reaction although it did not interfere with their viability as measured by a dye exclusion test. Macrophage electrophoretic mobility was unaffected by this treatment. Irradiated guinea pig exudate was therefore usable as an indicator system for the detection of macrophage migration slowing factor resulting from interaction of blood lymphocytes and antigen to which they might be sensitized.

The exudate of guinea pig peritoneal macrophages was elicited by intraperitoneal injection of sterile liquid paraffin and washing out the peritoneal cavity with Hanks balanced salt solution containing 5 units of heparin/ml some eight to 12 days later. This suspension was washed twice (spinning at 250 *g* for 10 minutes) in heparinized balanced salt solution and finally suspended in TC 199 (without heparin) the volume being adjusted to yield 10^7 cells/ml. Usually $80-200 \times 10^6$ macrophages were obtained from a guinea pig of 400-600 g.

Blood lymphocytes were obtained by venepuncture (or cardiac puncture where appropriate in animals). Twenty ml blood was defibrinated with glass beads, followed by removal of polymorphs with saccharated iron as used by Hughes and Caspary (1970) in their modification of the method of Coulson and Chalmers (1964). Generally there was a yield of about 10^6 lymphocytes/ml blood.

In carrying out a test 0.5×10^6 blood lymphocytes (in 0.5 ml TC 199) were added to 1.0 ml of TC 199 containing 10^7 irradiated guinea pig macrophages and 1.5 ml TC 199 added to bring the total volume to 3 ml. To this the antigen to be tested was added in 0.1 ml of saline. The actual amount of antigen may have to be determined, but in our own experiments we have used 100 μ g of encephalitogenic factor (Caspary and Field, 1970), or special aliquots of thyroid derived antigens (Field, Caspary, Hall, and Clark, 1970). This mixture was allowed to stand at laboratory temperature (18-20°C) for 90 minutes. Controls were set up with antigens not suspected, eg, kidney extract, zoster antigen. Electrophoretic mobility was measured in a Zeiss cytopherometer.

All measurements were made in TC 199 medium and experience has shown that different batches of this medium might give slightly different actual migration speeds. However, differences between controls and positives were preserved exactly. It should be added that some experience with the cytopherometer is needed before consistent results can be expected and that it is useless to begin a set of observations unless the apparatus is 'behaving itself'. In our experience common sources of trouble are the draining stop-cock or microleaks at the partitioning membranes. A cell was timed in both directions of the potential difference, care being taken that only macrophages (easily identified under phase contrast by their size and liquid paraffin content) are studied. Ten different cells were used for each specimen so that the mean of 20 readings could be computed. The ambient temperature of the electrophoretic cell was maintained at $23 \pm 0.05^\circ\text{C}$. It has been our practice to number specimens and make all measurements 'blind' so that possible bias in observation is eliminated.

Migration speed in the presence of antigen is compared with that in its absence, ie, in the control mixture of irradiated macrophages and lymphocytes under test, and may be expressed as a percentage change. Results have been clear cut in neurological (Caspary and Field, 1970) and thyroid (Field *et al.*, 1970) disease, but the method is applicable to a very wide spectrum of physiological and pathological conditions in which sensitization of circulating lymphocytes to possible antigens is suspected or is to be studied. Such conditions range from autoimmune disease to sarcoidosis (where delayed sensitivity anergy would be of special interest), rheumatoid arthritis, sensitization to viral antigens, sensitization to placental tissue in unexplained repeated abortion, systemic lupus erythematosus, and many other conditions. In addition we have shown that the method can be used as a rapid means of LATS assay in serum; as an *in vitro* method of titrating antilymphocytic serum (since ALS has a blocking action in the test) and as a possible means of controlling the effects of therapy in multiple sclerosis or in graft rejection.

References

- Bloom, B. R., and Bennett, B. (1966). Mechanism of a reaction *in vitro* associated with delayed-type hypersensitivity. *Science*, 153, 80-82.
- Caspary, E. A., and Field, E. J. (1970). Sensitization of blood lymphocytes to possible antigens in neurological disease. *Europ. Neurol.* In the Press.
- Caspary, E. A., Hughes, D., and Field, E. J. (1970). On the mode of action of antilymphocytic serum: experiments on electrophoretic mobility of macrophages in experimental allergic encephalomyelitis. *Clin. exp. Immunol.*, 7, 395-400.
- Coulson, A. S., and Chalmers, D. G. (1964). Separation of viable lymphocytes from human blood. *Lancet*, 1, 468-469.

Technical methods

- Diengdoh, J. V., and Turk, J. L. (1968). Electrophoretic mobility of guinea pig peritoneal exudate cells in hypersensitivity reactions. *Int. Arch. Allergy*, **34**, 297-302.
- Field, E. J., Caspary, E. A., Hall, R., and Clark, F. (1970). Circulating sensitized lymphocytes in Graves' disease. Observations on its pathogenesis. *Lancet*, **1**, 1144-1147.
- Field, E. J., Hughes, D., and Caspary, E. A. (1969). Mode of action of antilymphocyte serum. *Lancet*, **2**, 964.
- Hughes, D., and Caspary, E. A. (1970). Lymphocyte transformation *in vitro* measured by tritiated thymidine uptake. I. Lymphocyte culture techniques. *Int. Arch. Allergy*, **37**, 506-531.
- Hughes, D., Caspary, E. A., and Field, E. J. (1970). Delayed hypersensitivity studies in experimental allergic encephalomyelitis; a comparison of macrophage migration inhibition test and electrophoretic mobility. *Z. Immunitätsforsch.* In the Press.
- Kaltreider, H. B., Soghor, D., Taylor, J. B., and Decker, J. L. (1969). Capillary tube migration for detection of human delayed hypersensitivity: difficulties encountered with "buffy coat" cells and tuberculin antigen. *J. Immunol.*, **103**, 179-184.
- Søborg, M., and Bendixen, G. (1967). Human lymphocyte migration as a parameter of hypersensitivity. *Acta med. scand.*, **181**, 247-256.
- Sundaram, K., Phondke, G. P., and Ambrose, E. J. (1967). Electrophoretic mobilities of antigen-stimulated lymph node cells. *Immunology*, **12**, 21-26.

The February 1971 Issue

THE FEBRUARY 1971 ISSUE CONTAINS THE FOLLOWING PAPERS

The Journal in the 1970s

The Australia (hepatitis-associated) antigen in fibrinogen and other fractions of human plasma
A. J. ZUCKERMAN, PATRICIA E. TAYLOR, R. G. BIRD, AND SUSAN M. RUSSELL

Comparison of three serological methods for the detection of hepatitis-associated antigen G. B. BRUCE WHITE, RAGAA M. LASHEEN, MARGERY B. BAILLIE, AND G. C. TURNER

Immunofluorescent staining of rat gastric parietal cells by human antibody unrelated to pernicious anaemia H. K. MULLER, A. R. MCGIVEN, AND R. C. NAIRN

Activities of various cobalamins for *Euglena gracilis* with reference to vitamin B₁₂ assay with *Euglena*
J. F. ADAMS AND FIONA MCEWAN

Use of the 'Analmatic clinical system' in the microbiological assay of vitamin B₁₂ and folic acid in serum GILLIAN COLVIN, G. L. GIBSON, AND D. W. NEILL

Coagulation studies in haemophilic patients taking oral contraceptives A. L. BLOOM, PETER BECK, J. C. GIDDINGS, AND P. M. SWEETNAM

Congenital nephrotic syndrome and renal vein thrombosis in infancy F. ALEXANDER AND W. A. B. CAMPBELL

Filter paper blood samples used in serological studies of respiratory virus infections in children S. M. K. CHISHTY

Changes in the drug resistance of *Staphylococcus aureus* in a non-hospital population during a 20-year period D. J. GOLDIE, V. G. ALDER, AND W. A. GILLESPIE

Cloxacillin-resistant staphylococci in a children's hospital M. B. SKIRROW AND K. B. ROGERS

Isolation of fastidious mycoplasma from human sources ELLI JANSSON

Renal dysplasia R. A. RISDON

Part I. A clinico-pathological study of 76 cases

Part II. A necropsy study of 41 cases

Bronchiolitis obliterans, bronchiectasis, and other sequelae of adenovirus type 21 infection in young children D. M. O. BECROFT

Microscope slide electrophoresis of serum lipoproteins in agarose gel M. C. ELPHICK

Technical methods

Crossover electrophoresis for the rapid detection of serum hepatitis (Australia) antigen and antibody C. VERGANI

An economical way of measuring fibrin degradation products CONSTANCE R. SPITTLE

Adapting the glucose channel on the SMA 6/60 for creatinine determinations in blood and urine MAGDALENA MOLL AND BARBARA A. PRING

Present-day practice

Principles of protein estimation by radial immunodiffusion. G. H. GRANT

Letters to the Editor

Book reviews/Notices

Copies are still available and may be obtained from the PUBLISHING MANAGER,

BRITISH MEDICAL ASSOCIATION, TAVISTOCK SQUARE, WC1H 9JR price £1.05