Lability of human creatine kinase isoenzymes at 37°C: a complication of electrophoretic separation

D. A. NEALON AND A. R. HENDERSON

From the Department of Clinical Biochemistry, University Hospital (University of Western Ontario), London, Ontario N6G 2K3, Canada

SYNOPSIS The activity of the brain specific isoenzyme of creatine kinase is shown to fall off rapidly at 37°C, particularly in the presence of albumin. Dithiothreitol cannot reverse this lability. The implications of this finding suggest that electrophoretic techniques which use incubation methods to detect the brain specific isoenzyme of creatine kinase may underestimate the true activity.

We (Nealon and Henderson, 1975a) have shown that the use of DEAE-cellulose is superior to DEAE-Sephadex (Mercer, 1974) for the separation, in a mini-column, of the human creatine kinase (ATP: creatine phosphotransferase, EC 2.7.3.2, CK) isoenzymes. Subsequently we showed (Nealon and Henderson, 1975b) that CK₁¹ could be detected in the sera of patients who had undergone brain surgery. A survey of the literature dealing with the elevation of serum creatine kinase following the acute onset of serious neurological conditions such as head injury, cerebrovascular accidents, encephalitis, etc. shows that CK₁ has not been detected in the serum of these patients.

CK₁ has been found in the serum of persons suffering from the rare condition of malignant hyperpyrexia (Zsigmond *et al.*, 1972) although to date this finding has never been substantiated by other workers (Britt, 1974). Application of our technique to the sera of known cases of malignant hyperpyrexia established the presence, in appreciable quantities, of the brain specific isoenzyme (Nealon *et al.*, 1975). It occurred to us that the general failure to demonstrate the presence of CK₁ in sera of cases of malignant hyperpyrexia or acute neurological conditions could be due to the difference in technique between ourselves and other workers. Most workers had used an electrophoretic technique

for separation of the isoenzymes with subsequent incubation at 37°C with the CK assay media for periods of up to two hours whereas we have used a chromatographic technique. We have therefore investigated the heat stability of the three CK isoenzymes at 37°C.

Materials and Methods

ENZYME ASSAY

Creatine kinase was determined at 37°C by the Rosalki-Oliver method (Rosalki, 1967) using the CPK Fast-Pak (Calbiochem, San Diego, Calif. 92112). Assays were performed on a Unicam SP 1700 spectrophotometer (Canadian Laboratory Supplies Ltd, Toronto, Ontario, Canada). Normal serum activity, in this laboratory, extends up to 135 U/litre.

ISOENZYME PREPARATIONS

Human brain, heart, and skeletal muscle were used as sources of CK₁, CK₂, and CK₃ respectively. All tissues were homogenized in 100 mmol/litre KCl (1 g wet weight in 2 ml) and clarified by centrifugation at 1300 g for 30 min at 10°C.

The separation and partial purification of the isoenzymes was accomplished by the method previously described (Nealon and Henderson, 1975b). Solutions of each isoenzyme were prepared for analysis as follows:

(a) CK₃ (final activity 216 U/litre) and CK₂ (final activity 197 U/litre) were prepared in 50 mmol/litre Tris-HCl buffer, pH 7.0 (25°C), with and without 7.5 g/litre bovine serum albumin (BSA).

¹The CK isoenzymes are named on the basis of their electrophoretic mobility towards the anode. Thus the isoenzymes BB, MB, and MM of creatine kinase are named CK₁, CK₂, and CK₃ respectively (Commission on Biochemical Nomenclature, 1972).

Received for publication 14 April 1975.

(b) CK₁ (final activity 223 U/litre) was prepared in the same buffer described above, with and without BSA, in the following concentrations—1.5, 2.5, 4.0, 5.0, 6.5, 7.5, and 15 g/litre.

Results

Figure 1 illustrates the activity of the CK isoenzymes as a function of incubation time at 37°C. In the case of CK₃ and CK₂, albumin (BSA at a concentration of 7.5 g/litre) has very little effect on the activity of the isoenzymes over a two-hour period so that with or without albumin there is a loss of less than 20% of total activity. On the other hand, CK₁ is shown to be extremely unstable at 37°C, particularly in the presence of albumin. For clarity, only the results of the experiments with CK₁ in the

presence of 7.5 and 15 g/litre BSA are shown. Decreasing the protein concentration, as described in Materials and Methods, produced results which approached those obtained in the absence of protein.

Figure 2 illustrates that dithiothreitol in the presence, or absence, of BSA exerts a stabilizing effect on the isoenzymes. With each of these isoenzymes, however, there is an increase in lability in the presence of BSA. CK₁ is clearly stabilized by dithiothreitol (compare with fig 1) but the presence of BSA (at a concentration of 15 g/litre) makes the isoenzyme extremely heat labile. Again the extent of the heat lability is directly proportional to concentrations of BSA (compare with fig 1).

Experiments with purified human albumin produced exactly the same results, therefore the observed effects are albumin specific but not species specific.

Fig 1 Effect of incubation at 37°C on human creatine kinase isoenzymes with or without added bovine serum albumin.

Each isoenzyme was incubated at 37°C and sampled, for enzyme assay, at the time shown. The isoenzymes were either incubated without BSA (▲—▲) or with BSA at a concentration of 7.5 g/litre (△—△) or 15 g/litre (○—○). The other BSA concentrations used in the experiment are omitted for the sake of clarity.

Fig 2 Effect of incubation at 37°C on human creatine kinase isoenzymes in the presence of dithiothreitol with or without added bovine serum albumin.

Each isoenzyme was incubated at 37°C and sampled, for enzyme assay, at the times shown. The isoenzymes were either incubated with dithiothreitol alone (■—■) or with added BSA at a concentration of 15 g/litre (□—□). The other BSA concentrations used in the experiment are omitted for the sake of clarity.

Discussion

The results clearly show that CK₁, in the presence of albumin, in concentrations much less than those present in normal serum, is inactivated to such an extent that within 15 min at 37°C only about 20% of its original activity remains (fig 1). Even the presence of dithiothreitol, which is a known activator of creatine kinase (Bishop *et al*, 1971; Warren, 1972), has little influence on this heat lability (fig 2).

The lability of CK₁ is clearly related to the presence of albumin and it is known that serum albumin migrates with, or close to, CK₁ on electrophoresis (Wood, 1963; Trainer and Gruenig, 1968; Smith, 1972). It therefore seems very likely that unless the isoenzymes are eluted from the electrophoretic media immediately after separation for reaction rate assay (Roberts *et al*, 1974) any incubation of the electrophoresis medium is likely to destroy CK₁. Examination of the many current electrophoresis techniques indicates that incubation of the separated CK isoenzymes with the CK assay medium ranges over the following: 25 min at 37°C (Klein *et al*, 1973), 30 min at 37°C (Wolf *et al*, 1974; Goto *et al*, 1969), 1 h at 30°C (Roe *et al*, 1972), 1 h at 37°C (Zsigmond *et al*, 1972), 30 min-2 h at 37°C (Smith, 1972), and 2 h at 37°C (Somer and Konttinen, 1972b; Goto, 1974).

Therefore low levels of serum CK₁ could very well be missed by an electrophoretic technique, especially as it has been suggested that this technique will not detect less than about 2 U/litre (at 37°C) of CK activity (Yasminieh and Hanson, 1975). We suggest that the only satisfactory technique for detecting CK₁ is the column chromatographic method already referred to.

D. A. Nealon is supported by a Medical Research Council (Canada) Studentship which is gratefully acknowledged.

References

- Bishop, C., Chu, T. M., and Shihabi, Z. K. (1971). Single stable reagent for creatine kinase assay. *Clin. Chem.*, **17**, 548-550.
- Britt, B. A. (1974). Malignant hyperthermia: diagnosis and treatment. *Contemp. Surg.*, **4**, 97-104.
- Commission on Biochemical Nomenclature (1972). *Enzyme Nomenclature*, pp. 24-25. Elsevier, Amsterdam.
- Goto, I. (1974). Creatine phosphokinase isoenzymes in neuromuscular disorders. *Arch. Neurol. (Chic.)*, **31**, 116-119.
- Goto, I., Nagamine, M., and Katsuki, S. (1969). Creatine phosphokinase isoenzymes in muscles. *Arch. Neurol. (Chic.)*, **20**, 422-429.
- Klein, M. S., Shell, W. E., and Sobel, B. E. (1973). Serum creatine phosphokinase (CPK) isoenzymes after intramuscular injections, surgery, and myocardial infarction. *Cardiovasc. Res.*, **7**, 412-418.
- Mercer, D. W. (1974). Separation of tissue and serum creatine kinase isoenzymes by ion-exchange column chromatography. *Clin. Chem.*, **20**, 36-40.
- Nealon, D. A., Britt, B. A., and Henderson, A. R. (1975). (In preparation).
- Nealon, D. A. and Henderson, A. R. (1975a). Separation of creatine kinase isoenzymes in serum by ion-exchange chromatography (Mercer's method, modified to increase sensitivity). *Clin. Chem.*, **21**, 392-397.
- Nealon, D. A. and Henderson, A. R. (1975b). Quantification of brain-specific creatine kinase isoenzyme in serum. *Clin. Chem.*, **21** (In press).
- Roberts, R., Henry, P. D., Witteveen, S. A. G. J., and Sobel, B. E. (1974). Quantification of serum creatine phosphokinase isoenzyme activity. *Amer. J. Cardiol.*, **33**, 650-654.
- Roe, C. R., Limbird, L. E., Wagner, G. S. and Nerenberg, S. T. (1972). Combined isoenzyme analysis in the diagnosis of myocardial injury: application of electrophoretic methods for the detection and quantitation of the creatine phosphokinase MB isoenzyme. *J. Lab. clin. Med.*, **80**, 577-590.
- Rosalki, S. B. (1967). An improved procedure for serum creatine phosphokinase determination. *J. Lab. clin. Med.*, **69**, 696-705.
- Smith, A. F. (1972). Separation of tissue and serum creatine kinase isoenzymes on polyacrylamide gel slabs. *Clin. chim. Acta*, **39**, 351-359.
- Somer, H. and Konttinen, A. (1972a). A method allowing the quantitation of serum creatine kinase isoenzymes. *Clin. chim. Acta*, **36**, 531-536.
- Somer, H. and Konttinen, A. (1972b). Demonstration of serum creatine kinase isoenzymes by fluorescence technique. *Clin. chim. Acta*, **40**, 133-138.
- Trainer, T. D. and Gruenig, D. (1968). A rapid method for the analysis of creatine phosphokinase isoenzymes. *Clin. chim. Acta*, **21**, 151-154.
- Warren, W. A. (1972). Activation of serum creatine kinase by dithiothreitol. *Clin. Chem.*, **18**, 473-475.
- Wood, T. (1963). Adenosine triphosphate-creatine phosphotransferase from ox brain: purification and isolation. *Biochem. J.*, **87**, 453-462.
- Wolf, P. L., Kearns, T., Neuhoft, J., and Lauridson, J. (1974). Identification of CPK isoenzyme MB in myocardial infarction. *Lab. Med.*, **5**, 48-50.
- Yasminieh, W. G. and Hanson, N. Q. (1975). Electrophoresis on cellulose acetate and chromatography on DEAE-Sephadex A-50 compared in the estimation of creatine kinase isoenzymes. *Clin. Chem.*, **21**, 381-386.
- Zsigmond, E. K., Starkweather, W. H., Duboff, G. S., and Flynn, K. (1972). Elevated serum-creatine phosphokinase activity in a family with malignant hyperpyrexia. *Anesth. Analg. Curr. Res.*, **51**, 220-225.