An antibiotic susceptibility testing trial organised as part of the United Kingdom national external microbiological quality assessment scheme

JJS SNELL, DFJ BROWN,* PS GARDNER

*From the Division of Microbiological Reagents and Quality Control, Central Public Health Laboratory, Colindale, NW9 5HT, and the *Regional Public Health Laboratory, Level 6 Addenbrooke's Hospital, Hills Road, Cambridge CB2 2QW*

SUMMARY Organisms of known susceptibility to antimicrobial drugs were distributed for sensitivity testing to laboratories participating in the United Kingdom National External Microbiological Quality Assessment Scheme. The results obtained were correlated with the methods used. Laboratories differed in their standards of antimicrobial drug sensitivity testing. An association between error rates and particular methods and practices enabled recommendations to be made on disc content, method of methicillin testing, preparation of inoculum, use of controls and use of lysed blood for sulphonamide testing. Some media appeared significantly better than others but because of the many factors involved further information is being sought to clarify this.

The United Kingdom National External Quality Assessment Scheme has been described previously.¹ Its main emphasis has been on the supply of simulated clinical material for proficiency testing. In the bacteriology section, simulated specimens designed to test isolation and identification procedures have been sent to participants at approximately monthly intervals. During 1974–1980 participants were asked to perform antimicrobial drug sensitivity tests on one of the pathogens included in the simulated specimens. The strains were not specially selected for their susceptibility patterns and participants were unrestricted in their choice of drugs to test. The results obtained indicated a high overall discrepancy rate, particularly pronounced with some strain/drug combinations, confirming the earlier findings of Stokes and Whitby.² In order to investigate the causes of errors in results reported for these strains, a trial scheme was designed and implemented during September 1980–May 1981.

Material and methods

LABORATORIES PARTICIPATING

All participants enrolled in the UK National External Microbiology Quality Assessment Scheme and accepting bacteriology specimens were included in the trials (450 laboratories in September, 1980). The

geographical and functional distribution of participants has been described previously.¹

ORGANISMS

The organisms distributed and the recommended antibiotics to be tested are shown in Table 1. Before despatch, the drug susceptibility characteristics of the strains were determined in the Division of Microbiological Reagents and Quality Control (DMRQ) and the Regional Public Health Laboratory, Cambridge. Minimum inhibitory concentrations were determined by serial dilution of the drug in solid medium based on the method of Ericsson and Sherris³ but with Oxoid Isosensitest agar and with methicillin testing at 30°C. These reference laboratory results are shown in Table 1. Strains were sent to participants at approximately monthly intervals during September 1980–May 1981.

QUESTIONNAIRE

A questionnaire requesting details of antimicrobial susceptibility testing methods was sent to all participants at the beginning of the trial. Subsequently, the report form for each specimen contained a section requesting details of any changes in methods since completion of the questionnaire. When changes were notified, the details of the method before and after modification for the laboratory concerned were excluded from the analysis.

Table 1 Reference laboratories' and participants' results for the strains distributed

	MICs determined by reference labs (mg/l)	Designated correct result	Number of laboratories reporting strain as			% of labs correct
			S	I	R	
<i>E. coli</i> MQCL 60 (derived from NCTC 10418, the recommended control for sensitivity tests). Site: urine. Combined results from 2 distributions						
Ampicillin	2/4	S	815	24	28	94.0
Cephaloridine	4/4	S	674	46	25	90.5
Gentamicin	0.125/0.5	S	836	1	2	99.6
Sulphonamide	8/2	S	831	15	52	92.5
Trimethoprim	0.125/0.125	S	799	5	9	98.3
Cotrimoxazole	NT/NT	S*	744	2	16	97.6
<i>Proteus mirabilis</i> MQCL 11. Site: urine. Combined results from 2 distributions						
Ampicillin	2/2	S	844	3	21	97.2
Cephaloridine	8/8	S	562	96	93	74.8
Gentamicin	0.25/1	S	867	3	1	99.5
Sulphonamide	4/2	S	734	28	104	84.8
Trimethoprim	1/2	S†	563	109	147	68.7
Cotrimoxazole	NT/NT	S*	672	25	75	87.0
<i>Klebsiella aerogenes</i> MQCL 314. Site: urine						
Ampicillin	>512/>128	R	1	1	431	99.5
Cephaloridine	>128/128	R	58	16	293	79.8
Gentamicin	0.06/0.25	S	430	1	2	99.3
Sulphonamide	16/4	S	343	34	53	79.8
Trimethoprim	>128/>128	R	6	1	401	98.3
Cotrimoxazole	NT/NT	ND‡	97	59	209	—
<i>Pseudomonas aeruginosa</i> MQCL 356. Site: other than urine						
Carbenicillin	32/32	S	371	23	22	89.2
Gentamicin	0.5/0.5	S	432	1	0	99.8
Tobramycin	0.25/0.25	S	397	1	0	99.7
Amikacin	2/2	S	390	2	0	99.5
<i>Pseudomonas aeruginosa</i> MQCL 8. Site: other than urine						
Carbenicillin	64/64	S	265	81	70	63.7
Gentamicin	0.125/0.25	S	426	5	1	98.6
Tobramycin	0.125/0.125	S	392	2	0	99.4
Amikacin	0.5/0.5	S	365	1	0	99.7
<i>Staphylococcus aureus</i> MQCL 360. Site: other than urine						
Penicillin	0.06/0.06	S	434	0	9	97.9
Methicillin	2/2	S	424	1	3	99.0
Tetracycline	0.5/0.25	S	420	13	3	96.3
Erythromycin	0.25/0.25	S	438	3	1	99.0
Gentamicin	0.125/0.125	S	440	1	0	99.7
Fusidic acid	0.25/0.06	S	430	3	2	98.8
<i>Staphylococcus aureus</i> MQCL 192. Site: other than urine. Combined results from 2 distributions						
Penicillin	16/8	R	0	1	875	99.9
Methicillin	32/16	R	106	28	723	84.3
Tetracycline	64/64	R	6	11	847	98.0
Erythromycin§	2/4	R	23	90	759	87.0
Gentamicin	0.06/0.2	S	861	2	7	98.9
Fusidic acid	0.06/0.03	S	851	4	7	98.7

* Participants were not scored on their results for cotrimoxazole because of the difficulties of interpreting the various testing and reporting conventions used for this combination.

† Participants were not scored on their results for trimethoprim for this strain because although the strain was sensitive to trimethoprim for the species, it was more resistant than many coliforms. For the purpose of the analysis the strain was regarded as sensitive.

‡ A correct result was not designated for cotrimoxazole with the strain because it was sensitive to sulphonamide but resistant to trimethoprim.

§ This strain showed dissociated resistance to erythromycin.

S = sensitive; I = intermediate; R = resistant.

NT = not tested.

SCORING SCHEME

For each strain/drug combination a correct result was designated on the basis of the reference laboratory MIC results (Table 1). The results from each laboratory were assessed at the MQCL after each distribution and every laboratory was given a score for each result as follows: if the participant's result was the same as the designated correct result, a score of 2 was given; other results (including "intermediate") were given a score of 0; where a

correct result was not designated the participant's results were not scored—for example, cotrimoxazole due to the difficulties of interpreting the various testing and reporting conventions used for this combination. After each distribution participants were provided with a computer print-out showing details of their results and scores for the current specimen, their cumulative score for the previous specimens reported and the maximum possible score for these strain/drug combinations. Participants were

No of laboratories achieving various percentages correct of their total possible scores

also provided with the combined results of all laboratories reporting results of sensitive, intermediate or resistant for each strain/drug combination.

Results

DISCREPANCY RATES

Participants' results and error rates for the seven strains are shown in Table 1. The numbers of

laboratories achieving various percentages of their total possible scores are shown in the Figure. It is apparent that standards of performance in antimicrobial sensitivity testing vary considerably with 20% of laboratories achieving less than 90% correct. To ascertain whether performance in sensitivity testing was correlated to performance in examination of specimens for general bacteriology, the coefficient of correlation between the percentage of the total possible score for the sensitivity tests and

Table 2 Distribution of incorrect results according to media used (combined results for all specimens)

Medium used (manufacturer)	No of labs	All laboratories			Laboratories using the Stokes method			Laboratories using control method other than Stokes's		
		No of results Right	No of results Wrong	Ratio of right:wrong	No of results Right	No of results Wrong	Ratio of right:wrong	No of results Right	No of results Wrong	Ratio of right:wrong
DST (Oxoid and Mast Laboratories)	200	9391	722	13	4829	328	15	3011	229	13
Sensitest (Oxoid)	60	2922	197	15	1130	79	14	1186	87	14
Wellcotest (Wellcome Reagents)	1	45	3	15	45	3	15	0	0	—
SAF (Mast)	9	392	41	10	291	36	8	101	5	20
Isosensitest (Oxoid)	50	2456	130	19	1300	61	21	772	52	15
Mueller-Hinton (various)	34	1508	158	10	239	42	6	953	83	11
STA (London Analytical and Bacteriological Media)	12	616	27	23	353	18	20	263	9	29

All laboratories, significant differences in error rates

DST < Mueller-Hinton (χ^2 11.4, $p < 0.001$)

Sensitest < SAF (χ^2 6.0, $p < 0.02$)

and Mueller-Hinton (χ^2 15.9, $p < 0.001$)

Isosensitest < DST (χ^2 14.7, $p < 0.001$),

Sensitest (χ^2 4.3, $p < 0.05$), SAF (χ^2 13.7, $p < 0.001$)

and Mueller-Hinton (χ^2 31.9, $p < 0.001$)

STA < DST (χ^2 8.1, $p < 0.01$),

and Sensitest (χ^2 4.3, $p < 0.05$), SAF (χ^2 12.1, $p < 0.01$)

and Mueller-Hinton (χ^2 17.6, $p < 0.001$)

Laboratories using control method other than Stokes', significant differences in error rates

STA < DST (χ^2 5.6, $p < 0.05$), Sensitest (χ^2 4.7, $p < 0.05$), and Mueller-Hinton (χ^2 7.2, $p < 0.01$)

DST = Diagnostic sensitivity test agar

SAF = Sulphonamide antagonist free medium

STA = Sensitivity test agar

Laboratories using the Stokes method, significant differences in error rates

DST < Mueller-Hinton (χ^2 30.9, $p < 0.001$) and SAF (χ^2 10.7, $p < 0.01$)

Sensitest < SAF (χ^2 7.4, $p < 0.01$)

and Mueller-Hinton (χ^2 21.6, $p < 0.001$)

Isosensitest < DST (χ^2 6.7, $p < 0.01$), Sensitest (χ^2 5.2, $p < 0.05$),

SAF (χ^2 20.7, $p < 0.001$) and Mueller-Hinton (χ^2 43, $p < 0.001$)

STA < SAF (χ^2 9.2, $p < 0.01$) and Mueller-Hinton (χ^2 19, $p < 0.001$)

Table 3 Distribution of incorrect results for sulphonamide, trimethoprim and cotrimoxazole according to use of media with or without lysed blood (combined results for specimens 506, 521, 536, 549 and 570)

Medium	Lysed blood used	No of labs	No of results Right	Wrong	Ratio of right:wrong	χ^2	p
DST	No	54	530	149	4	69.9	<0.001
	Yes	143	1667	173	10		
Sensitest	No	42	486	66	7	1.73	NS
	Yes	16	182	17	11		
Wellcotest	No	1	9	0	—	—	—
	Yes	0	0	0	—		
SAF	No	3	29	7	4	0.39	NS
	Yes	6	58	10	6		
Isosensitest	No	41	487	44	11	0.4	NS
	Yes	7	89	6	15		
Mueller-Hinton	No	21	223	30	8	4.78	<0.05
	Yes	12	117	28	4		
STA	No	7	93	5	19	1.6	NS
	Yes	5	69	1	69		
All media	No	181	2002	329	6	22.4	<0.001
	Yes	203	2340	253	9		

NS = not significant.

the score obtained with the 27 general bacteriology specimens distributed in the same period for each laboratory was calculated. The method of assessing performance and scoring of laboratories in general bacteriology has been previously described.¹ A correlation coefficient of 0.482 was obtained suggesting significant association of performance as measured by the two features ($p < 0.001$).

METHOD

The Stokes method,⁴ in which test and control strains are inoculated on the same plate, was used by 182 laboratories. Another 116 laboratories apparently compared test and control zone sizes by other methods. Zone sizes were interpreted by the use of regression lines, tables, templates or similar devices in 37 laboratories. For 21 laboratories the basis of interpretation of zone sizes was not explained. The breakpoint method was used by three laboratories. The questionnaire was not detailed enough to elucidate sufficient details of the above general methods to justify comparison of performance.

MEDIA

The media used by participants are shown in Table 2 together with the distribution of right and wrong results for each of them. Several significant differences between the ratios of right and wrong results obtained with the various media are evident when considering separately, results from all laboratories, results from laboratories using the Stokes method and the results from laboratories using other methods in which controls are used (Table 2).

USE OF LYSED BLOOD

The error rates in testing sulphonamide,

trimethoprim and cotrimoxazole associated with the use of media with or without lysed blood are shown in Table 3. Considering the results on all media there were fewer errors made with these agents by laboratories adding lysed blood to the media than by those not adding lysed blood. The effect of adding lysed blood depends on the medium used. With the exception of Mueller-Hinton medium all media show fewer incorrect results when lysed blood was added, although only with DST medium is the effect statistically significant. The reverse effect was seen with Mueller-Hinton medium when addition of lysed blood was associated with an increase in the number of incorrect results.

METHODS FOR TESTING METHICILLIN SENSITIVITY

In 309 laboratories using media supplemented with NaCl and/or incubated at 30°C, 65/893 results were wrong when testing methicillin sensitivity (ratio right:wrong = 13). Neither of these methods was used in 45 laboratories and 43/125 results were wrong (ratio right:wrong = 2). The difference between these groups is highly significant ($\chi^2 85.0$, $p < 0.001$). These errors were almost exclusively associated with the failure to detect the presence of heteroresistance in *Staphylococcus aureus* MQCL 192.

STANDARDISATION OF INOCULUM

Laboratories professing to standardise the inoculum made fewer errors than those making no attempt at inoculum standardisation. In the 315 laboratories standardising the inoculum 1053/16129 of the combined results for all specimens were wrong (ratio right:wrong = 14) compared to 212/2592 wrong results (ratio right:wrong = 11) in the 52 laboratories not attempting to standardise the inoculum. The

Table 4 Distribution of incorrect results according to method of application of inoculum (combined results for all specimens)

Inoculum applied by:	No of labs	No of results		Ratio of
		Right	Wrong	right:wrong
Loop	14	688	50	14
Swab	255	12010	943	13
Loop followed by swab	70	3387	184	18
Flooding	21	986	110	9

difference between the two groups is significant (χ^2 9.6, $p < 0.01$). Details of methods used to standardise the inoculum were not ascertained.

INOCULUM

Laboratories emulsifying growth in fluid or subculturing to broth before inoculation made fewer errors than those using colonies directly as an inoculum. In the 323 laboratories emulsifying growth in fluid or subculturing to broth before inoculation 1181/16501 of the combined results for all specimens were wrong (ratio right:wrong = 13) compared to 188/1465 wrong results (ratio right:wrong = 7) in the 28 laboratories using colonies directly as an inoculum. The difference between these two groups is significant (χ^2 61.6, $p < 0.001$).

APPLICATION OF INOCULUM

The distribution of correct and incorrect results according to the method of application used is shown in Table 4. Laboratories applying the inoculum by flooding the plates made more errors than those using a loop (χ^2 5.9, $p < 0.02$), a swab (χ^2 11, $p < 0.001$) or a loop followed by a swab (χ^2 33.9, $p < 0.001$). Those using a loop followed by a swab made fewer errors than those using a swab alone (χ^2 19.9, $p < 0.001$).

USE OF CONTROLS

Laboratories using strains of known sensitivity to control their tests made fewer errors than those not using controls (Table 5). There was no significant difference between the number of errors made by laboratories using the Stokes method and those using controls with other methods (Table 5). *E. coli* strain

NCTC 10418, the recommended control strain for testing urinary isolates, was distributed in two specimens of the series. There was no significant difference between the number of errors made with this strain by laboratories using the Stokes method and those using controls with other methods (Table 5).

DISC CONTENT

The amount of antimicrobial agent in the discs used varied widely among laboratories even where methods such as Stokes's, which recommends particular disc contents, were used (Table 6). Association between the disc content and the number of incorrect results was significant for only three of the agents. For ampicillin, high content discs (mostly 25 μ g used) gave more reliable results for the organisms from urine than low content discs (10 μ g used) when controlled methods other than Stokes's were used (Table 7). For erythromycin, low content discs (mostly 5 μ g used) gave more reliable results for *Staph. aureus* from sites other than urine than high content discs (similar numbers of 10 and 15 μ g used) when controlled methods other than Stokes's were used (Table 7). For carbenicillin, high content discs (mostly 100 μ g used) gave more reliable results for *Pseudomonas aeruginosa* from sites other than urine than low content discs (mostly 5 μ g used) both when the Stokes method and other controlled methods were used (Table 7).

Discussion

Particularly with a small series of tests the error rates are influenced by the particular strains selected for testing and should not be taken as an indication of the overall error rates likely in routine laboratories. However, error rates with some tests appear high for this series of what were intended to be straightforward strains. It is possible to attribute some of the discrepancies to factors other than technical error. The high error rate with cephaloridine (17.9%) may have been to some extent associated with the undisclosed use of other cephalosporins. Laboratories were asked to test cephaloridine as class representative for the first

Table 5 Distribution of incorrect results according to use of controls (combined results for all specimens)

Comparison		No of labs	No of results		Ratio of	χ^2	p
			Right	Wrong	right:wrong		
Controls	Used	347	16531	1201	14	46.0	<0.001
	Not used	32	1415	181	8		
Control method (all specimens)	Stokes	182	8683	629	14	0.8	NS
	Other	143	6713	514	13		
Control method (<i>E. coli</i>) NCTC 10418	Stokes	182	1912	81	24	0.7	NS
	Other	143	1477	72	21		
Two distributions							

NS = not significant.

Table 6 Disc content used by laboratories using Stokes's method

	Amikacin	Ampicillin	Carbenicillin	Cephaloridine	Cotrimoxazole	Erythromycin	Fusidic Acid	Gentamicin	Methicillin	Penicillin*	Sulphonamide	Tetracycline	Tobramycin	Trimethoprim
Disc content (µg)	No of labs using stated disc content													
1	...	...	...	...	...	...	...	...	...	80+	...	...	...	...
1.25	...	...	...	...	...	...	...	...	...	1	...	...	...	(73)‡§
1.5	...	...	...	...	...	...	...	...	...	39	...	...	...	(1)
2	...	...	...	...	...	5	...	12	...	46§	...	...	1	(73)
2.5	...	...	...	...	...	...	...	...	...	...	...	...	...	...
4	...	...	...	...	...	...	...	...	...	5	...	...	...	...
5	...	...	...	...	...	106	...	1	18	...	...	...	1	(2)
10	132	(12)	1	...	...	34	161	160§ (168)§	36§	2	...	130§	160	(1)
15	...	...	...	...	...	29*	...	...	...	...	...	...	...	...
25	1	(159)§	16	(47)	(66)	...	...	...	...	...	(10)	31	...	...
30	19	(1)	...	(59)§	...	...	...	(1)	...	...	(1)	3	...	...
50	...	...	...	...	...	...	...	...	...	...	(21)	2	...	...
100	...	...	141§	...	...	...	...	(1)	...	...	(68)§	1	...	...
200	...	...	1	...	...	...	...	...	...	...	(49)	...	...	...
300	...	...	...	...	...	...	...	...	...	...	(9)	...	...	...
500	...	...	...	...	...	...	...	...	...	...	(9)	...	...	...
Response not clear/ >1/or other value	3	(4)	2	(12)	(1)	2	1	3 (4)	0	2	(2)	5	0	(3)
Not used/not stated	27	(6)	21	(64)	(115)	6	8	6 (8)	78	7	(13)	10	20	(29)

*Participants' responses often did not distinguish between µg and units.

‡§Results in parentheses indicate number of laboratories using stated disc contents for organisms from urine. Results without parentheses are disc contents for organisms from sites other than urines.

§Disc content recommended for the Stokes method*

generation cephalosporins. However, many laboratories were unhappy with using this compound for testing strains from urine because it is unlikely to be used clinically. When laboratories stated that other cephalosporins were used in the trial, the results were excluded from the analysis but it is probable that this information was not universally disclosed. However, the high error rates for cephaloridine cannot be explained solely on this basis because the errors were found even with strains sensitive to cephaloridine that were also sensitive to other cephalosporins. Laboratories were not given credit for reports of intermediate sensitivity in this trial

series because it was the intention that when possible the strains would be unequivocally sensitive or resistant. However, it is evident that room for differences in interpretation existed in some cases. The high error rate with carbenicillin (23.6%) was largely associated with interpretation of results from a strain with an MIC of 64 mg/l. The correct result was designated as sensitive but many participants considered that a report of intermediate would be more appropriate and were given a score of 0, the same as laboratories making the more serious error of calling the strain resistant. Similar problems may have arisen with the *Proteus* strain with an MIC of 8

Table 7 Distribution of incorrect results according to disc contents used (combined results for specimens on which relevant antibiotics were tested)

Antibiotic	Control method	Disc content (µg)	No of results Right	Wrong	Ratio of right:wrong	χ ²
Ampicillin	Stokes	10	60	0	—	2.3
		25/30	745	29	26	NS
	Method other than Stokes*	10	64	6	11	8.0
Erythromycin		25/30	535	13	41	p < 0.01
	Stokes	2/5	302	17	18	0.2
		10/15	173	8	22	NS
	Method other than Stokes*	2/5	193	6	32	23.4
		10/15	151	25	6	p < 0.001
Carbenicillin		10/25/50	13	16	0.8	21
	Stokes	100/200	226	50	5	p < 0.001
		10/25/50	14	15	0.9	21.2
	Method other than Stokes*	100/200	146	45	3	p < 0.001

mg/l for cephaloridine. Although this value indicates usual sensitivity for the species, it is close to the upper limit of the "sensitive" category. The errors with erythromycin may be due in part to laboratories failing to consider the possibility of dissociated resistance in one of the strains. Although it is possible to dispute the fairness of the marking system for individual strain/drug combinations, the differences in performance among laboratories and the correlation with the level of performance achieved in examining simulated specimens in the general bacteriology scheme indicate an uneven level of expertise in sensitivity testing.

It is apparent that a wide variety of techniques is in use. Even where techniques are used for which standard procedures have been described, many variations are practised. An example of this is the wide variety of disc contents used for the Stokes method. The number of incorrect results obtained for the *E coli* strain, NCTC 10418 even by laboratories using a method stipulating control by the same strain on the same plate, suggests that departures from recommended procedures are common. The details of methods used did not allow a form of analysis in which the superiority of any one general method of testing could be demonstrated. However, it is interesting to note that the results from laboratories using the Stokes method, the single most common method in use showed no fewer errors than in the combined results of laboratories using controls in methods other than Stokes's. The Stokes method appeared less sensitive to variations in disc content than those using controls in methods other than Stokes's but the error rates associated with the use of different media were similar for the Stokes and other methods using control organisms. The use of control strains evidently reduces errors although this may be because laboratories doing so are generally more aware of the problems of disc diffusion tests. A possible explanation of some of the minor errors with the *E coli* control strain NCTC 10418 is that stock control strains in individual laboratories may have become contaminated or transposed with other cultures although it is unlikely that this would result in the sensitive control organism being reported resistant. The importance of correct maintenance procedures and regular replacement of stock cultures is self evident.

Errors with the sulphonamide group were slightly reduced by the addition of lysed blood to the medium particularly where DST agar was used. The manufacturers of DST recommended the addition of lysed horse blood for sulphonamide and trimethoprim testing and there is little excuse for not following this advice. This effect is well known and the active agent in lysed horse blood has been

identified.⁵ An unexpected finding was the increase in errors when lysed blood was added to the Mueller-Hinton agar. Experience with the product of one manufacturer of Mueller-Hinton in one of our laboratories has shown that although the addition of lysed blood reduced "hazy" growth within zones, the edges of the zone were less distinct than on medium without lysed blood, causing difficulties in measurement. Detection of heteroresistance to methicillin is facilitated by incubation of tests at 30°C⁶ and/or the addition of sodium chloride to the medium.⁷ This was also evident with methicillin sensitivity testing of the heteroresistant strain distributed in the present series. At least one laboratory found heteroresistance in these strains but chose to report them as sensitive to methicillin, doubting the clinical significance of resistance at 30°C but not at 37°C. In the absence of conclusive evidence supporting this opinion it may be prudent to consider such strains resistant. The method of inoculum preparation and application exerted a marked influence on the error rate. The effects seen were probably attributable to the different inoculum densities achieved with the various methods. The importance of inoculum standardisation is demonstrated by the difference in error rates between laboratories standardising the inoculum and those not. The higher error rates associated with the direct use of the colony as an inoculum compared with suspension or subculture to liquid medium is probably a reflection of the difficulty in achieving a correct density of inoculum by the former method. The apparent superiority of inoculation by loop followed by spreading with a swab over flooding the medium was surprising because flooding is known to give an even lawn of growth with clearly defined zone edges. The apparent poor results achieved with flooding may be associated with other factors, for instance, it is not possible to use the Stokes method with inoculation by flooding.

The apparent association of error rates and the use of various media is interesting, although further evaluation with a larger series of strains is needed to substantiate these findings. The poor results associated with the use of Mueller-Hinton agar were unexpected and no explanation can be provided. This medium was originally designed for the growth of *Neisseria* and not all formulations are likely to be suitable for its more recent use in sensitivity testing. No attempt was made to associate results with the products of different manufacturers because the numbers involved were too small.

It is unlikely that the methodological information collected in this trial takes account of all the technical factors that can profoundly affect the results of sensitivity tests and care must be exercised when

interpreting the results of the trial. It is, however, disappointing to see laboratories continuing to use methods that fail to take account of sources of error that have been known and publicised for several years. The error rates found in this study are almost certainly artificially low because quality assessment specimens are a measure of the best that laboratories can perform. To remedy these deficiencies the following recommendations are made.

- 1 Disc strengths appropriate to the method of testing practised should be used. On the basis of the selected strains used in this study 25 µg ampicillin discs for organisms from urine, 100 µg carbenicillin discs for *Ps aeruginosa* and 5 µg erythromycin discs for *Staph aureus* appear most suitable.
- 2 For methicillin testing, incubation should be at 30°C and/or 5% NaCl should be added to the media.
- 3 The inoculum should be standardised to the density recommended for the method used (dense but not confluent growth for the methods most widely used by participants). Emulsification of the inoculum in fluid or subculture to broth followed by application with a loop and spreading by swab appeared to give the best results.
- 4 Daily controls should be used.
- 5 Lysed horse blood should be added to the medium if DST is used for testing sensitivity to sulphonamide and trimethoprim.

All participants have been informed of the results of this trial and it is hoped that they will introduce the recommended modifications. It is intended to continue the trial using similar organisms as well as some giving less clear cut results and reanalyse the results after a further year.

We thank members of the Steering Committee for Quality Control in Microbiology, Antibiotic Sensitivity Testing Subcommittee for advice on the planning of the trial and analysis of results. Members of the subcommittee are: Dr Joan Stokes (chairman), Dr PR Mortimer (secretary), Mr DFJ Brown, Dr PS Gardner and Dr GL Gibson.

References

- ¹ Snell JJS, de Mello JV, Gardner PS. The United Kingdom national microbiological quality assessment scheme. *J Clin Pathol* 1982;35:82-93.
- ² Stokes EJ, Whitby JL. Quality control in bacteriology: preliminary trials. *J Clin Pathol* 1971;24:790-7.
- ³ Ericsson HM, Sherris JC. Antibiotic sensitivity testing. Report of an International Collaborative Study. *Acta Pathol Microbiol Scand [B]* 1971;suppl 217:1-90.
- ⁴ Stokes EJ, Waterworth PM. Antibiotic sensitivity tests by diffusion methods. *ACP Broadsheet* 55, 1972.
- ⁵ Ferone R, Bushby SRM, Burchall JJ, Moore WD and Smith D. Identification of Harper-Cawston factor as thymidine phosphorylase and removal from media of substances interfering with susceptibility testing of sulphonamides and diaminopyrimidines. *Antimicrob Agents Chemother* 1975;7:91-8.
- ⁶ Annear DI. The effect of temperature on resistance of *Staphylococcus aureus* to methicillin and some other antibiotics. *Med J Aust* 1968;1:444-6.
- ⁷ Barber M. Naturally occurring methicillin-resistant staphylococci. *J Gen Microbiol* 1964;35:183-90.

Requests for reprints to: Dr PS Gardner, Director, Division of Microbiological Reagents and Quality Control, 175 Colindale Avenue, London NW9 5HT, England.