Evaluation of a micromethod gallery (API Staph) for the identification of staphylococci and micrococci

RR MARPLES, JF RICHARDSON

From the Division of Hospital Infection, Central Public Health Laboratory, Colindale Avenue, Colindale, London NW9 5HT

SUMMARY A collection of 300 well-characterised strains of staphylococci and micrococci was examined by a commercially available gallery micromethod (API Staph). The results were compared with biotyping by conventional methods. The gallery micromethod broadly agreed with the biotyping scheme used but gave an identification from the index supplied in less than 30% of the trials. Reproducibility was better after 48 h incubation than after 24 h but was poor for the tests for phosphatase and acetoin. When compared with the results of conventional tests, the tests for acetoin, phosphatase and urea were unsatisfactory.

Two schemes based mainly on simple biochemical tests have been widely used for the identification of coagulase-negative staphylococci. The first, described by Baird-Parker¹ and later modified by him^{2,3} and others,^{4,5} is based on the acidification of solid media containing a small number of "sugars", and the phosphatase and acetoin tests. The second, described by Schleifer, Kloos and co-workers,⁶⁻⁹ tests rather more "sugars" but employs different bases and incubation temperatures. This scheme was used as the basis of a micromethod identification gallery (API Staph) which is now widely available.¹⁰ In a study of the Micrococcaceae, Feltham¹¹ included the API 50E and 20 E galleries, but he did not use the API Staph gallery.

A collection of 300 well-characterised strains of staphylococci and micrococci available to us¹² was used to assess the new gallery.

Material and methods

SOURCE OF STRAINS

The 300 strains were selected to include approximately one-third of named strains from culture collections, one-third of strains from newborn babies, schoolchildren and adults isolated in this Division and one-third of strains isolated elsewhere.¹³ As far as possible strains were selected to include 10 of each Baird-Parker biotype. The strains were lyophilised and working cultures kept at -20°C in glycerol

broth. The strains were subcultured onto nutrient agar, and incubated at 30°C for 24 h.

MICROMETHOD IDENTIFICATION GALLERY (API STAPH)

The gallery consisted of 20 cupules containing substrates and appropriate indicators in a dried form. The order of the substrates was: negative control, glucose, fructose, mannose, maltose, lactose, trehalose, mannitol, xylitol, melibiose, nitrate, α naphthyl phosphate, pyruvate (for the production of acetoin), raffinose, xylose, sucrose, α methyl glucoside, N acetyl glucosamine, arginine and urea.

INOCULATION PROCEDURE

The procedure recommended by the manufacturers was followed. For each strain an ampoule of the suspending medium (modified from Brun *et al*¹⁰) was opened. A small amount of growth from nutrient agar was picked up on the tip of a pasteur pipette and suspended in the ampoule by repeatedly taking medium into the pipette, to an opacity equivalent to tube 2 on the McFarland scale. The tube parts of the cupules were then carefully filled without bubbles. Only the cupules containing arginine and urea were sealed with sterile liquid paraffin. Each gallery was placed in a plastic incubation tray containing 5 ml of water to maintain humidity, covered and incubated at 37°C.

RECORDING PROCEDURES

Readings of the tests containing pH indicators were made after 24 and 48 h and reagents added to the

cupules containing nitrate, α naphthyl phosphate and pyruvate after incubation for 48 h.

For the nitrate test one drop of 0.8% sulphanilic acid in 5N acetic acid and one drop of 0.6% naphthyl-1-amine in 5N acetic acid were added. A red colour developing immediately was taken as a positive test.

For the phosphatase test one drop of reagent A (trishydroxymethyl) aminomethane 250 g; HCl (37%) 110 g; lauryl sulphate 100 g, distilled water to one litre and one drop of reagent B (fast blue BB 3.5 g, 2 methoxyethanol to one litre) were added to the α naphthyl phosphate cupule. A violet colour developing after 10 min was taken as a positive test.

To the pyruvate cupule were added single drops of 40% KOH and of 6% α naphthol in ethanol. A red colour developing after 10 min was taken as a positive test. Strains showing anomalous results were inoculated into two galleries and reagents added at 24 h to one while the other was treated as before.

Four months later the glycerol broth suspensions were resubcultured and all 300 strains were retested in the same way.

In order to identify the organism from the index supplied by the manufacturer the results were expressed as a numerical code. The tests in order along the gallery were divided into six groups of three tests and a final group of two tests. A positive reaction in the first position in a group was given value 1, in the second position, 2, and in the third position, 4. The sum of these values within each group made up one digit of the seven figure code numbers. The index listed the codes in numerical order and indicated the appropriate species name.

The results were also tabulated according to the biotyping scheme of Marples.¹² This scheme is a modification of the scheme of Baird-Parker but is based on studies of the present strains by conventional methods.

CONVENTIONAL METHODS

The conventional methods used to characterise the strains were those of Baird-Parker¹ (method A), those of Kloos and co-workers⁸⁻⁹ (method B) or the standard methods, usually those of Cowan and Steel.¹⁴ Acid production from arabinose, fructose, lactose, maltose, mannitol, ribose, sucrose, trehalose, xylitol and xylose was sought by both methods A and B. In addition, acidification of cellobiose, fucose, galactose, mannose, melezitose, raffinose, salicin and turanose was tested by method A only. Reduction of nitrate was tested by methods A and B and method 1 of Cowan. Acetoin was sought by methods A and B. Phosphatase production was assessed by methods A and B and by the method of Pennock and Huddy.¹⁵ Arginine hydrolysis was tested by method A,

Møller's method and by incubation of arginine broth for 24 h.¹⁴ Method B was employed for the erythromycin-glycerol test, thioglycollate broth, and for resistance to lysozyme and lysostaphin. A variety of other tests including coagulase, gelatinase, urease, nuclease, splitting of Tweens and the egg yolk reaction were performed by standard methods.

ANALYSIS OF RESULTS

The reproducibility of the micromethod gallery was assessed by comparing the results recorded at 24 and 48 h incubation when tested on the two occasions. They were compared with the results of conventional tests, taking the results of method A where different methods gave conflicting results as these seemed the more sensitive.¹² Much of the analysis had necessarily to be restricted to biotypes of *Staphylococcus* spp as the micromethod was not designed for *Micrococcus* spp.

Results

IDENTIFICATION BY CONVENTIONAL METHODS

On the results of the relevant conventional tests all strains were identified by both identification schemes. Then the strains were reclassified on all the results to form a new extension of the scheme of Baird-Parker in which the major divisions were subdivided to define more than 46 biotypes.¹² The correlation of the three schemes for the staphylococcal biotypes is shown in Table 1. In the Table the number of strains allocated to the new biotypes are listed and within each biotype the numbers and biotype of strains identified by the other two schemes are shown. As the identifications were very strictly made according to the published tables a few individual strains were probably misidentified. Single appearances of a different biotype have therefore been omitted from the Table.

It is evident from the Table that there was good agreement between the three schemes for some biotypes. The subdivisions of SI, SII and SIII showed a good correlation. There was relatively poor agreement over strains allocated to the subdivisions of SIV and SV and several anomalies in the subdivisions of the biotypes of SVI and M3 that were novobiocin sensitive though M3 g and *S. capitis* seemed identical. The saprophyticus species group (SVC, M3 a, b and MXa, b) and the xylosus species group correlated well.

RESULTS OF THE MICROMETHOD

The instructions for reading the gallery suggested that the glucose cupule be examined for acidification at 24 h and incubation continued to 48 h only in the

Table 1 Correlation between biotyping systems for staphylococcal strains

No of strains	Marples biotype ¹²	Baird-Parker biotype ¹⁵ and No of strains in that biotype	Schleifer and Kloos species ⁸ and No of strains in that species
20	SIa	SI	<i>S aureus</i>
8	SIb	SI	<i>S sp</i>
23	SIIfa	SIIf	<i>S epidermidis</i>
8	SIIfb	SIIf lac ⁻	<i>S epidermidis</i>
6	SIIfc	SV	<i>S hominis</i> 3; <i>S epidermidis</i>
13	SIIfa	SIIf mal ⁺	<i>S simulans</i>
12	SIIfb	SIIf	<i>S simulans</i>
6	SIIfc	<i>S sp</i>	<i>S simulans</i>
9	SIVa	SIv 7; M1	<i>S warneri</i> 6; <i>S hominis</i>
9	SIvb	M1 6; SIv	<i>S hominis</i> 6; <i>S warneri</i>
4	SIvc	SIv 2; M1	<i>S haemolyticus</i>
4	SVaa	M2	<i>S haemolyticus</i> 2; <i>S hominis</i>
7	SVab	SV	<i>S hominis</i>
2	SVb	M2	<i>S haemolyticus</i>
3	SVd	SV	<i>S haemolyticus</i>
3	SVe	<i>S sp</i> 1; SIIf 1; SV	<i>S hominis</i>
6	SV1a	SV1 (1) 4; SV1 (2)	<i>S haemolyticus</i>
3	SV1b	SV1 (2)	<i>S warneri</i> 1; <i>S capitis</i> 1; <i>S haemolyticus</i>
8	SV1c	SV1 (1) 5; M3 (1)	<i>S haemolyticus</i> 6; <i>S warneri</i>
13	SV1d	SV1 (2)	<i>S warneri</i> 11; <i>S haemolyticus</i>
2	SV1e	SV	<i>S cohnii</i>
6	M3c	M3 (2)	<i>S haemolyticus</i> 3; <i>S hominis</i>
4	M3e	SV1 (1)	<i>S haemolyticus</i>
4	M3f	SV1 (2)	<i>S haemolyticus</i>
11	M3g	M3 (3) 7; SV1 (3)	<i>S capitis</i>
4	M3h	SV1 (3) 2; M3 (3)	<i>S capitis</i> 2; <i>S haemolyticus</i>
3	SVc	M2	<i>S saprophyticus</i>
13	M3a	M3 (1)	<i>S saprophyticus</i>
5	M3b	M3 (2)	<i>S saprophyticus</i>
2	MXa	MX	<i>S cohnii</i>
14	MXb	MX	<i>S cohnii</i>
6	M5a + b	M5	<i>S xylosus</i>
10	M6a + b	M6	<i>S xylosus</i>

absence of such acid production. All but three of the 251 strains of staphylococci produced acid from glucose within 24 h and all by 48 h, but 45 of the 98 tests on the 49 strains of micrococci remained negative at 48 h and in 15 trials all tests were negative. The micrococcal strains were not considered further.

MICROMETHOD RESULTS RELATED TO BIOTYPE

Table 2 shows the results of each of the 19 tests for each of the staphylococcal biotypes after incubation for 48 h. Each strain was tested twice and the result was recorded as positive if more than 80% of the tests were positive, negative if more than 80% of the tests were negative and variable if the number of positive tests lay between 20% and 80%. This procedure was selected because of the small number of strains in some biotypes and because of the problems of reproducibility which are discussed later.

The results with the micromethod gallery largely confirmed the biotyping system. This was not

surprising as the biotypes were defined in terms of some of the same tests, though carried out by conventional methods. The biotypes of SIIf were defined on lactose negativity for SIIf b and phosphatase negativity for SIIf c. These were distinguished by the micromethod. The biotypes of SIIf were defined on maltose positivity for SIIf a and mannitol positivity for SIIf c. These differences showed in the micromethod results but some anomalies were evident, for example SIIf was defined as acetoin-negative but gave variable results with the micromethod. The greatest numbers of variable results were seen in the biotypes of *warneri-hominis* species group (SIv and SV) and in the *haemolyticus-capitis* group (SV1 and M3 c-h). These are the least well defined in both classifications but within the biotypes some, such as M3 g (*S capitis*), were reasonably characterised.

The *saprophyticus-cohnii* species group (M3 ab, Mx and SV c) strains were well characterised by the micromethod and the variable positive results with xylitol could be expected. However, strains of Mx (*S cohnii*), defined as phosphatase-positive, were variable by the micromethod.

Strains of the *xylosus* species group (M5 and M6), defined as acetoin-negative were positive in the majority of micromethod trials, but the acidification of xylose was confirmed.

IDENTIFICATION FROM THE INDEX

After 24 h, 93 (18.5%) of 502 tests gave an acceptable identification from the index rising to 137 (27.5%) using the 48 h results. Only 10 strains were clearly misidentified at 24 h, in seven trials this was an identification as *S. capitis* of strains that later acidified additional substrates. The 48 h results misidentified one strain of *S. epidermidis* (SII) as *S. aureus* and one strain of biotype SV ab as *S. saprophyticus* each on one occasion. The improvement in identification from the index using the 48 h results extended over all biotypes with the exception of *S. capitis* and *S. saprophyticus* where sucrose and xylitol respectively were not included as positive in the listed codes. Biotypes that were reasonably identified were restricted to *S. aureus*, *S. epidermidis* (SII a), *S. capitis* (M3 g) and *S. saprophyticus* (M3 a) but it was clear that the index could be extended to improve identification if needed.

REPRODUCIBILITY

The results of the 16 tests read at 24 h on each time of testing were compared, a total of 251 comparisons for each test. Discrepancies of greater than 5% were found in the results for mannose, maltose, mannitol, N-acetyl glucosamine, arginine and urea. Only urea gave more than 10% discrepancies in the two results. Reproducibility was excellent (<1% discrepancies) for glucose, fructose, raffinose, xylose, and α methyl glucoside.

The results of all 19 tests could only be compared at 48 h since reagents for nitrate, phosphatase and acetoin were added to the majority of the galleries at that time. More than 5% discrepancies were found in the results for mannose, phosphatase, acetoin, N acetyl glucosamine and arginine. Reproducibility of the phosphatase test was least with 49 of 251 (19.5%) results discordant and acetoin was little better with 10.8% of the results discordant. Reproducibility was improved at 48 h over the 24 h results for maltose, mannitol and urea but worsened for mannose. As with the 24 h results, excellent reproducibility was found for glucose, fructose, raffinose, xylose and α methyl glucoside.

COMPARISON OF THE MICROMETHODS WITH CONVENTIONAL TESTS

Only 17 of the 19 tests could be compared with conventional tests, usually to the method of Baird-Parker.¹ If the micromethod gave a positive result but the conventional method gave a negative result

the comparison was deemed a "false-positive" and the reverse "false-negative" (Table 3). More than 5% false-positives were found in the results of mannose, maltose, phosphatase and acetoin. The acetoin test gave the greatest number of false-positives with 75 of the 502 tests (14.9%) discordant. False-negative results in greater than 5% of the comparisons were recorded for phosphatase and urea. Urea was falsely recorded as negative in 57 of the 502 tests (11.4%). Good comparability (<1% of the tests false) was seen in tests for glucose, fructose, raffinose and xylose. The test for the reduction of nitrate showed a number of false-negative results at the first replication of the study. Repeat testing showed that reduction past nitrate had occurred. For repeat testing and the second replication zinc dust was added with a wooden applicator stick to each apparently negative nitrate test and the result recorded as negative only if nitrate was reduced by the zinc and the colour developed. This confirmatory step was sometimes necessary even after only 24 h incubation.

It was possible to examine the frequency of "late-positives" at the same time where the micromethod gave a negative result at 24 h but a positive result at 48 h and the conventional method was positive. More than 5% of late-positives were seen for mannose, maltose, mannitol, arginine and urea. In 86 (17.1%) of the urea tests a correct positive was delayed.

Discussion

The classification of the staphylococci and micrococci remains a matter of controversy but a consensus appears to be evolving. The schemes of Baird-Parker¹ and Kloos and co-workers⁷⁻⁹ do show some correlation of biotypes.⁵ Present schemes, particularly that of Marples¹² may be subdividing true species into biotypes that have epidemiological but not taxonomic validity. The reversal of the "splitting" movement has created the concept of species group¹⁶ and this trend may continue. It therefore becomes necessary to review the reasons for identifying species and biotypes of staphylococci and to consider the value of a convenient micromethod gallery in medical microbiology.

The main reason for identification of staphylococci from clinical sources is to separate *S. aureus*, a frequent pathogen, from the other less pathogenic species. At a higher level the identification of *S. epidermidis* and *S. saprophyticus* may define clinical syndromes. While *S. epidermidis* is the commonest coagulase-negative staphylococcus species associated with hospital-acquired infection¹⁷ *S. saprophyticus* is frequently associated with urinary tract infection in

Table 2 Results of API Staph on strains of different biotypes after incubation for 48 h

	No of tests	Glucose	Fructose	Mannose	Maltose	Lactose	Trehalose	Mannitol	Xylitol	Melibiose
SI	a 39	+	+	+	+	+	+	+	-	-
	b 16	+	+	+	+	+	+	V	-	-
SII	a 44	+	+	V	+	+	-	-	-	-
	b 16	+	+	V	+	-	-	-	-	-
	c 12	+	+	-	+	+	-	-	-	-
SIII	a 26	+	+	+	+	+	+	-	-	-
	b 24	+	+	V	-	+	+	-	-	-
	c 12	+	+	V	-	+	+	+	-	-
SIV	a 18	+	+	V	+	-	+	-	-	-
	b 18	+	+	V	+	-	+	-	-	-
	c 8	+	-	-	+	-	+	V	-	-
SV	aa 8	+	+	V	+	+	+	-	-	-
	ab 14	+	+	-	+	V	+	-	-	-
	b 4	+	+	-	+	+	+	-	-	-
	d 6	+	-	-	+	+	+	-	-	-
	e 6	+	+	+	+	+	+	-	-	-
SVI	a 12	+	+	V	+	V	V	V	-	-
	b 6	+	V	V	+	-	-	+	-	-
	c 16	+	+	V	+	+	+	+	-	-
	d 26	+	+	V	+	-	+	+	-	-
	e 4	+	+	+	+	+	+	+	-	-
M3	c 12	+	+	-	+	-	+	+	-	-
	e 8	+	-	-	+	+	+	+	-	-
	f 8	+	-	-	+	-	-	+	-	-
	g 22	+	+	+	-	-	-	V	-	-
	h 8	+	+	-	-	-	+	+	-	-
SV	c 6	+	+	-	+	+	+	-	V	-
M3	a 26	+	+	-	+	+	+	+	V	-
	b 10	+	+	-	+	-	+	+	V	-
Mx	a 4	+	+	+	+	+	+	+	V	-
	b 28	+	+	-	+	-	+	+	V	-
M5	a 6	+	+	+	+	+	+	+	-	-
	b 6	+	+	+	+	V	+	+	-	-
M6	a 16	+	+	+	+	+	+	+	-	-
	b 4	+	+	+	+	V	+	+	-	-

+ >80% of the tests positive; - <20% of the tests positive.
V 20-80% of the tests positive.

Table 3 Comparison of API Staph with conventional tests

Test	Percentage of 502 tests classified as		
	False-negative	False-positive	Late-positive
Glucose	0	0	0
Fructose	0	0.4	0
Mannose	2.0	8.6	5.4
Maltose	0.6	6.0	5.2
Lactose	0.6	2.4	4.6
Trehalose	0.2	1.8	2.6
Mannitol	3.4	1.8	7.2
Xylitol	1.4	4.6	1.6
Nitrate	1.6	1.4	NA
Phosphatase	6.6	12.0	NA
Acetoin	3.8	14.9	NA
Raffinose	0.4	0.6	0
Xylose	0	0.2	0.6
Sucrose	0.6	3.8	4.6
Arginine	3.6	3.4	7.6
Urea	11.4	0.6	17.1

NA = Not applicable.

young women.^{18 19} Other species may cause infection or have to be distinguished from the three most important species but usually do not need to be identified taxonomically.²⁰

Biotyping implies an epidemiological purpose: to distinguish between closely allied organisms in order to determine the likely cause of infection or course of an epidemic. Like all typing schemes typability, reproducibility and discrimination become characters to be examined. Once the three common species are defined biotyping may subdivide the groups and, by including biotypes of the less frequent species, can be helpful in all infections caused by staphylococci. Other typing methods such as phage typing also may be helpful in epidemiological studies.

The micromethod gallery (API Staph) offers many advantages in that the results can be recorded as a code for epidemiological purposes, and by using an index, a taxonomic identification can be reached.

For many years it has been known that the ability to split urea shows considerable variation in the rate of development of a positive test. For example, a presumptive identification of *Proteus* is often made on the ability of this group to alkalise urea solutions within four hours. Staphylococcal strains vary widely in the speed of this reaction and it is therefore not surprising to find urea figuring prominently in both the late-positive and false-negative groups.

The test giving the largest number of positive results where the conventional method was negative was the production of acetoin. In the method of Baird-Parker the primary substrate is glucose in an unbuffered medium. In the micromethod the substrate is pyruvate. The difference in substrate may explain the high false-positive rate. Some biotypes have been defined on the basis of acetoin negativity by the method of Baird-Parker, so the false-positives were concentrated in these biotypes. Further work is necessary to determine whether these false-positives were truly false.

The micromethod gallery appears to have value as an epidemiological biotyping method and for the identification by index of the three species that should be identified in all laboratories concerned with diagnostic medical microbiology. The tests included need improvement. Melibiose, raffinose, and α methyl glucoside were acidified very rarely and contributed little. The reproducibility and reliability of the test for phosphatase were unsatisfactory. Many tests required 48 h incubation but to recommend this would increase the false-positivity rate of mannose. The procedure employed here of reading the sugar reactions at 24 and 48 h and adding reagents at 48 h may be recommended.

We thank API Laboratories, La Balme les Grottes, France for donating the API Staph galleries, and Mrs J Malone for technical assistance.

References

- ¹ Baird-Parker AC. A classification of micrococci and staphylococci based on physiological and biochemical tests. *J Gen Microbiol* 1963;**30**:409-27.
- ² Baird-Parker AC. The classification of staphylococci and micrococci from world-wide sources. *J Gen Microbiol* 1965;**38**:363-87.
- ³ Baird-Parker AC. The basis for the present classification of staphylococci and micrococci. *Ann NY Acad Sci* 1974;**236**:7-14.
- ⁴ Pelzer K, Pulverer G, Jeljaszewicz J, Pillich J. Modification of Baird-Parker's classification system of *Staphylococcus albus*. *Med Microbiol Immunol* 1973;**158**:249-57.
- ⁵ Marples RR. Coagulase-negative staphylococci—their classification and problems. *RSM Int Cong Symp* 1980;**23**:57-64.
- ⁶ Kloos WE, Tornabene TG, Schleifer KH. Isolation and characterisation of micrococci from human skin including two new species: *Micrococcus lylae* and *Micrococcus kristinae*. *Int J Syst Bacteriol* 1974;**24**:79-101.
- ⁷ Schleifer KH, Kloos WE. Isolation and characterisation of staphylococci from human skin. I. *Int J Syst Bacteriol* 1975;**25**:50-61.
- ⁸ Kloos WE, Schleifer KH. Isolation and characterization of staphylococci from human skin. II. *Int J Syst Bacteriol* 1975;**25**:62-79.
- ⁹ Kloos WE, Schleifer KH. Simplified scheme for routine identification of human *Staphylococcus* species. *J Clin Microbiol* 1975;**1**:82-8.
- ¹⁰ Brun Y, Fleurette J, Forey F. Micromethod for biochemical identification of coagulase-negative staphylococci. *J Clin Microbiol* 1978;**8**:503-8.
- ¹¹ Feltham RKA. A taxonomic study of the Micrococcaceae. *J Appl Bacteriol* 1979;**47**:243-54.
- ¹² Marples RR. Taxonomic studies of staphylococci and micrococci. *Zentralbl Bakteriol* 1981; suppl 10, 9-13.
- ¹³ McKay-Ferguson E. *Organic acids in microbiology*. University of London MD thesis, 1977.
- ¹⁴ Cowan ST. *Cowan & Steel's manual for the identification of medical bacteria*. 2nd ed. Cambridge: Cambridge University Press, 1974.
- ¹⁵ Pennock CA, Huddy RB. Phosphatase reaction of coagulase-negative staphylococci and micrococci. *J Pathol Bacteriol* 1967;**93**:685-7.
- ¹⁶ Kloos WE, Schleifer KH, Noble WC. Estimation of character parameters in coagulase-negative *Staphylococcus* species. *Zentralbl Bakteriol* suppl 5, 1976;23-41.
- ¹⁷ Holt RJ. The pathogenic role of coagulase-negative staphylococci. *Br J Dermatol* 1972;**86** Suppl 8:42-7.
- ¹⁸ Mitchell RG. Classification of *Staphylococcus albus* strains isolated from the urinary tract. *J Clin Pathol* 1968;**21**:93-6.
- ¹⁹ Maskell R. Importance of coagulase-negative staphylococci as pathogens in the urinary tract. *Lancet* 1974;**ii**:1155-8.
- ²⁰ Subcommittee on the taxonomy of staphylococci and micrococci. Minutes. *Int J Syst Bacteriol* 1976;**26**:332-4.

Requests for reprints to: Dr RR Marples, Central Public Health Laboratory, Division of Hospital Infection, 175 Colindale Avenue, London NW9 5HY, England.