Technical method

A reduced incubation period for the destruction of somatic adenosine triphosphate in urine specimens

PW MCWALTER,* CA SHARP, GE STERN *Department of Bacteriology, University Hospital of South Manchester, Withington Hospital, Manchester M20 8LR*

The use of luciferin-luciferase adenosine triphosphate (ATP) assay as a rapid alternative to cultural methods for bacteriuria screening has been recently reported.^{1,2} Measurement of bacterial ATP in urine is dependent on effective removal of host-derived somatic ATP. Various methods have been described¹⁻⁸ which destroy or remove unwanted somatic ATP, however, many of the earlier techniques³⁻⁷ involved complex or time-consuming methodology. The advent of a commercially available somatic nucleotide releasing agent (NRS; Lumac BV, Schaesberg, The Netherlands) has greatly simplified these earlier procedures. The addition of NRS in equal volume to a sample of urine allows the quantitative release of ATP from somatic cells within a period of 10-15 s.⁸ Subsequent destruction of released somatic ATP using a high quality ATPase enzyme usually involves incubation at room temperature for 45 min.^{1,2,8} This study was carried out to determine whether higher temperature incubation (37°C) for a reduced period (20 min) would significantly alter the sensitivity of the screening technique.

Material and methods

One thousand routine urine specimens were screened for bacteriuria using two preincubation methods for the destruction of somatic ATP, 45 min at room temperature and 20 min at 37°C. The method, equipment and reagents used to carry out the 45 min room temperature preincubation technique have been described.¹ The same method was used for the rapid technique with preincubation being carried out in a 37°C incubator for 20 min. Specimens giving a relative light unit (RLU) value of greater than 711 (10⁴ organisms/ml)¹ by either or both techniques were cultured on salt-free McCon-

key agar (Oxoid code CM7B) using a standard loop technique (0.002 ml) and examined microscopically for the presence of cells. Culture plates were examined for significant bacterial growth after overnight incubation at 37°C.

Results

The results of 1000 specimens screened by bioluminescence using both preincubation methods are shown in the Table. There appeared to be good correlation between the two methods ($p < 0.001$). Nineteen specimens (1.9%) giving a bioluminescence-positive result at 37°C were negative using the routine room temperature method. All specimens negative by the 20-minute method were also negative by the 45-minute method.

Discussion

The correlation between the two preincubation methods was excellent ($p < 0.001$). Variation of results between the two techniques was noted in only 19 specimens (1.9%), all of which were bioluminescence-positive using the rapid method but negative using the 45-minute method and gave no bacterial growth on cultural investigation. Microscopical examination of these specimens showed that all had increased white and/or red cell counts ($>100/\text{mm}^3$). Although the somase would have an increased reaction rate at the higher incubation temperature, we felt in these cases the period of incubation was insufficient to allow complete removal of released somatic ATP giving rise to "false-positive" bioluminescence results.

It has previously been shown¹ that the use of a 1×10^4 organism/ml screening threshold removes the occurrence of false-negative (bioluminescence-negative/culture-positive) results. No increase in false-negative results occurred using the rapid technique. The 20-minute at 37°C preincubation method reduces the time previously required^{1,2,8} to remove somatic ATP without significantly altering the effectiveness of the reagents involved. The previously described advantages of rapid bioluminescence bacteriuria screening¹⁻⁴ may be further enhanced by the use of this shortened preincubation technique. These advantages must however be balanced against the slightly increased number (1.9%) of specimens requiring cultural investigation. We feel this rapid preincubation technique is a useful adjunct to bioluminescence bacteriuria screening.

Results of 1000 urine specimens screened by bioluminescence using two preincubation techniques

	37°C for 20 min	Room temp for 45 min
Bioluminescence-negative	608 (60.8%)	627 (62.7%)
Bioluminescence-positive	392 (39.2%)	373 (37.3%)

We would like to thank Dr DM Jones and Mr P Davis for advice and encouragement, Mrs B Jinks who typed the manuscript and our colleagues who ran the screening system on a routine basis.

References

- McWalter PW, Sharp CA. Evaluation of a commercially available semiautomated bioluminescence system for bacteriuria screening. *Eur J Clin Microbiol* 1982;14:223-7.
- Ruokenen A, Koskinen M, Leinonen M, Tillikainen A, Vihko R. Detection of urinary infection by rapid luciferase assay of adenosine triphosphate: A comparative study using urine sediment findings and various culture techniques. In: 3rd International Symposium on Rapid methods and Automation in Microbiology, American Society of Microbiology, Washington DC: 1981.
- Johnston HH, Mitchell CJ, Curtis GDW. An automated test for the detection of significant bacteriuria. *Lancet* 1976;ii:400-8.
- Curtis GDW, Johnston HH. A rapid screening test for bacteriuria. In: Schram E, Stanley PE, eds. Proceedings of International Symposium on Analytical Applications of Bioluminescence and Chemiluminescence. Westlake Village, California 91361: State Printing Publishing, 1978:448-57.
- Alexander DN, Ederer GM, Matsen JM. Evaluation of an adenosine 5'-triphosphate assay as a screening method to detect significant bacteriuria. *J Clin Microbiol* 1976;3:1:42-6.
- Conn RB, Charache P, Chappelle EW. Limits of applicability of the firefly luminescence ATP assay for the detection of bacteria in clinical specimens. *Am J Clin Pathol* 1975;63:493-501.
- Thore, A, Lundin AA, Bergman S. Detection of bacteriuria by luciferase assay of adenosine triphosphate. *J Clin Microbiol* 1975;1:1-8.
- Vanstaen H. Current developments in bioluminescence and chemiluminescence techniques for the medical laboratory. *Med Lab World* 1979;December;16-23.

Requests for reprints to: Mr PW McWalter, Chief Medical Laboratory Scientific Officer, Withington Hospital, West Didsbury, Manchester M20 8LR, England.

Letters to the Editor

Hypothermia and pancreatitis

Dr Foulis' recent study¹ relating hypothermia with the morphology of the associated acute pancreatitis is of great interest. This work, together with Dr Foulis' previous comprehensive survey² relating the histological pattern of pancreatitis with various clinical diagnoses may shed some light on the pathogenesis of pancreatitis.

We would like to report certain findings of our own and consider their relevance to Dr Foulis' studies.

We have shown that at temperatures below 37°C the inhibitory action of prostacyclin (PGI₂) on in vitro platelet aggregation is diminished.^{3,4} Furthermore, the in vitro synthesis of PGI₂ by vascular

endothelium is decreased at low temperatures.⁵ We have also suggested that at low temperatures the vasodilatory action of PGI₂ is diminished.⁶ The raised serum non-esterified fatty acid (NEFA) concentrations reported in hypothermic patients⁷ may be of relevance since we have shown that high NEFA concentrations inhibit vascular PGI₂ synthesis⁸ and accelerate PGI₂ decay in albumin solutions.⁹

Hypothermia may thus impair synthesis, increase the rate of decay and diminish the potency of PGI₂. These changes would result in platelet activation and predisposition to thrombosis and ischaemia. Abnormal platelet behaviour has been demonstrated in an animal study where thrombocytopenia occurred during induced hypothermia.¹⁰ An increased incidence of

thrombotic phenomena in patients with hypothermia is evident from Dr Foulis' own observations and from the literature he cites.¹

Ischaemic damage to the pancreas may occur in hypothermic patients as a result of platelet activation (as discussed above) or of "microcirculatory shock" as suggested by Dr Foulis. Such damage to the pancreas may cause the release of various enzymes. Amongst these, elastase, is thought to play a role in the destruction of vessel walls in patients with pancreatitis.¹¹ Damage to vascular endothelium would further impair PGI₂ synthesis and expose deeper layers of vessels which may initiate platelet aggregation due to their collagen content. The release of trypsin is thought to activate the coagulation cascade, enhancing the ten-