Evaluation of Questor urine screening system for bacteriuria and pyuria

M Stevens, C J Mitchell, S A Livsey, C A MacDonald

Abstract

Aims—To evaluate the Questor automated bacteriuria and pyuria screening system; to compare its performance with that of a reference method; and to assess its usefulness in a routine clinical laboratory.

Methods—The Questor urine screening system was compared with a comprehensive regimen to detect urinary tract infection, using pour-plate viable counts to determine the numbers of bacteria present in urine samples, a wide range of other cultural methods, microscopic findings and clinical information.

Results—The optimal performance in detecting significant growths was a sensitivity of 93%, a specificity of 74%, a positive predictive value of 43% and a negative predictive value of 98%. The list price per test is £0.17 and the capital cost of the system is £39 950. Questor can test 50 samples an hour and can be operated by one member of the laboratory staff, who is not required to make interpretative judgments—for example, a medical laboratory assistant.

Conclusions—The sensitivity and specificity of the Questor was better than that obtained from other screening systems using the same protocol. The system was easy to use and is a useful addition to the methods available for screening for bacteriuria.

(J Clin Pathol 1993;46:817-821)

Department of Health
Evaluation Unit,
Public Health
Laboratory, The Royal
Infirmary, Leicester
LE1 5WW
M Stevens
C J Mitchell
S A Livsey

Medical Devices
Directorate
Department of
Health, Russell
Square, London
C A MacDonald

Correspondence to:
M Stevens

Accepted for publication
7 April 1993

Urinary infections are second only in prevalence to those of the upper respiratory tract. But the number of requests submitted to clinical microbiology laboratories for the microbiological examination of urine far exceeds the number of upper respiratory tract specimens. Bacteriuria may be chronic, persistent, and possibly influence structure and function outside the urinary tract. It has an important role in disease throughout life, from prematurity to hypertension and renal failure.¹ Many methods for screening for bacteriuria have been described and reviewed,² but there remains a need for an accurate, rapid, cost effective system for use in routine clinical laboratories. Questor (Difco Laboratories Ltd, East Molesey, Surrey) is an automated system which is designed to categorise and enumerate bacteria and other cells found in urine using particle counting technology.

An evaluation of the system was commissioned by the Medical Devices Directorate of the Department of Health and carried out at the Leicester Public Health Laboratory.

The system's performance was compared with a comprehensive reference method which included a variety of cultural techniques, microscopic examination, and clinical information. A full report of the evaluation has been produced.³

Methods

The Questor system (fig 1) is an automated single-channel impedance particle counter which uses a direct physical method to count and categorise particles. A 100 µl aliquot of the urine to be tested is pipetted into one well of an eight-well disposable plastic tray which is then placed in the instrument. The following automatic process is then carried out. The specimen is diluted with a conductive electrolyte solution and offered to a particle counting probe (fig 2). There is a small orifice (A) in the glass end of the probe which also contains two platinum electrodes, one internal (B) and one external (C). An electrical field is established between the electrodes via the electrolyte in the orifice. The sample is drawn through the orifice under vacuum. As a particle is drawn through the orifice and enters the electrical field, it displaces its own volume of electrolyte, causing a change in the impedance of the electrical circuit. The changes are detected as pulses, the number of which indicate the quantity of particles present in the sample.

The height and area of the pulse is propor-

Figure 1 Complete Questor system with computer.

Figure 2 Schematic representation of the Questor probe: (A) orifice; (B) internal electrode; (C) external electrode; (D) movable insert showing direction of movement (E).

tional to the mean spherical volume of the particle. A histogram of the particles and their sizes is formed and analysed by the computer software to give enumeration of the particles of different size bands. In Questor, the volume of the orifice is varied by inserting a tapered spike (the insert) (D) into and out of the orifice (A). Thus the characteristics of the orifice can be optimised for the size of the particles of interest (from 0.5 to 40 μm). The analyser records the volume distributions for four different particle size ranges corresponding to organisms, erythrocytes, leucocytes and epithelial cells.

Threshold settings to define a positive result can be selected by the user. If the organism count is above the selected threshold—for example, 1.0×10^5 —the interpretation given is “positive”. If the leucocyte count is above the selected threshold— $\geq 50/\mu\text{l}$ —the interpretation given is “pos-white cell”. If the erythrocyte count is above the selected threshold— $\geq 250/\mu\text{l}$ —the interpretation given is “pos-red blood cell”. The user has the option of choosing a threshold value for any combination of organisms, leucocytes, and erythrocytes above which results are interpreted as positive.

A previous evaluation of the Ramus urine screening system,⁴ which utilised particle counting, showed that the fixed, small size orifice used was prone to blocking. In the Questor system the manufacturers claim that, “putting the insert into the orifice forms an annulus. The flow pattern sets up a standing wave at the entrance of the annulus. This wave sweeps the large particles away from the annulus but allows the small particles through. This is called the ‘filter effect’, which greatly reduces the incidence of probe blockage.”

Apart from the Questor system, the items required for testing urines are plastic trays and particle-free isotonic diluent (both of which are supplied by the manufacturer), a

micropipette and tips, plastic waste disposal bags, and computer consumables: paper, ribbons and floppy disks.

Evaluation methods

The urine samples examined during the evaluation were random, unselected, consecutive specimens and were representative of a cross section of the types submitted to the routine clinical laboratory. A total of 1023 urine specimens were examined by Questor and by comprehensive comparative methods. Of the 1023, 1003 were received preserved in boric acid. The remaining 20 were unpreserved but were refrigerated and tested by Questor and the reference methods concurrently.

A pour-plate viable bacterial counting technique was used as the basis of the reference method. Urines were diluted in quarter strength Ringer solution to produce dilutions of 10^2 and 10^4 . One millilitre of each dilution and of the undiluted urine was mixed with 20 ml of sterile, molten Columbia agar base (Oxoid CM331) held at a temperature of 45°C. Plates were incubated overnight at 37°C and colonies were counted. The range of dilutions was designed to allow bacterial numbers in the range 0 to 10^7 cfu/ml to be counted. To facilitate the identification of organisms, a calibrated 1 μl loop was used to seed urine on to one half of a 9 cm diameter agar plate containing cystine-lactose-electrolyte deficient (CLED) medium (Oxoid CM301). Inoculated plates were incubated at 37°C for 18 hours. Isolates were identified using a variety of standard methods.⁵

Growths were considered to be clinically significant if a count of $\geq 10^5$ cfu/ml of no more than two types of organisms were grown or there was a predominant organism in a mixed culture. Counts between 10^4 and 10^5 cfu/ml were considered to be clinically significant if they were in pure culture and the patient was receiving antimicrobial treatment, or the urine contained ≥ 10 leucocytes/ μl .

Cultures designed to detect and enumerate carboxyphilic and obligately anaerobic bacteria were prepared as follows: each urine sample was cultured, using a 1/ μl loop on to blood agar and heated blood agar plates. These were incubated anaerobically and in an atmosphere of 10% CO_2 in air, respectively, for 48 hours at 37°C.

All urine samples were tested for the presence of antibacterial agents. A multipoint inoculator was used to apply urine to lawn cultures of *Staphylococcus aureus* (NCTC 6571) and *Escherichia coli* (NCTC 10418). These strains were chosen because of their susceptibility to antibiotics commonly used in the treatment of urinary tract infections. After overnight incubation, the presence of antibacterial agents was recorded where zones of inhibition occurred.

Statistical analyses

The results of individual tests from Questor and the reference method were categorised

Table 1 Categories of results comparing reference and Questor methods

Reference test result	Questor result	
	Positive	Negative
Positive	True positive	False negative
Negative	False positive	True negative

(table 1). Using these categories, the sensitivity, specificity, and positive and negative predictive values were calculated as follows⁶:

$$\text{Sensitivity (\%)} = \frac{\text{True positives}}{\text{True positives} + \text{false negatives}}$$

$$\text{Specificity (\%)} = \frac{\text{True negatives}}{\text{True negatives} + \text{false positives}}$$

$$\text{Positive predictive value (\%)} = \frac{\text{True positives}}{\text{True positives} + \text{false positives}}$$

$$\text{Negative predictive value (\%)} = \frac{\text{True negatives}}{\text{True negatives} + \text{false negatives}}$$

The performance of Questor was calculated by using varying threshold settings for the organism counts between 1.0×10^4 to 1.1×10^5 in steps of 0.1 log, together with leucocyte counts of 20, 30, 40 and 50 per μl and an erythrocyte count of ≥ 250 (table 2). From the many possible combinations of thresholds, the optimal performance was calculated using the method of Koenig *et al.*⁷ In this method sensitivity is weighted more heavily than specificity, because the consequences of reporting a false negative result are potentially more serious than those of a

Table 2 Performance of Questor at different positive thresholds

Questor organisms	Sensitivity (%)	Specificity (%)	Performance index
<i>Leucocytes 50 μl</i>			
5.0×10^4	93.2	59.0	83.0
6.0×10^4	93.2	65.3	84.8
7.0×10^4	92.7	71.3	86.3
8.0×10^4	91.0	75.8	86.4
9.0×10^4	89.8	77.7	86.2
1.0×10^5	89.3	78.4	86.0
1.1×10^5	88.7	79.3	86.9
<i>Leucocytes 40 μl</i>			
5.0×10^4	93.2	58.0	82.7
6.0×10^4	93.2	64.2	84.5
7.0×10^4	93.2	70.0	86.3
8.0×10^4	92.7	74.0	87.1
9.0×10^4	91.5	75.9	86.9
1.0×10^5	91.5	76.5	87.0
1.1×10^5	91.0	77.2	86.8
<i>Leucocytes 30 μl</i>			
5.0×10^4	93.8	57.1	82.8
6.0×10^4	93.8	63.0	84.6
7.0×10^4	93.8	68.3	86.2
8.0×10^4	93.2	71.8	86.8
9.0×10^4	92.1	74.0	86.7
1.0×10^5	92.1	74.0	86.7
1.1×10^5	91.5	74.5	86.4
<i>Leucocytes 20 μl</i>			
5.0×10^4	94.9	54.6	82.8
6.0×10^4	94.9	59.7	84.4
7.0×10^4	94.9	64.1	85.7
8.0×10^4	94.4	67.3	86.2
9.0×10^4	93.8	68.4	86.2
1.0×10^5	93.8	68.9	86.3
1.1×10^5	93.8	69.2	86.4

false positive result. The optimal performance was determined by calculating an index that emphasised sensitivity more than specificity in a ratio of 70:30–70% of the sensitivity was added to 30% of the specificity at each threshold combination (table 2). The 70:30 ratio is arbitrary; small changes in the ratio do not alter the optimal thresholds.

Results

Of the 1023 urine samples tested, 225 (22%) contained $\geq 10^5$ cfu/ml of bacteria as determined by the pour-plate counts. Of these 225, 167 were considered to be clinically significant growths. An additional 10 specimens contained between 10^4 and 10^5 cfu/ml and were considered to be clinically significant. Because the growth was pure, it was associated with the presence of 10 or more leucocytes and the clinical information indicated an infection. There were therefore a total of 177 significant growths.

Using the method of Koenig *et al.*⁷; the optimal performance of Questor was with the following thresholds: bacteria 8.0×10^4 , or leucocytes 40/ μl , or erythrocytes 250/ μl . The performance achieved at these thresholds is shown in table 3.

Of the 13 false negative results 10 grew pure cultures of *E. coli*, all of which had less than 10 leucocytes/ μl . One specimen grew *Enterococcus faecalis* and another *Proteus mirabilis*, associated with leucocyte counts of 1/ μl and 6/ μl , respectively. The clinical significance of these growths may be questioned. The remaining false negative grew *E. coli* and had a leucocyte count of 22/ μl .

Questor assigned the following interpretations to the false positive results: (n = number of samples in category): "positive" 109, "pos white cell" 84, and "pos-red blood cell" 27.

The leucocyte counts of the false positive results determined by microscopy is shown in table 4. Questor was thus able to detect a large number of leucocytes (≥ 50) in almost 50% of the false positive results.

Of the 220 false positive results, 86 (39%) contained anaerobic or carboxyphilic organisms and a further 74 (34%) contained a mixed bacterial population. Thus in 73% of

Table 3 Optimal performance of Questor

True positives	164
True negatives	626
False positives	220
False negatives	13
Sensitivity	92.7%
Specificity	74.0%
Positive predictive value	42.7%
Negative predictive value	98.0%

Table 4 False positive samples containing leucocytes

Number of leucocytes	Number of samples
0–9	72
10–19	17
20–29	5
30–39	8
40–49	9
50–99	21
≥ 100	88

the false positive results, organisms were present in numbers greater than 10^5 cfu/ml. Antibiotic substances were detected in a significantly higher percentage (47 of 220, or 21%) of the false positive samples compared with 147/1023 (14%) of all the samples tested ($p = 0.001$).

Questor assigned the following interpretation to the true positive results: (n = number of samples in category): "positive" 105, "pos-white cells" 48, and "pos-red blood cells" 10. Comparisons between the Questor leucocyte, erythrocyte, and epithelial cells counts and microscopy are given in tables 5, 6, and 7 respectively.

Discussion

The Questor system was more sensitive and specific than other bacteriuria screening systems evaluated using the same protocol.^{4,8,9}

The results showed that the system was able selectively to detect bacteria, leucocytes, erythrocytes and epithelial cells. The bacterial counts given were accurate, except below 1.0×10^4 and above 1.1×10^5 , and the evaluators suggest that as part of the installation process for individual laboratories, the software should be configured so that results outside these figures should read $<1.0 \times 10^4$ and $>1.0 \times 10^5$, respectively.

Although Questor recorded a false positive result for 220 specimens, many of these could be accounted for by the presence of anaerobic or carboxyphilic organisms, or the presence of antimicrobial substances and therefore the possibility of the presence of non-viable bacteria.

A total of 109 of the 220 false positive results contained high numbers of leucocytes

and although no organisms were detected in these samples, these high leucocyte counts have diagnostic value.

Questor has a tendency to give higher leucocyte and erythrocyte counts than is found by microscopy (tables 5 and 6).

Individual laboratories using Questor have the option of choosing any combination of different positive threshold levels for organisms, leucocytes, and erythrocytes. Changing these thresholds will alter the performance, increasing the sensitivity and lowering the specificity or vice versa. This may be important when testing urine samples from particular groups of patients. In a patient population with a low incidence of urinary tract infection the thresholds could be changed to increase the specificity; in a symptomatic group, a high sensitivity is required. It is recommended that each user laboratory should calibrate the system by comparing the results obtained with those of the best possible comparative method.

Because the operator involvement in the analytical process consists only of pipetting 100 μ l of urine sample into a tray and entering the sample number into the computer system, with no interpretive steps, a medical laboratory assistant can be used to operate the system.

In this evaluation it was possible for a medical laboratory assistant to operate the Questor system and to process 230 urine samples a day. As Questor takes about 1 minute to process one sample, in theory it would be possible for the system to test over 400 samples in a normal working day. This, however, would rely on specimens being received at regular intervals and therefore available for testing as a continuous workload throughout the day. In practice this did not occur, with most samples being received in mid-afternoon. Questor is a system designed to screen urines, allowing a negative result to be sent from the laboratory in most cases on the same day as specimen receipt.

This could have some beneficial impact on patient care, although this would depend greatly on both the speed at which the specimen is sent to the laboratory after voiding and how quickly the report is sent to the requesting physician. A benefit to the laboratory would be that the negative specimens would not be cultured and therefore not require the preparation and interpretation of culture plates by laboratory staff. Specimens recorded as positive by Questor would require culture, to identify causative organisms, to provide cultures for antimicrobial susceptibility tests, and to detect mixed, possibly insignificant growths.

The list price capital cost of the Questor system is £39 950. The list price costs of the consumables required, diluent, and trays total £0.17 a sample. The WELCAN workload system, when applied to all procedures associated with urine screening by Questor and a routine microscopical and cultural method, showed a saving of staff time of 31% when using the automated method.

Table 5 Leucocyte counts, Questor compared with microscopy Questor white cell count (per μ l)

Microscopy	0-9	10-29	30-49	≥ 50	Totals
0-9	486	129	29	14	658
10-29	14	44	9	16	83
30-49	0	13	18	13	44
≥ 50	1	8	8	221	238
Totals	501	194	64	264	1023

Table 6 Erythrocyte counts, Questor compared with microscopy Questor red blood cell count (per μ l)

Microscopy	0-9	10-29	30-49	≥ 50	Totals
0-9	524	141	30	37	732
10-29	17	43	13	5	78
30-49	1	10	9	11	31
≥ 50	6	23	28	125	182
Totals	548	217	80	155	1023

Table 7 Epithelial cells, Questor compared with microscopy Questor epithelial cell count (per μ l)

Microscopy	0-9	10-29	30-49	≥ 50	Totals
0-9	687	34	1	1	723
10-29	84	57	3	1	145
30-49	9	34	1	0	44
≥ 50	3	16	14	22	55
Totals	783	141	19	24	967

*Epithelial cells were not recorded by Questor for the remaining 56 specimens, due to the presence of an excess number of leucocytes.

The system was simple to use, was liked by laboratory staff, and had excellent, user-friendly software. The ability of the system to provide an accurate rapid, result, the savings in staff time taken to process samples make the Questor system an important and useful addition to the methods available for screening for bacteriuria.

This evaluation was supported financially by the Medical Devices Directorate, Department of Health.

- 1 Kass EH. Pyelonephritis and bacteriuria: a major problem in preventative medicine. *Ann Intern Med* 1962;56: 46-53.
- 2 Stevens M. Screening urines for bacteriuria. *Med Lab Sci*

1989;46:194-206.

- 3 Medical Devices Directorate. *An evaluation of the Questor urine screening system*. Report MDD/92/48. London: Department of Health, 1992.
- 4 Medical Devices Directorate. *An evaluation of the Orbec Ramus 256*. Report STD/88/13. London: Department of Health, 1988.
- 5 Cowan & Steel's *Manual for the Identification of Medical Bacteria*. 2nd Edn, Cambridge: Cambridge University Press, 1975.
- 6 Galen RS, Gambino SR. *Beyond Normality: The Predictive Value and Efficiency of Medical Diagnoses*. New York: John Wiley and Sons, 1975.
- 7 Koenig C, Tick LJ, Hanna BA. Analyses of the FlashTrack DNA probe and UTIscreen bioluminescence tests for bacteriuria. *J Clin Microbiol* 1992;30: 342-5.
- 8 Medical Devices Directorate. *An evaluation of the Autotrak urine screening system*. Report STD/88/28. London: Department of Health, 1988.
- 9 Medical Devices Directorate. *An evaluation of the Sensitive LeukoBact urine screen*. Report MDD/92/03. London: Department of Health, 1992.