- Scott, J. M., Ghanta, V., and Herbert, V., (1974). Trouble-free microbiologic serum and red cell folate assays. *American Journal of Medical Technology*, **40**, 125-134.
- Tennant, G. B. (1977). Continuous-flow automation of the *Lactobacillus casei* serum folate assay. *Journal of Clinical Pathology*, **30**, 1168-1174.
- Tennant, G. B., Newberry, C. H., Davies, J., and Dziedzic, H. (1974). A comparison of seven strains of *Lactobacillus casei* var. *rhamnosus* in relation to their use in the microbiological assay of serum folate. *Journal of Applied Bacteriology*, **37**, 373-389.
- Tennant, G. B., and Withey, J. L. (1972). An assessment of work simplified procedures for the microbiological assay of serum vitamin B₁₂ and serum folate. *Medical Laboratory Technology*, **29**, 171-181.

Letter to the Editor

Assay of antibacterial drugs using *Bacillus stearothermophilus*

In the description of the rapid microbiological gentamicin assay (Wahlig and Holt, 1976) it is suggested that the *Bacillus stearothermophilus* method could have wide applications. It is also suggested that the method is simple, accurate, and fast in both technical time and result. We have investigated this method and do not agree.

Sixteen specimens of serum spiked with known amounts of gentamicin were prepared from either known potency material or were samples used in previous quality control surveys: they ranged from 1.25 mg/l to 20 mg/l. The specimens were assayed by the acetylase method (Haas and Davies, 1973) or by the method described by the instructions in the Merck Refobacin-Test.

Using the same analysis as used in the Public Health Laboratory Service quality control survey (Reeves and Bywater, 1975) the mean per cent error plus two standard deviations was 33.3 for the microbiological test and 21.3 for the acetylase method. Thus the microbiological method would fall into the 'poor' group.

There were other problems using the high temperature method: (a) there was a tendency for the media to dry out; (b) getting the plates to float was not very easy; (c) the high humidity led to excess moisture being present and on several occasions the discs were washed off; (d) we found that the method took two hours to perform before incubation and considerable technical expertise was needed to set it up; (e) in our hands, no growth was visible for four hours in comparison with a total time to perform the acetylase method of 1½-2 hours.

We feel that this method cannot be recommended. It has little or no advantage over the acetylase method or other rapid

microbiological techniques (Shanson and Hince, 1977).

R. WISE

J. M. ANDREWS

Department of Medical Microbiology,
Dudley Road Hospital,
Dudley Road,
Birmingham B18 7QH,
UK

References

- Haas, M. J. and Davies, J. (1973). Enzymatic acetylation as a means of determining serum aminoglycoside concentrations. *Antimicrobial Agents and Chemotherapy*, **4**, 497-499.
- Reeves, D. S. and Bywater, M. J. (1975). Quality control of serum gentamicin assays—experience of national surveys. *Journal of Antimicrobial Chemotherapy*, **1**, 103-116.
- Shanson, D. C. and Hince, C. (1977). Serum gentamicin assays of 100 clinical serum samples by a rapid 40°C *Klebsiella* method compared with overnight plate diffusion and acetyltransferase assays. *Journal of Clinical Pathology*, **30**, 521-525.
- Wahlig, H. and Holt, R. J. (1976). A rapid microbiological procedure using *Bacillus stearothermophilus* for the assay of antibacterial drugs. *Journal of Clinical Pathology*, **29**, 858-861.

The authors have commented as follows:

It is difficult to account for the wide range of error in the results of Drs Wise and Andrews using the high temperature method described by us; in our individual hands and in those of several experienced colleagues in both Germany and the UK, much greater accuracy has been regularly attained during the performance of many hundred, even thousands, of clinical assays by this procedure.

We are also surprised by the technical difficulties met by Drs Wise and Andrews. In our hands the plates, plastic of course, float readily on water, growth appears quite obviously well within three hours, and during this time the plates do not become excessively damp nor do the discs float off. The length of preparation time required by these workers seemed to us very excessive and may be attributable to their unfamiliarity with this method. The obvious advantages of the high temperature method are not mentioned in their letter.

When writing this paper we recognized that the method and its variations reported by us would probably appear unorthodox and even heretical to some workers on microbiological assay procedures, of which we have between us over 50 years' continuous clinical experience, are notoriously subject to personal views and prejudices. Nevertheless we felt that a simple and rapid method, however unusual in technique, which gave reliable and accurate results in our and other experienced hands, should be available for workers in the field.

R. J. HOLT

Department of Microbiology,
Queen Mary's Hospital for Children,
Carshalton, Surrey, UK
HELMUT WAHLIG
E. Merck, Darmstadt,
Germany