

Technical methods

Serogrouping of *Legionella pneumophila* by double diffusion and counter immunoelectrophoresis

MALCOLM G HOLLIDAY,* JOHN BARKER†
*Department of Microbiology, St George's Hospital
Stafford, and the †Public Health Laboratory, East
Birmingham Hospital, Birmingham

Ten serogroups of *Legionella pneumophila* have now been described.¹ Although the direct fluorescent antibody technique is widely used to identify and serologically differentiate *L pneumophila* isolates,² some cross reactivity between serogroups³ and some staining of non-legionella organisms with legionella antisera⁴ have been reported. Several


simple alternatives to the direct fluorescent antibody technique have been described which do not require fluorescence microscopes and competent microscopists. These include slide agglutination,² staphylococcal coagglutination,⁵ latex agglutination,⁶ and antiserum-agar plates.⁷

Counter immunoelectrophoresis and double diffusion have been used to detect human antibodies to *L pneumophila*.^{8,9} The studies reported here suggest that both double diffusion and counter immunoelectrophoresis can be used as alternatives to direct fluorescent antibody when serogrouping isolates of *L pneumophila*.

Material and methods

Antisera to *L pneumophila* serogroups 1 to 6 were obtained from the Division of Microbiological

Accepted for publication 30 May 1984



(a) Counter immunoelectrophoresis slide (unstained) and (b) double diffusion slide (stained) showing urea extracts of *L pneumophila* strains tested against DMRQC serogroup 1 antiserum.

Counter immunoelectrophoresis wells: (A) antiserum, (1) serogroup 6 strain (Oxford-1) (2) serogroup 1 strain (Philadelphia-1), (3) serogroup 1 strain (Pontiac-1).

Double diffusion wells: (A) antiserum, (1) serogroup 1 strain (Philadelphia-1), (2) serogroup 1 strain (East Birmingham EB-1), (3) serogroup 2 strain (EB-2), (4) serogroup 3 strain (EB-3), (5) serogroup 4 strain (EB-4), (6) serogroup 5 strain (Cambridge-2), (7) serogroup 1 strain (Pontiac-1), (8) serogroup 6 strain (Oxford-1).

Technical methods

Results of testing 35 *Legionella pneumophila* strains, 132 non-*Legionella* strains, and calorifier and lung samples containing *L pneumophila* by counter immunoelectrophoresis and double diffusion

<i>L pneumophila</i> antisera	<i>L pneumophila</i> strains						Other bacterial species	Calorifier* deposit	Lung samples†
	Serogroup 1	2	3	4	5	6			
Serogroup 1	+	-	-	-	-	-	-	-	-
Serogroup 2	-	+	-	-	-	-	NT	-	-
Serogroup 3	-	-	+	-	-	-	NT	-	-
Serogroup 4	-	-	-	+	-	-	NT	-	-
Serogroup 5	-	-	-	-	+	-	NT	-	-
Serogroup 6	-	-	-	-	-	+	NT	-	-
Pooled 1-6	-	-	-	-	-	-	NT	-	-
Normal rabbit serum	-	-	-	-	-	-	NT	-	-
Total	12	5	3	1	3	11	132	1	4

**L pneumophila* serogroup 1 isolated.

†All from positive cases of Legionnaires' disease.

NT = not tested.

+ = positive reaction.

- = negative reaction.

Reagents and Quality Control (DMRQC), Colindale, London. Fluorescent antibody titres were between 4000 and 16 000.

An antiserum to *L pneumophila* serogroup 1 was also prepared in rabbits immunised with an ultrasonically disrupted antigen.⁸ This was designated Phil-1 antiserum. Fluorescent antibody titre was 5000.

Antigens for counter immunoelectrophoresis and double diffusion were extracted from 24 or 48 h old cultures on buffered charcoal yeast extract agar (Oxoid Ltd, Basingstoke, England). Organisms under test were suspended in 35 µl of phosphate buffered saline (PBS) in a glass bijou bottle to about 3 × 10¹⁰ colony forming units (CFU) /ml (roughly equivalent to two 3 cm sweeps through confluent growth). Diethyl ether (100 µl) was added, and the contents were shaken. The bottle was incubated at room temperature for 10 min and then 200 µl of 9 M urea was added, mixed, and the bottle left for a further 10 min. The mixture was centrifuged for 5 min at 2000 g and the supernatant was used as antigen.

Counter immunoelectrophoresis was performed as described previously⁸ except that the antigen and antibody wells were 3 mm in diameter and 3 mm apart and 10 µl volumes were employed in each (Fig (a)).

Double diffusion was carried out on glass microscope slides with 3 ml of 1% agarose (BDH) in PBS, pH 7.2. Nine wells (3 mm diameter) were made: one central antiserum well and eight others at 3 mm from it for antigen extracts (Figure (b)). The slides were incubated at room temperature overnight in a moist chamber.

Results

Serogroups determined by counter immunoelec-

trophoresis and double diffusion showed 100% correlation with previously determined fluorescent antibody serogroups when 35 strains of *L pneumophila* were tested (Table). No cross reactions between serogroups were seen. No reactions were noted when these 35 strains were tested against pooled serogroup 1 to 6 antiserum, or when 132 non-legionella strains were tested against serogroup 1 antiserum. A calorifier deposit known to contain *L pneumophila* and four lung samples from confirmed cases of Legionnaires' disease were repeatedly negative with all antisera by both counter immunoelectrophoresis and double diffusion.

Discussion

A positive counter immunoelectrophoresis or double diffusion result was denoted by the appearance of a single slightly diffuse, almost central, precipitin line (Figure), although one serogroup 3 strain and four serogroup 1 strains occasionally produced negative results when tested. These strains were all positive when repeated. No morphological differences were seen between these strains and others apart from slightly less luxuriant growth and a tendency for the urea extract to be slightly sticky after centrifugation; however, these occasional false negative results mean that all organisms examined by these techniques should be extracted and tested in duplicate in an attempt to eliminate this problem. Strains which persist in giving a negative reaction should be examined by conventional means or sent to a reference laboratory. *L pneumophila* strains cultured on legionella blood agar (Oxoid) gave comparable results to strains grown on charcoal yeast extract agar when tested by counter immunoelectrophoresis and double diffusion. Occasional non-legionella isolates such as *Pseudomonas*

fluorescens and *Bacteroides fragilis* have been shown to give positive reactions with *L pneumophila* antisera by double diffusion and direct fluorescent antibody,⁴ owing to shared antigens, but none of the 132 non-legionella strains tested here (which included 13 *P fluorescens* and 33 *B fragilis*) produced any reaction with serogroup 1 antisera (DMRQC or Phil-1) by counter immunoelectrophoresis or double diffusion.

The results indicate that counter immunoelectrophoresis and double diffusion are not sufficiently sensitive to detect the relatively small amount of antigen presumably present in clinical or environmental samples such as calorifier deposits and lung tissues. The lack of sensitivity of counter immunoelectrophoresis in detecting legionella antigens has been noted previously.¹⁰

Both counter immunoelectrophoresis and double diffusion can be used to group and identify strains of *L pneumophila* accurately. Double diffusion is a much simpler and more economic test than counter immunoelectrophoresis when testing multiple strains, as up to eight extracts can be tested simultaneously against 10 µl of each antiserum, whereas each extract tested requires 10 µl of antiserum by counter immunoelectrophoresis. Thus although both techniques could be used in laboratories where fluorescent antibody techniques are not available, double diffusion should be the method of choice.

The fact that none of the *L pneumophila* strains reacted with the pooled serogroup 1-6 antiserum is probably due to the consequent dilution of each individual antiserum, and it is suggested that only neat, high titre antisera such as those obtained from DMRQC should be used for counter immunoelectrophoresis and double diffusion.

We thank Dr AG Taylor, Division of Microbial Reagents and Quality Control, Colindale, London, for kindly supplying antisera and strains of *L*

pneumophila, and for much helpful advice and encouragement; Dr AD Macrae, Queen's Medical Centre, Nottingham, for kindly supplying positive lung tissue; and Dr J Gray and Mr A Barratt, PHLS, Stoke on Trent, for kindly supplying Phil-1 antiserum. Mrs D Johnson kindly typed the manuscript.

References

- Macfarlane JT. Legionnaires' disease: update. *Br Med J* 1983;287:443-4.
- Wilkinson HW, Fikes BJ. Slide agglutination test for serogrouping *Legionella pneumophila* and atypical Legionella-like organisms. *J Clin Microbiol* 1980;11:99-101.
- McKinney RM, Wilkinson HW, Sommers HM, et al. *Legionella pneumophila* serogroup six: Isolation from cases of Legionellosis. Identification by immunofluorescence staining and immunological response to infection. *J Clin Microbiol* 1980;12:395-401.
- Orrison LH, Bibb WF, Cherry WB, Thaker L. Determination of antigenic relationships among Legionellae and non-Legionellae by direct fluorescent-antibody and immunodiffusion tests. *J Clin Microbiol* 1983;17:332-7.
- Wilkinson HW, Fikes BJ. Detection of cell-associated or soluble antigens of *Legionella pneumophila* serogroups 1-6. *Legionella bozemanii*. *Legionella dumoffii*. *Legionella gormanii* and *Legionella micdadei* by Staphylococcal coagglutination tests. *J Clin Microbiol* 1981;14:322-5.
- Sedgwick AK, Tilton RC. Identification of *Legionella pneumophila* by latex agglutination. *J Clin Microbiol* 1983;17:365-8.
- Janssen WA, Hedlund KW. Antiserum-agar plate method for simultaneous detection and direct isolation of *Legionella* species in clinical and environmental specimens. *J Clin Microbiol* 1982;15:1176-7.
- Holliday MG. Counterimmunoelectrophoresis in the serodiagnosis of Legionnaires' disease. *J Clin Pathol* 1983;36:466-70.
- Soriano F, Aguilar L, Garces JLG. Simple immunodiffusion test for detecting antibodies against *Legionella pneumophila* serotype 1. *J Clin Microbiol* 1982;15:330-1.
- Smith RA, Digiorgio S, Darner J, Wilhelm A. Detection of *Legionella pneumophila* capsular like envelope antigens by counterimmunoelectrophoresis. *J Clin Microbiol* 1981;13:637-42.

Requests for reprints to: Mr MG Holliday, Department of Microbiology, St George's Hospital, Stafford, England.

New medium for enhancing pigment production of group B streptococci

SAMANIYA SUKROONGREUNG,
CHURAIRATANA NILAKUL,
SRISURANG TUNTIMAVANICH, SOM-
CHAI SANTIWATANAKUL *From the Department of Clinical Microbiology, Faculty of Medical Technology, Mahidol University, Bangkok 10700, Thailand*

Group B streptococci are important pathogens in neonates^{1,2} and in adults.^{3,4} It is important to identify group B streptococci either by serogrouping or by demonstrating one of their unique physiological characteristics. The sensitive and specific physiological method is pigment production.^{5,6} Specific media⁶⁻⁸ for enhancing pigment production of group B streptococci need anaerobic conditions and contain serum. The purpose of this study was to develop a new medium which does not require serum and anaerobic conditions.