

Some take more than a day to perform while others require expensive equipment. The agglutination method is rapid, reproducible, inexpensive and easy to perform in a laboratory with a platelet aggre-gometer. Because single measurements of CRP concentrations give limited information this method could be used as a screening method in the serial monitoring of at risk patients—for example, those with leukaemia or those in the neonatal intensive care unit.

#### References

- 1 Osmond AP, Friedman H, Gewurz H, Painter RH, Hofman T, Shelton E. Characterization of C-reactive protein and the complement component C1t as homologous proteins displaying cyclic pentameric symmetry (pentaxins). *Proc Natl Acad Sci USA* 1977;74:739.
- 2 Kushner I, Broder ML, Karp D. Control of the acute phase response. *J Clin Invest* 1978;61:235.
- 3 Gozzard DI, Lin Yin JA, Delamore LW. The clinical usefulness of C-reactive protein measurements. *Br J Haematol* 1986;63:411-13.
- 4 Tillet WS, Francis T. Serological reaction in pneumonia with a non-protein somatic fraction of non-pneumococcus. *J Exp Med* 1930;52:561-71.
- 5 McCord FB, Jenkins JG, Lim JHA. C-reactive protein concentration as screening test for bacterial infection in febrile children. *Br Med J* 1985;291:1685-86.
- 6 Mackie PH, Crockson RA, Stuart J. C-reactive protein for rapid diagnosis of infection in leukemia. *J Clin Pathol* 1979; 32:1253-56.
- 7 Sabel KG, Wadsworth C. C-reactive protein in early diagnosis of neonatal septicemia. *Acta Paediatr Scand* 1979;68:825-31.
- 8 Hulman G, Pearson HJ, Fraser I, Bell PRF. Agglutination of intralipid by sera of acutely ill patients. *Lancet* 1982;ii:1426-27.
- 9 Tsujimoto M, Inoue K, Nojima S. C-reactive protein induced agglutination of lipid suspensions prepared in the presence and absence of phosphatidylcholine. *J Biochem* 1980;87:1531-37.
- 10 Adam A, Ers P, Herman G, Stas JL. Analytical evaluation of a new latex agglutination test for quantitative determination of C-reactive protein by laser nephelometry. *J Clin Chem Clin Biochem* 1985;23:787-89.
- 11 Rose PE, Johnson SA, Meakin M, et al. Serial study of C-reactive protein during infection in leukemia. *J Clin Pathol* 1981;34:263-6.
- 12 Mancini G, Carbonara AO, Heremans JF. Immunochemical quantitation of antigens by single radial immunodiffusion. *Immunochemistry* 1965;2:235-54.
- 13 Wadsworth CH. A rapid spot immunoprecipitate assay method applied to quantitating C-reactive protein (CRP) in pediatric sera. *Scand J Immunol* 1977;6:1263.

Requests for reprints to: Dr PA Gordon, Department of Laboratory Medicine, University of Alberta Hospitals, 8440-112th Street, Edmonton, Alberta, T6G 2B7, Canada.

## Letters to the Editor

### Spurious increase in plasma potassium concentration and reduction in plasma calcium due to in vitro contamination with liquid potassium edetic acid at phlebotomy

Recently an increased number of spuriously raised plasma potassium values were noted in the routine workload of our biochemistry laboratory, an incidence of about 1 in 1000 samples. Six months previously a blood tube (Sarstedt) containing liquid potassium edetic acid anticoagulant had been introduced by the haematology department. Obvious sources of plasma contamination, such as aged samples, haemolysis, sampling from intravenous lines, and batch contamination of the lithium heparin tubes used for plasma potassium determination were excluded. It seemed plausible that the increased potassium values could have been caused by droplet transfer of liquid potassium edetic acid anticoagulant from the Sarstedt tubes to lithium heparin tubes when the latter were being filled from a syringe after phlebotomy.

To test this hypothesis 30 Sarstedt tubes containing visible droplets of liquid potassium edetic acid at their apertures and 40 lithium heparin bottles from a single batch were selected. Ten millilitres of pooled heparinised blood was drawn up into each of 40

fresh syringes. Thirty test samples were prepared by injecting 2 ml of blood from each syringe into a blood tube containing liquid potassium edetic acid, with the syringe tip resting against the side of the tube, and then dispensing the remaining 8 ml of blood into a lithium heparin bottle. Pooled blood dispensed directly from 10 fresh syringes into 10 lithium heparin bottles served as controls.

Blood in the lithium heparin tubes was analysed for potassium using flame photometry and for calcium by the o-cresolphthalein complexone dye method, which detects only non-chelated calcium and thus gives a low plasma calcium value in the presence of edetic acid.

The table shows our results. Obvious differences were noted between control and test mean plasma potassium values and mean plasma calcium values measured by the dye method.

Spurious hyperkalaemia and hypo-

calcaemia can clearly arise due to in vitro contamination with liquid potassium edetic acid. This may therefore be an unsuitable form of anticoagulant for clinical practice unless strict precautions are taken to prevent cross contamination.

MF FITZPATRICK\*  
J NEWELL†  
H GRIMES†  
EL EGAN\*

Departments of †Biochemistry and  
\*Haematology,  
Regional Hospital,  
Galway,  
Ireland

#### References

- 1 Wills MJ, Fraser ID. Spurious hyperkalaemia. *J Clin Pathol* 1964;17:649-50.
- 2 Zilva JF, Pannall PR. *Clinical chemistry - in diagnosis and treatment*. 4th ed. London: Lloyd-Luke Ltd, 1984:505.

Table Mean (SEM) plasma potassium and mean plasma calcium values in control and test plasma samples

Group	No of patient samples	Plasma potassium (mmol l <sup>-1</sup> )	Plasma calcium* (mmol l <sup>-1</sup> )
Control	10	4.65 (0.1)	2.10 (0.08)
Test	30	5.65 (0.59) range 4.5-7.2	1.65 (0.24) range 0.83-2.02

\*Plasma calcium estimated using Technicon SMA12/60 dye method (o-cresolphthalein complexone)

J Clin Pathol: first published as 10.1136/jcp.40.5.588 on 1 May 1987. Downloaded from http://jcp.bmj.com/ on March 19, 2023 by guest. Protected by copyright.